


LB 116990

SFAM01.1

INSTRUMENT 111, CV1



A= 0.0000e+000

B= 2.5246e-004

C= 7.3386e-003

Rho= 0.9999840

Accept=Accepted

slope
y=mx+ceps

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%d
0.00	0.000	0.003	0.003	-17	0.000	-17					=
0.05	0.050					205	0				6
0.20	0.200	0.212	0.012	810	0.0 %	810					-
2.50	2.500	2.463	-0.037	9729	0.0 %	9729					1
5.00	5.000	5.032	0.032	19902	0.0 %	19902					1
7.50	7.500	7.489	-0.011	29634	0.0 %	29634					0
10.0	10.000	10.001	0.001	39587	0.0 %	39587					0

LB116990
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-17	-		0 SFAM01.1	PPB	10/23/2021 13:33	Std
	0.2 S01	810	-		0.2 SFAM01.1	PPB	10/23/2021 13:36	Std
	2.5 S02	9729	-		2.5 SFAM01.1	PPB	10/23/2021 13:38	Std
	5 S03	19902	-		5 SFAM01.1	PPB	10/23/2021 13:40	Std
	7.5 S04	29634	-		7.5 SFAM01.1	PPB	10/23/2021 13:43	Std
	10 S05	39587	-		10 SFAM01.1	PPB	10/23/2021 13:47	Std
ICV085	ICV085	15970	4.0391 -		SFAM01.1	PPB	10/23/2021 13:50	SMPL
ICB085	ICB085	-58	-0.0073 -		SFAM01.1	PPB	10/23/2021 13:52	SMPL
CCV032	CCV032	20991	5.3066 -		SFAM01.1	PPB	10/23/2021 13:55	SMPL
CCB032	CCB032	-71	-0.0106 -		SFAM01.1	PPB	10/23/2021 13:57	SMPL
PB140172BL	PBW172	67	0.0243 -		SFAM01.1	PPB	10/23/2021 14:02	SMPL
M4231-01	MH1T45	283	0.0788 -		SFAM01.1	PPB	10/23/2021 14:04	SMPL
M4231-02	MH1T47	87	0.0293 -		SFAM01.1	PPB	10/23/2021 14:06	SMPL
M4231-03	MH1T49	92	0.0306 -		SFAM01.1	PPB	10/23/2021 14:08	SMPL
M4231-04	MH1T51	73	0.0258 -		SFAM01.1	PPB	10/23/2021 14:11	SMPL
M4231-05	MH1T53	97	0.0318 -		SFAM01.1	PPB	10/23/2021 14:13	SMPL
M4231-06	MH1T55	832	0.2174 -		SFAM01.1	PPB	10/23/2021 14:15	SMPL
M4231-07	MH1T57	69	0.0248 -		SFAM01.1	PPB	10/23/2021 14:17	SMPL
M4231-08	MH1T59	81	0.0278 -		SFAM01.1	PPB	10/23/2021 14:20	SMPL
M4231-09	MH1T61	74	0.026 -		SFAM01.1	PPB	10/23/2021 14:22	SMPL
M4231-10	MH1T63	112	0.0356 -		SFAM01.1	PPB	10/23/2021 14:24	SMPL
M4231-11	MH1T65	102	0.0331 -		SFAM01.1	PPB	10/23/2021 14:27	SMPL
M4231-12	MH1T67	86	0.029 -		SFAM01.1	PPB	10/23/2021 14:29	SMPL
M4231-13	MH1T69	619	0.1636 -		SFAM01.1	PPB	10/23/2021 14:31	SMPL
M4231-14	MH1T71	304	0.0841 -		SFAM01.1	PPB	10/23/2021 14:33	SMPL
M4231-15	MH1T73	19	0.0121 -		SFAM01.1	PPB	10/23/2021 14:36	SMPL
M4231-16	MH1T75	29	0.0147 -		SFAM01.1	PPB	10/23/2021 14:38	SMPL
M4231-17	MH1T77	67	0.0243 -		SFAM01.1	PPB	10/23/2021 14:40	SMPL
M4231-18	MH1T79	45	0.0187 -		SFAM01.1	PPB	10/23/2021 14:42	SMPL
M4231-19	MH1T81	74	0.026 -		SFAM01.1	PPB	10/23/2021 14:45	SMPL
CCV033	CCV033	21333	5.393 -		SFAM01.1	PPB	10/23/2021 14:47	SMPL
CCB033	CCB033	-12	0.0043 -		SFAM01.1	PPB	10/23/2021 14:49	SMPL
M4231-20	MH1T83	425	0.1146 -		SFAM01.1	PPB	10/23/2021 14:51	SMPL
M4231-21	MH1T83D	431	0.1161 -		SFAM01.1	PPB	10/23/2021 14:54	SMPL
M4231-22	MH1T83S	3703	0.9422 -		SFAM01.1	PPB	10/23/2021 14:56	SMPL
PB140173BL	PBW173	-17	0.003 -		SFAM01.1	PPB	10/23/2021 14:58	SMPL
M4232-01	MH1T85	394	0.1068 -		SFAM01.1	PPB	10/23/2021 15:00	SMPL
M4232-02	MH1T87	434	0.1169 -		SFAM01.1	PPB	10/23/2021 15:03	SMPL
M4232-03	MH1T89	404	0.1093 -		SFAM01.1	PPB	10/23/2021 15:05	SMPL
M4232-04	MH1T91	384	0.1043 -		SFAM01.1	PPB	10/23/2021 15:07	SMPL
M4232-05	MH1T93	350	0.0957 -		SFAM01.1	PPB	10/23/2021 15:10	SMPL
M4232-06	MH1T95	216	0.0619 -		SFAM01.1	PPB	10/23/2021 15:12	SMPL
M4232-07	MH1T97	35	0.0162 -		SFAM01.1	PPB	10/23/2021 15:14	SMPL
M4232-08	MH1T99	54	0.021 -		SFAM01.1	PPB	10/23/2021 15:16	SMPL
M4232-09	MH1U02	40	0.0174 -		SFAM01.1	PPB	10/23/2021 15:19	SMPL
M4232-10	MH1U04	43	0.0182 -		SFAM01.1	PPB	10/23/2021 15:21	SMPL

LB116990
INSTRUMENT ID : CV1

M4232-11	MH1U06	38	0.0169 -	SFAM01.1	PPB	10/23/2021 15:23 SMPL
M4232-12	MH1U08	41	0.0177 -	SFAM01.1	PPB	10/23/2021 15:25 SMPL
M4232-13	MH1U10	34	0.0159 -	SFAM01.1	PPB	10/23/2021 15:28 SMPL
M4232-14	MH1U12	30	0.0149 -	SFAM01.1	PPB	10/23/2021 15:30 SMPL
M4232-15	MH1U14	38	0.0169 -	SFAM01.1	PPB	10/23/2021 15:32 SMPL
M4232-16	MH1U16	29	0.0147 -	SFAM01.1	PPB	10/23/2021 15:34 SMPL
CCV034	CCV034	21412	5.4129 -	SFAM01.1	PPB	10/23/2021 15:37 SMPL
CCB034	CCB034	-29	0 -	SFAM01.1	PPB	10/23/2021 15:39 SMPL
M4232-17	MH1U18	28	0.0144 -	SFAM01.1	PPB	10/23/2021 15:41 SMPL
M4232-18	MH1U20	490	0.131 -	SFAM01.1	PPB	10/23/2021 15:44 SMPL
M4232-19	MH1U22	371	0.101 -	SFAM01.1	PPB	10/23/2021 15:46 SMPL
M4232-20	MH1V33	452	0.1214 -	SFAM01.1	PPB	10/23/2021 15:48 SMPL
M4232-21	MH1V33D	440	0.1184 -	SFAM01.1	PPB	10/23/2021 15:50 SMPL
M4232-22	MH1V33S	3794	0.9652 -	SFAM01.1	PPB	10/23/2021 15:53 SMPL
PB140174BL	PBW174	17	0.0116 -	SFAM01.1	PPB	10/23/2021 15:55 SMPL
M4233-01	MH1U24	30	0.0149 -	SFAM01.1	PPB	10/23/2021 15:57 SMPL
M4233-02	MH1U26	53	0.0207 -	SFAM01.1	PPB	10/23/2021 15:59 SMPL
M4233-03	MH1U28	63	0.0232 -	SFAM01.1	PPB	10/23/2021 16:02 SMPL
M4233-04	MH1U30	68	0.0245 -	SFAM01.1	PPB	10/23/2021 16:04 SMPL
M4233-05	MH1U32	86	0.029 -	SFAM01.1	PPB	10/23/2021 16:06 SMPL
M4233-06	MH1U34	103	0.0333 -	SFAM01.1	PPB	10/23/2021 16:09 SMPL
M4233-07	MH1U36	88	0.0296 -	SFAM01.1	PPB	10/23/2021 16:11 SMPL
M4233-08	MH1U38	81	0.0278 -	SFAM01.1	PPB	10/23/2021 16:13 SMPL
M4233-09	MH1U40	62	0.023 -	SFAM01.1	PPB	10/23/2021 16:15 SMPL
M4233-10	MH1U42	98	0.0321 -	SFAM01.1	PPB	10/23/2021 16:18 SMPL
M4233-11	MH1U44	72	0.0255 -	SFAM01.1	PPB	10/23/2021 16:20 SMPL
M4233-12	MH1U46	93	0.0308 -	SFAM01.1	PPB	10/23/2021 16:22 SMPL
M4233-13	MH1U47	72	0.0255 -	SFAM01.1	PPB	10/23/2021 16:24 SMPL
CCV035	CCV035	20908	5.2857 -	SFAM01.1	PPB	10/23/2021 16:27 SMPL
CCB035	CCB035	0	0.0073 -	SFAM01.1	PPB	10/23/2021 16:29 SMPL
M4233-14	MH1U48	67	0.0243 -	SFAM01.1	PPB	10/23/2021 16:31 SMPL
M4233-15	MH1U49	97	0.0318 -	SFAM01.1	PPB	10/23/2021 16:33 SMPL
M4233-16	MH1U50	893	0.2328 -	SFAM01.1	PPB	10/23/2021 16:36 SMPL
M4233-17	MH1U51	333	0.0914 -	SFAM01.1	PPB	10/23/2021 16:38 SMPL
M4233-18	MH1U52	55	0.0212 -	SFAM01.1	PPB	10/23/2021 16:40 SMPL
M4233-19	MH1U53	87	0.0293 -	SFAM01.1	PPB	10/23/2021 16:43 SMPL
M4233-20	MH1U54	97	0.0318 -	SFAM01.1	PPB	10/23/2021 16:45 SMPL
M4233-21	MH1U54D	97	0.0318 -	SFAM01.1	PPB	10/23/2021 16:47 SMPL
M4233-22	MH1U54S	4610	1.1712 -	SFAM01.1	PPB	10/23/2021 16:52 SMPL
PB140175BL	PBW175	19	0.0121 -	SFAM01.1	PPB	10/23/2021 16:54 SMPL
M4234-01	MH1U55	86	0.029 -	SFAM01.1	PPB	10/23/2021 16:57 SMPL
M4234-02	MH1V32	78	0.027 -	SFAM01.1	PPB	10/23/2021 16:59 SMPL
M4234-03	MH1U57	79	0.0273 -	SFAM01.1	PPB	10/23/2021 17:01 SMPL
M4234-04	MH1U58	69	0.0248 -	SFAM01.1	PPB	10/23/2021 17:03 SMPL
M4234-05	MH1U59	89	0.0298 -	SFAM01.1	PPB	10/23/2021 17:06 SMPL
M4234-06	MH1U60	94	0.0311 -	SFAM01.1	PPB	10/23/2021 17:08 SMPL
M4234-07	MH1U61	91	0.0303 -	SFAM01.1	PPB	10/23/2021 17:10 SMPL

LB116990
INSTRUMENT ID : CV1

M4234-08	MH1U62	83	0.0283 -	SFAM01.1	PPB	10/23/2021 17:12 SMPL
M4234-09	MH1U63	78	0.027 -	SFAM01.1	PPB	10/23/2021 17:15 SMPL
M4234-10	MH1U64	84	0.0285 -	SFAM01.1	PPB	10/23/2021 17:17 SMPL
CCV036	CCV036	20705	5.2344 -	SFAM01.1	PPB	10/23/2021 17:19 SMPL
CCB036	CCB036	4	0.0083 -	SFAM01.1	PPB	10/23/2021 17:22 SMPL
M4234-11	MH1U65	77	0.0268 -	SFAM01.1	PPB	10/23/2021 17:24 SMPL
M4234-12	MH1U66	87	0.0293 -	SFAM01.1	PPB	10/23/2021 17:26 SMPL
M4234-13	MH1U67	99	0.0323 -	SFAM01.1	PPB	10/23/2021 17:28 SMPL
M4234-14	MH1U68	67	0.0243 -	SFAM01.1	PPB	10/23/2021 17:31 SMPL
M4234-15	MH1U69	72	0.0255 -	SFAM01.1	PPB	10/23/2021 17:33 SMPL
M4234-16	MH1U70	69	0.0248 -	SFAM01.1	PPB	10/23/2021 17:35 SMPL
M4234-17	MH1U71	71	0.0253 -	SFAM01.1	PPB	10/23/2021 17:37 SMPL
M4234-18	MH1U72	70	0.025 -	SFAM01.1	PPB	10/23/2021 17:40 SMPL
M4234-19	MH1U73	81	0.0278 -	SFAM01.1	PPB	10/23/2021 17:42 SMPL
M4234-20	MH1U74	106	0.0341 -	SFAM01.1	PPB	10/23/2021 17:44 SMPL
M4234-21	MH1U74D	108	0.0346 -	SFAM01.1	PPB	10/23/2021 17:47 SMPL
M4234-22	MH1U74S	4282	1.0884 -	SFAM01.1	PPB	10/23/2021 17:51 SMPL
CCV037	CCV037	20487	5.1794 -	SFAM01.1	PPB	10/23/2021 17:54 SMPL
CCB037	CCB037	-4	0.0063 -	SFAM01.1	PPB	10/23/2021 17:56 SMPL