


SFAM01.1

INSTRUMENT ID: CY1

Linear



A= 0.0000e+000

B = 2.3365e-004 *slope*

C = 1.9250×10^{-2} $y = \text{mole} \cdot \text{cm}^{-3}$

Rho= 0.9996143

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.050	0.050	131	0.000	131					-
0.05	0.050					205	0				-
0.20	0.200	0.204	0.004	790	0.0 %	790					2
2.50	2.500	2.487	-0.013	10560	0.0 %	10560					-1
5.00	5.000	5.027	0.027	21432	0.0 %	21432					1
7.50	7.500	7.297	-0.203	31148	0.0 %	31148					-3
10.0	10.000	10.136	0.136	43297	0.0 %	43297					.

$$\begin{array}{r} 0/00 \\ \hline - \\ 2 \\ -1 \\ 1 \\ -3 \\ 1 \end{array}$$

LB116994
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0	131	-		0 SFAM01.1	PPB	10/24/2021 12:32	Std
	0.2	790	-		0.2 SFAM01.1	PPB	10/24/2021 12:34	Std
	2.5	10560	-		2.5 SFAM01.1	PPB	10/24/2021 12:36	Std
	5	21432	-		5 SFAM01.1	PPB	10/24/2021 12:39	Std
	7.5	31148	-		7.5 SFAM01.1	PPB	10/24/2021 12:41	Std
	10	43297	-		10 SFAM01.1	PPB	10/24/2021 12:43	Std
ICV086	ICV086	16567	3.8902	-	SFAM01.1	PPB	10/24/2021 12:46	SMPL
ICB086	ICB086	16	0.023	-	SFAM01.1	PPB	10/24/2021 12:48	SMPL
CCV038	CCV038	21081	4.9449	-	SFAM01.1	PPB	10/24/2021 12:53	SMPL
CCB038	CCB038	23	0.0246	-	SFAM01.1	PPB	10/24/2021 12:55	SMPL
PB140194BL	PBW194	100	0.0426	-	SFAM01.1	PPB	10/24/2021 12:57	SMPL
M4235-01	MH1U75	89	0.04	-	SFAM01.1	PPB	10/24/2021 13:00	SMPL
M4235-02	MH1U76	78	0.0375	-	SFAM01.1	PPB	10/24/2021 13:02	SMPL
M4235-03	MH1V17	89	0.04	-	SFAM01.1	PPB	10/24/2021 13:04	SMPL
M4235-04	MH1V17D	110	0.045	-	SFAM01.1	PPB	10/24/2021 13:06	SMPL
M4235-05	MH1V17S	4954	1.1768	-	SFAM01.1	PPB	10/24/2021 13:09	SMPL
M4235-06	MH1V18	362	0.1038	-	SFAM01.1	PPB	10/24/2021 13:11	SMPL
M4235-07	MH1V19	78	0.0375	-	SFAM01.1	PPB	10/24/2021 13:13	SMPL
M4235-08	MH1V20	82	0.0384	-	SFAM01.1	PPB	10/24/2021 13:16	SMPL
M4235-09	MH1V21	1433	0.3541	-	SFAM01.1	PPB	10/24/2021 13:18	SMPL
M4235-10	MH1V22	58	0.0328	-	SFAM01.1	PPB	10/24/2021 13:20	SMPL
M4235-11	MH1V23	110	0.045	-	SFAM01.1	PPB	10/24/2021 13:22	SMPL
M4235-12	MH1V24	99	0.0424	-	SFAM01.1	PPB	10/24/2021 13:25	SMPL
M4235-13	MH1V25	91	0.0405	-	SFAM01.1	PPB	10/24/2021 13:27	SMPL
M4235-14	MH1V26	113	0.0457	-	SFAM01.1	PPB	10/24/2021 13:29	SMPL
M4235-15	MH1V27	140	0.052	-	SFAM01.1	PPB	10/24/2021 13:31	SMPL
M4235-16	MH1V28	116	0.0464	-	SFAM01.1	PPB	10/24/2021 13:34	SMPL
M4235-17	MH1V29	97	0.0419	-	SFAM01.1	PPB	10/24/2021 13:36	SMPL
M4235-18	MH1V30	123	0.048	-	SFAM01.1	PPB	10/24/2021 13:38	SMPL
M4235-19	MH1V31	130	0.0496	-	SFAM01.1	PPB	10/24/2021 13:40	SMPL
CCV039	CCV039	20238	4.7479	-	SFAM01.1	PPB	10/24/2021 13:43	SMPL
CCB039	CCB039	57	0.0326	-	SFAM01.1	PPB	10/24/2021 13:45	SMPL
M4235-20	MH1V35	126	0.0487	-	SFAM01.1	PPB	10/24/2021 13:47	SMPL
M4235-21	MH1V37	144	0.0529	-	SFAM01.1	PPB	10/24/2021 13:50	SMPL
M4235-22	MH1V39	136	0.051	-	SFAM01.1	PPB	10/24/2021 13:52	SMPL
PB140195BL	PBW195	122	0.0478	-	SFAM01.1	PPB	10/24/2021 13:54	SMPL
M4236-01	MH1S99	125	0.0485	-	SFAM01.1	PPB	10/24/2021 13:56	SMPL
M4236-02	MH1T01	68	0.0351	-	SFAM01.1	PPB	10/24/2021 13:59	SMPL
M4236-03	MH1T02	138	0.0515	-	SFAM01.1	PPB	10/24/2021 14:01	SMPL
M4236-04	MH1T03	159	0.0564	-	SFAM01.1	PPB	10/24/2021 14:03	SMPL
M4236-05	MH1T04	138	0.0515	-	SFAM01.1	PPB	10/24/2021 14:05	SMPL
M4236-06	MH1T05	156	0.0557	-	SFAM01.1	PPB	10/24/2021 14:08	SMPL
M4236-07	MH1T06	78	0.0375	-	SFAM01.1	PPB	10/24/2021 14:10	SMPL
M4236-08	MH1T07	73	0.0363	-	SFAM01.1	PPB	10/24/2021 14:12	SMPL
M4236-09	MH1T08	72	0.0361	-	SFAM01.1	PPB	10/24/2021 14:14	SMPL
M4236-10	MH1T10	79	0.0377	-	SFAM01.1	PPB	10/24/2021 14:17	SMPL

LB116994
INSTRUMENT ID : CV1

M4236-11	MH1T12	71	0.0358 -	SFAM01.1	PPB	10/24/2021 14:19	SMPL
M4236-12	MH1T14	97	0.0419 -	SFAM01.1	PPB	10/24/2021 14:21	SMPL
M4236-13	MH1T16	89	0.04 -	SFAM01.1	PPB	10/24/2021 14:23	SMPL
M4236-14	MH1T18	103	0.0433 -	SFAM01.1	PPB	10/24/2021 14:26	SMPL
M4236-15	MH1T20	105	0.0438 -	SFAM01.1	PPB	10/24/2021 14:28	SMPL
M4236-16	MH1T22	90	0.0403 -	SFAM01.1	PPB	10/24/2021 14:30	SMPL
CCV040	CCV040	19892	4.6671 -	SFAM01.1	PPB	10/24/2021 14:33	SMPL
CCB040	CCB040	41	0.0288 -	SFAM01.1	PPB	10/24/2021 14:35	SMPL
M4236-17	MH1T24	97	0.0419 -	SFAM01.1	PPB	10/24/2021 14:37	SMPL
M4236-18	MH1T26	112	0.0454 -	SFAM01.1	PPB	10/24/2021 14:39	SMPL
M4236-19	MH1T28	94	0.0412 -	SFAM01.1	PPB	10/24/2021 14:42	SMPL
M4236-20	MH1T30	139	0.0517 -	SFAM01.1	PPB	10/24/2021 14:44	SMPL
M4236-21	MH1T30D	110	0.045 -	SFAM01.1	PPB	10/24/2021 14:46	SMPL
M4236-22	MH1T30S	4008	0.9557 -	SFAM01.1	PPB	10/24/2021 14:48	SMPL
PB140196BL	PBW196	78	0.0375 -	SFAM01.1	PPB	10/24/2021 14:51	SMPL
M4237-01	MH1T32	279	0.0844 -	SFAM01.1	PPB	10/24/2021 14:53	SMPL
M4237-02	MH1T34	535	0.1443 -	SFAM01.1	PPB	10/24/2021 14:55	SMPL
M4237-03	MH1T36	203	0.0667 -	SFAM01.1	PPB	10/24/2021 14:57	SMPL
M4237-04	MH1T38	159	0.0564 -	SFAM01.1	PPB	10/24/2021 15:00	SMPL
M4237-05	MH1T40	391	0.1106 -	SFAM01.1	PPB	10/24/2021 15:02	SMPL
M4237-06	MH1T42	134	0.0506 -	SFAM01.1	PPB	10/24/2021 15:04	SMPL
M4237-07	MH1T44	126	0.0487 -	SFAM01.1	PPB	10/24/2021 15:07	SMPL
M4237-08	MH1T46	93	0.041 -	SFAM01.1	PPB	10/24/2021 15:09	SMPL
M4237-09	MH1T48	95	0.0414 -	SFAM01.1	PPB	10/24/2021 15:11	SMPL
M4237-10	MH1T50	151	0.0545 -	SFAM01.1	PPB	10/24/2021 15:13	SMPL
M4237-11	MH1T52	114	0.0459 -	SFAM01.1	PPB	10/24/2021 15:16	SMPL
M4237-12	MH1T54	201	0.0662 -	SFAM01.1	PPB	10/24/2021 15:18	SMPL
M4237-13	MH1T56	102	0.0431 -	SFAM01.1	PPB	10/24/2021 15:20	SMPL
CCV041	CCV041	19616	4.6026 -	SFAM01.1	PPB	10/24/2021 15:22	SMPL
CCB041	CCB041	59	0.033 -	SFAM01.1	PPB	10/24/2021 15:25	SMPL
M4237-14	MH1T58	94	0.0412 -	SFAM01.1	PPB	10/24/2021 15:27	SMPL
M4237-15	MH1T60	110	0.045 -	SFAM01.1	PPB	10/24/2021 15:29	SMPL
M4237-16	MH1T62	112	0.0454 -	SFAM01.1	PPB	10/24/2021 15:32	SMPL
M4237-17	MH1T64	121	0.0475 -	SFAM01.1	PPB	10/24/2021 15:34	SMPL
M4237-18	MH1T66	86	0.0393 -	SFAM01.1	PPB	10/24/2021 15:36	SMPL
M4237-19	MH1T68	377	0.1073 -	SFAM01.1	PPB	10/24/2021 15:38	SMPL
M4237-20	MH1T70	234	0.0739 -	SFAM01.1	PPB	10/24/2021 15:41	SMPL
M4237-21	MH1T70D	172	0.0594 -	SFAM01.1	PPB	10/24/2021 15:43	SMPL
M4237-22	MH1T70S	3852	0.9193 -	SFAM01.1	PPB	10/24/2021 15:45	SMPL
PB140197BL	PBW197	5	0.0204 -	SFAM01.1	PPB	10/24/2021 15:50	SMPL
M4238-01	MH1T72	145	0.0531 -	SFAM01.1	PPB	10/24/2021 15:53	SMPL
M4238-02	MH1T74	81	0.0382 -	SFAM01.1	PPB	10/24/2021 15:55	SMPL
M4238-03	MH1T76	89	0.04 -	SFAM01.1	PPB	10/24/2021 15:57	SMPL
M4238-04	MH1T78	65	0.0344 -	SFAM01.1	PPB	10/24/2021 15:59	SMPL
M4238-05	MH1T80	58	0.0328 -	SFAM01.1	PPB	10/24/2021 16:02	SMPL
M4238-06	MH1T82	75	0.0368 -	SFAM01.1	PPB	10/24/2021 16:04	SMPL
M4238-07	MH1T84	72	0.0361 -	SFAM01.1	PPB	10/24/2021 16:06	SMPL

LB116994
INSTRUMENT ID : CV1

M4238-08	MH1T86	69	0.0354 -	SFAM01.1	PPB	10/24/2021 16:09	SMPL
M4238-09	MH1T88	65	0.0344 -	SFAM01.1	PPB	10/24/2021 16:11	SMPL
M4238-10	MH1T90	75	0.0368 -	SFAM01.1	PPB	10/24/2021 16:13	SMPL
CCV042	CCV042	19534	4.5834 -	SFAM01.1	PPB	10/24/2021 16:18	SMPL
CCB042	CCB042	40	0.0286 -	SFAM01.1	PPB	10/24/2021 16:20	SMPL
M4238-11	MH1T92	48	0.0305 -	SFAM01.1	PPB	10/24/2021 16:23	SMPL
M4238-12	MH1T94	54	0.0319 -	SFAM01.1	PPB	10/24/2021 16:25	SMPL
M4238-13	MH1T96	61	0.0335 -	SFAM01.1	PPB	10/24/2021 16:27	SMPL
M4238-14	MH1T98	64	0.0342 -	SFAM01.1	PPB	10/24/2021 16:29	SMPL
M4238-15	MH1U01	71	0.0358 -	SFAM01.1	PPB	10/24/2021 16:32	SMPL
M4238-16	MH1U03	61	0.0335 -	SFAM01.1	PPB	10/24/2021 16:34	SMPL
M4238-17	MH1U05	94	0.0412 -	SFAM01.1	PPB	10/24/2021 16:36	SMPL
M4238-18	MH1U07	72	0.0361 -	SFAM01.1	PPB	10/24/2021 16:39	SMPL
M4238-19	MH1U09	72	0.0361 -	SFAM01.1	PPB	10/24/2021 16:41	SMPL
M4238-20	MH1U11	89	0.04 -	SFAM01.1	PPB	10/24/2021 16:43	SMPL
M4238-21	MH1U11D	88	0.0398 -	SFAM01.1	PPB	10/24/2021 16:45	SMPL
M4238-22	MH1U11S	4199	1.0004 -	SFAM01.1	PPB	10/24/2021 16:51	SMPL
CCV043	CCV043	22126	5.1891 -	SFAM01.1	PPB	10/24/2021 16:54	SMPL
CCB043	CCB043	67	0.0349 -	SFAM01.1	PPB	10/24/2021 16:56	SMPL