

LB116997


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

37247



A= 0.0000e+000

B= 2.6583e-004

C= -3.3057e-002

Rho= 0.9996325

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.018	0.018	192	0.000	192					-
0.2	0.200	0.199	-0.001	874	0.0 %	874					-1
2.5	2.500	2.412	-0.088	9196	0.0 %	9196					-4
5.0	5.000	5.018	0.018	18999	0.0 %	18999					0
7.5	7.500	7.685	0.185	29034	0.0 %	29034					2
10.0	10.000	9.868	-0.132	37247	0.0 %	37247					-1

LB116997 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 25 Oct 2021 09:41:20				Date	Type	Type
Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units				
0.0 - 1	50	192	-	0.0000	7470A	PPB	25 Oct 2021	10:02:01	S	Std
0.2 - 1	50.2	874	-	0.2000	7470A	PPB	25 Oct 2021	10:04:18	S	Std
2.5 - 1	52.1	9196	-	2.5000	7470A	PPB	25 Oct 2021	10:06:36	S	Std
5.0 - 1	55	18999	-	5.0000	7470A	PPB	25 Oct 2021	10:08:53	S	Std
7.5 - 1	57.5	29034	-	7.5000	7470A	PPB	25 Oct 2021	10:11:09	S	Std
10.0 - 1	510	37247	-	10.0000	7470A	PPB	25 Oct 2021	10:18:34	S	Std
ICV56 - 1	ICV56	14787	3.8978	-	7470A	PPB	25 Oct 2021	10:21:18	U	SMPL
ICB56 - 1	ICB56	74	-0.0134	-	7470A	PPB	25 Oct 2021	10:23:33	U	SMPL
CCV44 - 1	CCV44	17977	4.7458	-	7470A	PPB	25 Oct 2021	10:25:48	U	SMPL
CCB44 - 1	CCB44	72	-0.0139	-	7470A	PPB	25 Oct 2021	10:28:03	U	SMPL
CRA - 1	CRA	866	0.1972	-	7470A	PPB	25 Oct 2021	10:30:18	U	SMPL
HighStd - 1	HighStd	36716	9.7273	-	7470A	PPB	25 Oct 2021	10:32:33	U	SMPL
ChkStd - 1	ChkStd	25413	6.7226	-	7470A	PPB	25 Oct 2021	10:37:23	U	SMPL
PB140113BL - 1	PBW	157	0.0087	-	7470A	PPB	25 Oct 2021	10:42:50	U	SMPL
PB140113BS - 1	LCSW	14663	3.8649	-	7470A	PPB	25 Oct 2021	10:45:05	U	SMPL
M4305-01 - 1	GW-2(6-9)	49	-0.0200	-	7470A	PPB	25 Oct 2021	10:47:22	U	SMPL
M4305-02 - 1	GW-2(14-17)	117	-0.0020	-	7470A	PPB	25 Oct 2021	10:49:38	U	SMPL
M4305-03 - 1	GW-2(24-27)	850	0.1929	-	7470A	PPB	25 Oct 2021	10:56:43	U	SMPL
M4305-03DUP - 1	GW-2(24-27)DUP	853	0.1937	-	7470A	PPB	25 Oct 2021	10:58:59	U	SMPL
M4305-03MS - 1	GW-2(24-27)MS	14272	3.7609	-	7470A	PPB	25 Oct 2021	11:01:15	U	SMPL
CCV45 - 1	CCV45	17463	4.6092	-	7470A	PPB	25 Oct 2021	11:03:30	U	SMPL
CCB45 - 1	CCB45	72	-0.0139	-	7470A	PPB	25 Oct 2021	11:05:45	U	SMPL
M4305-03MSD - 1	GW-2(24-27)MSD	14715	3.8787	-	7470A	PPB	25 Oct 2021	11:08:00	U	SMPL
M4313-07 - 1	1-3-TOTES	667	0.1443	-	7470A	PPB	25 Oct 2021	11:10:16	U	SMPL
PB140170BL - 1	PBW	167	0.0113	-	7470A	PPB	25 Oct 2021	11:12:31	U	SMPL
PB140170BS - 1	LCSW	14280	3.7631	-	7470A	PPB	25 Oct 2021	11:14:47	U	SMPL
M4306-02 - 1	MH-PIT-SOIL	121	-0.0009	-	7470A	PPB	25 Oct 2021	11:17:03	U	SMPL
M4312-01 - 1	5262	199	0.0198	-	7470A	PPB	25 Oct 2021	11:19:19	U	SMPL
M4312-02 - 1	5255	203	0.0209	-	7470A	PPB	25 Oct 2021	11:21:35	U	SMPL
M4312-03 - 1	5266	206	0.0217	-	7470A	PPB	25 Oct 2021	11:23:52	U	SMPL
M4315-02 - 1	MH-D-S	226	0.0270	-	7470A	PPB	25 Oct 2021	11:26:09	U	SMPL
M4315-05 - 1	MH-D-D	237	0.0299	-	7470A	PPB	25 Oct 2021	11:28:24	U	SMPL
CCV46 - 1	CCV46	17507	4.6209	-	7470A	PPB	25 Oct 2021	11:30:40	U	SMPL
CCB46 - 1	CCB46	92	-0.0086	-	7470A	PPB	25 Oct 2021	11:32:55	U	SMPL
M4315-08 - 1	MH-DD-S	215	0.0241	-	7470A	PPB	25 Oct 2021	11:35:11	U	SMPL
M4315-11 - 1	MH-DD-D	212	0.0233	-	7470A	PPB	25 Oct 2021	11:37:26	U	SMPL
M4316-02 - 1	WC-2-IS	297	0.0459	-	7470A	PPB	25 Oct 2021	11:39:41	U	SMPL
M4316-05 - 1	WC-2A-ID	236	0.0297	-	7470A	PPB	25 Oct 2021	11:41:57	U	SMPL
M4316-05DUP - 1	WC-2A-IDDUP	237	0.0299	-	7470A	PPB	25 Oct 2021	11:44:12	U	SMPL
M4316-05MS - 1	WC-2A-IDMS	15449	4.0738	-	7470A	PPB	25 Oct 2021	11:51:33	U	SMPL
M4316-05MSD - 1	WC-2A-IDMSD	15271	4.0265	-	7470A	PPB	25 Oct 2021	11:53:49	U	SMPL
PB140171BL - 1	PBW	120	-0.0012	-	7470A	PPB	25 Oct 2021	11:56:06	U	SMPL
PB140171BS - 1	LCSW	14665	3.8654	-	7470A	PPB	25 Oct 2021	11:58:23	U	SMPL
M4299-06 - 1	1694-1696	399	0.0730	-	7470A	PPB	25 Oct 2021	12:00:39	U	SMPL
CCV47 - 1	CCV47	17557	4.6342	-	7470A	PPB	25 Oct 2021	12:02:54	U	SMPL
CCB47 - 1	CCB47	69	-0.0147	-	7470A	PPB	25 Oct 2021	12:05:10	U	SMPL
M4312-04 - 1	5261	456	0.0882	-	7470A	PPB	25 Oct 2021	12:07:25	U	SMPL
M4312-05 - 1	comp	286	0.0430	-	7470A	PPB	25 Oct 2021	12:09:40	U	SMPL
M4312-06 - 1	5234	418	0.0781	-	7470A	PPB	25 Oct 2021	12:11:56	U	SMPL
M4317-01 - 1	MIDTECH-C-OUT-SLUDGE	194	0.0185	-	7470A	PPB	25 Oct 2021	12:14:11	U	SMPL
M4317-01DUP - 1	MIDTECH-C-OUT-SLUDGEDUP	241	0.0310	-	7470A	PPB	25 Oct 2021	12:19:12	U	SMPL
M4317-01MS - 1	MIDTECH-C-OUT-SLUDGEMS	14462	3.8114	-	7470A	PPB	25 Oct 2021	12:21:27	U	SMPL
M4317-01MSD - 1	MIDTECH-C-OUT-SLUDGEMSD	15131	3.9893	-	7470A	PPB	25 Oct 2021	12:23:44	U	SMPL
PB140170TB - 1	PB140119TB	143	0.0050	-	7470A	PPB	25 Oct 2021	12:26:02	U	SMPL
PB140171TB - 1	PB140121TB	400	0.0733	-	7470A	PPB	25 Oct 2021	12:28:20	U	SMPL
CCV48 - 1	CCV48	18403	4.8591	-	7470A	PPB	25 Oct 2021	12:30:36	U	SMPL
CCB48 - 1	CCB48	95	-0.0078	-	7470A	PPB	25 Oct 2021	12:32:52	U	SMPL
M4305-03LX5 - 1		208	0.0222	-	7470A	PPB	25 Oct 2021	12:35:07	U	SMPL
M4305-03A - 1		14511	3.8245	-	7470A	PPB	25 Oct 2021	12:37:23	U	SMPL
M4316-05LX5 - 1		17	-0.0285	-	7470A	PPB	25 Oct 2021	12:39:39	U	SMPL
M4316-05A - 1		16650	4.3931	-	7470A	PPB	25 Oct 2021	12:41:54	U	SMPL
M4317-01LX5 - 1		22	-0.0272	-	7470A	PPB	25 Oct 2021	12:44:09	U	SMPL
M4317-01A - 1		15085	3.9771	-	7470A	PPB	25 Oct 2021	12:46:25	U	SMPL
CCV49 - 1	CCV49	18103	4.7793	-	7470A	PPB	25 Oct 2021	12:48:41	U	SMPL
CCB49 - 1	CCB49	74	-0.0134	-	7470A	PPB	25 Oct 2021	12:50:58	U	SMPL