



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 10/26/2021

OVEN IN TEMP1: 104 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 10/25/2021 16:30 DATE & TIME OUT1: 10/26/2021 08:00

OVEN IN TEMP2: 103 °C

OVEN OUT TEMP2: 103 °C

DATE & TIME IN2: 10/26/2021 08:30 DATE & TIME OUT2: 10/26/2021 10:00

Run Number: LB117000

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB117000BL	LB117000BL	75.0367	75.0367	100	75.0368	75.0368	75.0368	0.0001	1
2	M4330-03	JET-GROUT-SPOILS	90.0491	90.0491	100	90.1198	90.1198	90.1198	0.0707	707
3	M4330-03DUP	JET-GROUT-SPOILSDUP	90.0089	90.0089	100	90.0766	90.0766	90.0766	0.0677	677

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

Result mg/L = ((C - B) / A) * 1000 * 1000

117000

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ASTM TS M4330 WorkList ID : 153626 Department : Wet-Chemistry Date : 10-25-2021 10:47:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/29/2021	Solid	M4330-03	ASTM TS	Cool 4 deg C	TULL02	D11	JET-GROUT-SPOILS	10/21/2021	SM2540 B

Date/Time 10/25/21 16:00
Raw Sample Received by: JO 1WC
Raw Sample Relinquished by: JO 1WC

Date/Time 10/25/21 17:00
Raw Sample Received by: JO 1WC
Raw Sample Relinquished by: JO 1WC