



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117007

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	10/25/2021	13:55
2	M4326-04		1	10	100.00	>212.0	>212.0	10/25/2021	14:25
3	M4326-04DUP		1	10	100.00	>212.0	>212.0	10/25/2021	14:50
4	M4326-05		1	12	100.00	>212.0	>212.0	10/25/2021	15:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 117007

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp102421 WorkList ID : 153615 Department : Wet-Chemistry Date : 10-24-2021 14:57:34

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/29/2021	Water	M4326-04	Flash Point	Cool 4 deg C	PSEG03	K51	17066	10/22/2021	1010B
10/29/2021	Water	M4326-05	Flash Point	Cool 4 deg C	PSEG03	K51	17060	10/22/2021	1010B

Date/Time 10/25/21 13:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: SP (WC)

Date/Time 10/25/21 15:20
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)