


LB117013

7471B

INSTRUMENT ID: CV3

 μ Abs.:

48930



Linear

A= 0.0000e+000

B= 2.0548e-004

C= -3.9127e-002

Rho= 0.9999493

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.002	0.002	200	0.000	200					-19
0.2	0.200	0.163	-0.037	984	0.0 %	984					2
2.5	2.500	2.541	0.041	12557	0.0 %	12557					1
5.0	5.000	5.038	0.038	24709	0.0 %	24709					-1
7.5	7.500	7.440	-0.060	36400	0.0 %	36400					0
10.0	10.000	10.015	0.015	48930	0.0 %	48930					

LB117013 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 25 Oct 2021 13:56:48

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	200	-	0.0000	7471B	PPB	25 Oct 2021 13:59:44	S	Std
0.2 - 1	50-2	984	-	0.2000	7471B	PPB	25 Oct 2021 14:02:01	S	Std
2.5 - 1	52-5	12557	-	2.5000	7471B	PPB	25 Oct 2021 14:04:18	S	Std
5.0 - 1	55	24709	-	5.0000	7471B	PPB	25 Oct 2021 14:06:35	S	Std
7.5 - 1	57-1	36400	-	7.5000	7471B	PPB	25 Oct 2021 14:08:51	S	Std
10.0 - 1	510	48930	-	10.0000	7471B	PPB	25 Oct 2021 14:13:45	S	Std
ICV57 - 1	ICV57	20405	4.1538	-	7471B	PPB	25 Oct 2021 14:16:30	U	SMPL
ICB57 - 1	ICB57	12	-0.0367	-	7471B	PPB	25 Oct 2021 14:18:45	U	SMPL
CCV50 - 1	CCV50	25548	5.2106	-	7471B	PPB	25 Oct 2021 14:21:00	U	SMPL
CCB50 - 1	CCB50	10	-0.0371	-	7471B	PPB	25 Oct 2021 14:23:15	U	SMPL
CRA - 1	CRA	1250	0.2177	-	7471B	PPB	25 Oct 2021 14:25:30	U	SMPL
HighStd - 1	HighStd	49145	10.0594	-	7471B	PPB	25 Oct 2021 14:27:45	U	SMPL
ChkStd - 1	ChkStd	34376	7.0246	-	7471B	PPB	25 Oct 2021 14:30:00	U	SMPL
PB140199BL - 1	PB140199BL	15	-0.0360	-	7471B	PPB	25 Oct 2021 14:35:24	U	SMPL
PB140199BS - 1	PB140199BS	20745	4.2236	-	7471B	PPB	25 Oct 2021 14:37:39	U	SMPL
M4315-01 - 1	MH-D-S	88	-0.0210	-	7471B	PPB	25 Oct 2021 14:39:56	U	SMPL
M4315-01DUP - 1	MH-D-SDUP	367	0.0363	-	7471B	PPB	25 Oct 2021 14:42:13	U	SMPL
M4315-01MS - 1	MH-D-SMS	20278	4.1277	-	7471B	PPB	25 Oct 2021 14:44:29	U	SMPL
M4315-01MSD - 1	MH-D-SMSD	20032	4.0771	-	7471B	PPB	25 Oct 2021 14:46:47	U	SMPL
M4315-04 - 1	MH-D-D	1651	0.3001	-	7471B	PPB	25 Oct 2021 14:49:03	U	SMPL
CCV51 - 1	CCV51	25058	5.1099	-	7471B	PPB	25 Oct 2021 14:51:18	U	SMPL
CCB51 - 1	CCB51	3	-0.0385	-	7471B	PPB	25 Oct 2021 14:53:33	U	SMPL
M4315-07 - 1	MH-DD-S	728	0.1105	-	7471B	PPB	25 Oct 2021 14:55:49	U	SMPL
M4315-10 - 1	MH-DD-D	396	0.0422	-	7471B	PPB	25 Oct 2021 14:58:04	U	SMPL
M4316-01 - 1	WC-2-IS	27868	5.6873	-	7471B	PPB	25 Oct 2021 15:00:20	U	SMPL
M4316-04 - 1	WC-2A-ID	5074	1.0035	-	7471B	PPB	25 Oct 2021 15:02:35	U	SMPL
M4324-01 - 1	FK-SF-02-007-A	7603	1.5232	-	7471B	PPB	25 Oct 2021 15:04:52	U	SMPL
M4324-03 - 1	FK-SF-02-007-B	7157	1.4315	-	7471B	PPB	25 Oct 2021 15:07:09	U	SMPL
M4325-01 - 1	RED-BOX-1-225057	152	-0.0079	-	7471B	PPB	25 Oct 2021 15:09:24	U	SMPL
M4325-03 - 1	RED-BOX-2-1934	186	-0.0009	-	7471B	PPB	25 Oct 2021 15:11:41	U	SMPL
M4329-01 - 1	PSC124019	4011	0.7851	-	7471B	PPB	25 Oct 2021 15:13:58	U	SMPL
M4329-03 - 1	PSC124042	2655	0.5064	-	7471B	PPB	25 Oct 2021 15:16:14	U	SMPL
CCV52 - 1	CCV52	25257	5.1508	-	7471B	PPB	25 Oct 2021 15:18:29	U	SMPL
CCB52 - 1	CCB52	14	-0.0362	-	7471B	PPB	25 Oct 2021 15:20:45	U	SMPL
M4332-02 - 1	BB-SAMPLE-1	2798	0.5358	-	7471B	PPB	25 Oct 2021 15:23:00	U	SMPL
M4332-06 - 1	BB-SAMPLE-2	2602	0.4955	-	7471B	PPB	25 Oct 2021 15:25:16	U	SMPL
M4332-10 - 1	BB-SAMPLE-3	17892	3.6374	-	7471B	PPB	25 Oct 2021 15:27:31	U	SMPL
M4332-14 - 1	BB-SAMPLE-4	1862	0.3435	-	7471B	PPB	25 Oct 2021 15:29:47	U	SMPL
PB140200BL - 1	PB140200BL	10	-0.0371	-	7471B	PPB	25 Oct 2021 15:32:03	U	SMPL
PB140200BS - 1	PB140200BS	16963	3.4465	-	7471B	PPB	25 Oct 2021 15:34:20	U	SMPL
M4306-01 - 1	MH-PIT-SOIL	3213	0.6211	-	7471B	PPB	25 Oct 2021 15:36:36	U	SMPL
M4306-01DUP - 1	MH-PIT-SOILDUP	3614	0.7035	-	7471B	PPB	25 Oct 2021 15:38:53	U	SMPL
M4306-01MS - 1	MH-PIT-SOILMS	21611	4.4016	-	7471B	PPB	25 Oct 2021 15:41:09	U	SMPL
M4306-01MSD - 1	MH-PIT-SOILMSD	24088	4.9106	-	7471B	PPB	25 Oct 2021 15:43:25	U	SMPL
CCV53 - 1	CCV53	25146	5.1280	-	7471B	PPB	25 Oct 2021 15:45:41	U	SMPL
CCB53 - 1	CCB53	-11	-0.0414	-	7471B	PPB	25 Oct 2021 15:47:56	U	SMPL
M4308-01 - 1	COMP-1	15771	3.2016	-	7471B	PPB	25 Oct 2021 15:50:12	U	SMPL
M4308-02 - 1	COMP-2	14913	3.0253	-	7471B	PPB	25 Oct 2021 15:52:27	U	SMPL
M4308-03 - 1	COMP-3	23351	4.7591	-	7471B	PPB	25 Oct 2021 15:54:43	U	SMPL
M4309-01 - 1	L-1	2336	0.4409	-	7471B	PPB	25 Oct 2021 15:56:58	U	SMPL
M4311-01 - 1	OR-2-102121	2744	0.5247	-	7471B	PPB	25 Oct 2021 15:59:14	U	SMPL
M4313-01 - 1	124027	934	0.1528	-	7471B	PPB	25 Oct 2021 16:01:31	U	SMPL
M4313-03 - 1	3990	6552	1.3072	-	7471B	PPB	25 Oct 2021 16:03:47	U	SMPL
M4313-05 - 1	4743	1089	0.1846	-	7471B	PPB	25 Oct 2021 16:06:04	U	SMPL
CCV54 - 1	CCV54	24361	4.9667	-	7471B	PPB	25 Oct 2021 16:08:22	U	SMPL
CCB54 - 1	CCB54	109	-0.0167	-	7471B	PPB	25 Oct 2021 16:10:38	U	SMPL
M4315-01LX5 - 1		37	-0.0315	-	7471B	PPB	25 Oct 2021 16:12:54	U	SMPL
M4315-01A - 1		20192	4.1100	-	7471B	PPB	25 Oct 2021 16:15:09	U	SMPL
M4306-01LX5 - 1		552	0.0743	-	7471B	PPB	25 Oct 2021 16:17:25	U	SMPL
M4306-01A - 1		22523	4.5890	-	7471B	PPB	25 Oct 2021 16:19:40	U	SMPL
CCV55 - 1	CCV55	25308	5.1613	-	7471B	PPB	25 Oct 2021 16:21:56	U	SMPL
CCB55 - 1	CCB55	-3	-0.0397	-	7471B	PPB	25 Oct 2021 16:24:12	U	SMPL