



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/26/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 10/25/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/26/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117015

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M4089-06	20-SB-2-20211005-18.5-19.0	1	1.14	9.86	8.2	81.0	
M4089-08	20-SB-1-20211005-7.5-8.0	2	1.15	9.98	8.82	86.9	
M4336-02	C1	3	1.12	9.58	8.09	82.4	
M4336-03	C2	4	1.16	9.97	8.2	79.9	
M4336-04	C3	5	1.13	9.59	8.58	88.1	
M4336-05	C4	6	1.16	9.81	8.25	82.0	
M4336-06	G1	7	1.13	9.58	8.37	85.7	
M4336-07	G2	8	1.16	9.79	8.82	88.8	
M4336-08	G3	9	1.16	9.58	8.51	87.3	
M4336-09	G4	10	1.16	9.69	8.31	83.8	
M4340-01	TH-ES	11	1.16	9.62	8.57	87.6	
M4340-03	TH-ES-EPH-2	12	1.11	9.71	8.89	90.5	
M4340-04	TH-ED	13	1.16	9.92	8.02	78.3	
M4340-06	TH-ED-EPH-2	14	1.15	9.96	7.92	76.8	
M4340-07	TH-EES	15	1.16	9.95	9.00	89.2	
M4340-09	TH-EES-EPH-2	16	1.11	9.78	8.93	90.2	
M4340-10	TH-EED	17	1.16	9.98	8.97	88.5	
M4340-12	TH-EED-EPH-2	18	1.16	9.51	8.13	83.5	
M4340-12D	TH-EED-EPH-2DUP	19	1.18	9.53	8.15	83.5	
M4341-01	WASTE	20	1.13	9.98	8.62	84.6	
M4342-01	CLI-2021-1-A	51	1.00	2.00	2.00	100.0	PILC REEL
M4342-02	CLI-2021-2-A	52	1.00	2.00	2.00	100.0	PILC REEL
M4342-03	CLI-2021-2-B	53	1.00	2.00	2.00	100.0	PILC REEL
M4342-04	CLI-2021-3-A	54	1.00	2.00	2.00	100.0	PILC REEL
M4342-05	CLI-2021-3-B	55	1.00	2.00	2.00	100.0	PILC REEL
M4342-06	CLI-2021-4-A	56	1.00	2.00	2.00	100.0	PILC REEL
M4342-07	CLI-2021-4-B	57	1.00	2.00	2.00	100.0	PILC REEL



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/26/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 10/25/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/26/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117015

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4342-08	CLI-2021-5-A	58	1.00	2.00	2.00	100.0	PILC REEL
M4342-09	CLI-2021-6-A	59	1.00	2.00	2.00	100.0	PILC REEL
M4342-10	CLI-2021-7-A	60	1.00	2.00	2.00	100.0	PILC REEL
M4343-01	METRO-SP	61	1.13	9.59	8.82	90.9	
M4343-02	METRO-SP-EPH-2	62	1.13	9.67	8.91	91.1	
M4346-01	JEL-354B-(1-1)	21	1.00	2.00	2.00	100.0	PILC REEL
M4346-02	JEL-354B-(1-2)	22	1.00	2.00	2.00	100.0	PILC REEL
M4346-03	JEL-354B-(2-1)	23	1.00	2.00	2.00	100.0	PILC REEL
M4346-04	JEL-354B-(2-2)	24	1.00	2.00	2.00	100.0	PILC REEL
M4346-05	BC-260596-(1-1)	25	1.00	2.00	2.00	100.0	PILC REEL
M4346-06	BC-260596-(1-2)	26	1.00	2.00	2.00	100.0	PILC REEL
M4346-07	SEC-2021-1A	27	1.00	2.00	2.00	100.0	PILC REEL
M4346-08	SEC-2021-1B	28	1.00	2.00	2.00	100.0	PILC REEL
M4346-09	SEC-2021-2A	29	1.00	2.00	2.00	100.0	PILC REEL
M4346-10	SEC-2021-2B	30	1.00	2.00	2.00	100.0	PILC REEL
M4346-11	SEC-2021-3A	31	1.00	2.00	2.00	100.0	PILC REEL
M4346-12	SEC-2021-3B	32	1.00	2.00	2.00	100.0	PILC REEL
M4346-13	SEC-2021-4A	33	1.00	2.00	2.00	100.0	PILC REEL
M4346-14	SEC-2021-4B	34	1.00	2.00	2.00	100.0	PILC REEL
M4346-15	SEC-2021-5A	35	1.00	2.00	2.00	100.0	PILC REEL
M4346-16	SEC-2021-5B	36	1.00	2.00	2.00	100.0	PILC REEL
M4346-17	SEC-2021-6A	37	1.00	2.00	2.00	100.0	PILC REEL
M4346-18	SEC-2021-6B	38	1.00	2.00	2.00	100.0	PILC REEL
M4346-19	SEC-2021-7A	39	1.00	2.00	2.00	100.0	PILC REEL
M4346-20	SEC-2021-7B	40	1.00	2.00	2.00	100.0	PILC REEL
M4346-21	SEC-2021-8A	41	1.00	2.00	2.00	100.0	PILC REEL
M4346-22	SEC-2021-8B	42	1.00	2.00	2.00	100.0	PILC REEL
M4346-23	SEC-2021-9A	43	1.00	2.00	2.00	100.0	PILC REEL



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/26/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 10/25/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/26/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117015

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4346-24	SEC-2021-9B	44	1.00	2.00	2.00	100.0	PILC REEL
M4348-01	354	63	1.15	9.96	7.02	66.6	GEL MATRIX
M4348-02	354	64	1.16	9.76	6.33	60.1	GEL MATRIX
M4348-03	358	65	1.12	9.52	6.07	58.9	GEL MATRIX
M4348-04	358	66	1.19	9.94	6.36	59.1	GEL MATRIX
M4348-05	VNJ-229	67	1.14	9.91	6.39	59.9	GEL MATRIX
M4348-06	VNJ-229	68	1.13	9.62	6.54	63.7	GEL MATRIX
M4348-07	287	69	1.16	9.77	5.75	53.3	GEL MATRIX
M4348-08	287	70	1.18	9.96	6.21	57.3	GEL MATRIX
M4348-09	370	71	1.15	9.94	5.69	51.6	GEL MATRIX
M4348-10	370	72	1.13	9.58	5.2	48.2	GEL MATRIX
M4349-01	ASPHALT-GRAB	45	1.14	9.96	9.94	99.8	ASPHALT SAMPLE
M4350-01	CONCRETE-GRAB	46	1.15	9.69	9.63	99.3	CONCRETE
M4351-01	ASPHALT-GRAB	47	1.16	9.96	9.93	99.7	ASPHALT SAMPLE
M4352-01	CONCRETE-GRAB	48	1.12	9.79	9.7	99.0	CONCRETE
M4353-01	CONCRETE-GRAB-II	49	1.12	9.59	9.52	99.2	CONCRETE
M4354-01	ASPHALT-GRAB	50	1.15	9.95	9.93	99.8	ASPHALT SAMPLE
M4355-01	OK-2-102521	73	1.19	9.6	8.9	91.7	
M4355-02	OK-2-102521-EPH-2	74	1.11	9.58	8.94	92.4	
M4356-01	ETGI-282	75	1.15	9.95	9.19	91.4	
M4356-02	ETGI-282-EPH-2	76	1.15	9.95	9.19	91.4	
M4356-03	VNJ-221	77	1.14	9.98	9.13	90.4	
M4356-04	VNJ-221-EPH-2	78	1.14	9.98	9.13	90.4	
M4357-01	CL-01-102521	79	1.11	9.66	9.07	93.1	
M4357-02	CL-01-102521-EPH-2	80	1.12	9.52	9.02	94.0	
M4358-01	TH-FS	81	1.14	9.94	8.22	80.5	
M4358-03	TH-FS-EPH-2	82	1.19	9.97	8.69	85.4	
M4358-04	TH-FD	83	1.13	9.6	7.97	80.8	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/26/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 10/25/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 10/26/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117015

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4358-06	TH-FD-EPH-2	84	1.13	9.93	8.92	88.5	
M4358-07	TH-GS	85	1.17	9.96	8.41	82.4	
M4358-09	TH-GS-EPH-2	86	1.12	9.6	8.54	87.5	
M4358-10	TH-GD	87	1.19	9.84	8.00	78.7	
M4358-12	TH-GD-EPH-2	88	1.1	9.64	7.9	79.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

117015

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-102521 WorkList ID : 153616 Department : Wet-Chemistry Date : 10-25-2021 08:31:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/15/2021	Solid	M4089-06	Percent Solids	Cool 4 deg C	PARS02	J61	20-SB-2-20211005-18.5-1	10/05/2021	Chemtech -SO
10/15/2021	Solid	M4089-08	Percent Solids	Cool 4 deg C	PARS02	J61	20-SB-1-20211005-7.5-8.1	10/05/2021	Chemtech -SO
10/29/2021	Solid	M4336-02	Percent Solids	Cool 4 deg C	DVIR01	K61	C1	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4336-03	Percent Solids	Cool 4 deg C	DVIR01	K61	C2	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4336-04	Percent Solids	Cool 4 deg C	DVIR01	K61	C3	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4336-05	Percent Solids	Cool 4 deg C	DVIR01	K61	C4	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4336-06	Percent Solids	Cool 4 deg C	DVIR01	K61	G1	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4336-07	Percent Solids	Cool 4 deg C	DVIR01	K61	G2	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4336-08	Percent Solids	Cool 4 deg C	DVIR01	K61	G3	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4336-09	Percent Solids	Cool 4 deg C	DVIR01	K61	G4	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4340-01	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-ES	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4340-03	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-ES-EPH-2	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4340-04	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-ED	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4340-06	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-ED-EPH-2	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4340-07	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-EES	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4340-09	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-EES-EPH-2	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4340-10	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-EED	10/22/2021	Chemtech -SO
10/29/2021	Solid	M4340-12	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-EED-EPH-2	10/22/2021	Chemtech -SO
10/31/2021	Solid	M4341-01	Percent Solids	Cool 4 deg C	RMJE02	D11	WASTE	10/21/2021	Chemtech -SO
11/01/2021	Solid	M4342-01	Percent Solids	Cool 4 deg C	PSEG03	K61	CLI-2021-1-A	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4342-02	Percent Solids	Cool 4 deg C	PSEG03	K61	CLI-2021-2-A	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4342-03	Percent Solids	Cool 4 deg C	PSEG03	K61	CLI-2021-2-B	10/25/2021	Chemtech -SO

Date/Time	10/25/21	17:20
Raw Sample Received by:	CP SM	
Raw Sample Relinquished by:	JD WCC	

WORKLIST(Hardcopy Internal Chain)

117015

WorkList Name : %1-102521

WorkList ID : 153616

Department : Wet-Chemistry

Date : 10-25-2021 08:31:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/01/2021	Solid	M4342-04	Percent Solids	Cool 4 deg C	PSEG03	K61	CLI-2021-3-A	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4342-05	Percent Solids	Cool 4 deg C	PSEG03	K61	CLI-2021-3-B	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4342-06	Percent Solids	Cool 4 deg C	PSEG03	K61	CLI-2021-4-A	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4342-07	Percent Solids	Cool 4 deg C	PSEG03	K61	CLI-2021-4-B	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4342-08	Percent Solids	Cool 4 deg C	PSEG03	K61	CLI-2021-5-A	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4342-09	Percent Solids	Cool 4 deg C	PSEG03	K61	CLI-2021-6-A	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4342-10	Percent Solids	Cool 4 deg C	PSEG03	K61	CLI-2021-7-A	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4343-01	Percent Solids	Cool 4 deg C	PSEG03	K11	METRO-SP	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4343-02	Percent Solids	Cool 4 deg C	PSEG03	K11	METRO-SP-EPH-2	10/25/2021	Chemtech -SO
10/30/2021	Solid	M4346-01	Percent Solids	Cool 4 deg C	PSEG03	K11	JEL-354B-(1-1)	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-02	Percent Solids	Cool 4 deg C	PSEG03	K11	JEL-354B-(1-2)	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-03	Percent Solids	Cool 4 deg C	PSEG03	K11	JEL-354B-(2-1)	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-04	Percent Solids	Cool 4 deg C	PSEG03	K11	JEL-354B-(2-2)	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-05	Percent Solids	Cool 4 deg C	PSEG03	K11	BC-260596-(1-1)	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-06	Percent Solids	Cool 4 deg C	PSEG03	K11	BC-260596-(1-2)	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-07	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-1A	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-08	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-1B	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-09	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-2A	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-10	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-2B	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-11	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-3A	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-12	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-3B	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-13	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-4A	10/20/2021	Chemtech -SO

Date/Time 10/25/21 15:50
 Raw Sample Received by: JO (WOC)
 Raw Sample Relinquished by: CP SM

Date/Time 10/25/21 17:20
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JO (WOC)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-102521 WorkList ID : 153616 Department : Wet-Chemistry Date : 10-25-2021 08:31:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/30/2021	Solid	M4346-14	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-4B	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-15	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-5A	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-16	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-5B	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-17	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-6A	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-18	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-6B	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-19	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-7A	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-20	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-7B	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-21	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-8A	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-22	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-8B	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-23	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-9A	10/20/2021	Chemtech -SO
10/30/2021	Solid	M4346-24	Percent Solids	Cool 4 deg C	PSEG03	K11	SEC-2021-9B	10/20/2021	Chemtech -SO
11/01/2021	Solid	M4348-01	Percent Solids	Cool 4 deg C	PSEG03	K11	354	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4348-02	Percent Solids	Cool 4 deg C	PSEG03	K11	354	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4348-03	Percent Solids	Cool 4 deg C	PSEG03	K11	358	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4348-04	Percent Solids	Cool 4 deg C	PSEG03	K11	358	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4348-05	Percent Solids	Cool 4 deg C	PSEG03	K11	VNJ-229	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4348-06	Percent Solids	Cool 4 deg C	PSEG03	K11	VNJ-229	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4348-07	Percent Solids	Cool 4 deg C	PSEG03	K11	287	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4348-08	Percent Solids	Cool 4 deg C	PSEG03	K11	287	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4348-09	Percent Solids	Cool 4 deg C	PSEG03	K11	370	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4348-10	Percent Solids	Cool 4 deg C	PSEG03	K11	370	10/25/2021	Chemtech -SO
10/21/2021	Solid	M4349-01	Percent Solids	Cool 4 deg C	EART12	D11	ASPHALT-GRAB	10/11/2021	Chemtech -SO

Date/Time 10/25/21 1540
Raw Sample Received by: JP (W)
Raw Sample Relinquished by: JP sm

Date/Time 10/25/21 17120
Raw Sample Received by: CP sm
Raw Sample Relinquished by: JP (W)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M4340 WorkList ID : 153633 Department : Wet-Chemistry Date : 10-25-2021 13:13:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/22/2021	Solid	M4340-02	Corrosivity	Cool 4 deg C	PSEG03	K11	TH-ES	10/22/2021	9045D
10/22/2021	Solid	M4340-05	Corrosivity	Cool 4 deg C	PSEG03	K11	TH-ED	10/22/2021	9045D
10/22/2021	Solid	M4340-08	Corrosivity	Cool 4 deg C	PSEG03	K11	TH-EES	10/22/2021	9045D
10/22/2021	Solid	M4340-11	Corrosivity	Cool 4 deg C	PSEG03	K11	TH-EED	10/22/2021	9045D
10/25/2021	Solid	M4358-02	Corrosivity	Cool 4 deg C	PSEG03	K11	TH-FS	10/25/2021	9045D
10/25/2021	Solid	M4358-05	Corrosivity	Cool 4 deg C	PSEG03	K11	TH-FD	10/25/2021	9045D
10/25/2021	Solid	M4358-08	Corrosivity	Cool 4 deg C	PSEG03	K11	TH-GS	10/25/2021	9045D
10/25/2021	Solid	M4358-11	Corrosivity	Cool 4 deg C	PSEG03	K11	TH-GD	10/25/2021	9045D

Date/Time 10/25/21 16:20
 Raw Sample Received by: Ja (WC)
 Raw Sample Relinquished by: CP SM

Date/Time 10/25/21 18:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JB (WC)

WORKLIST(Hardcopy Internal Chain)

W2117015

WorkList Name : %1-102521 WorkList ID : 153616 Department : Wet-Chemistry Date : 10-25-2021 08:31:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/21/2021	Solid	M4350-01	Percent Solids	Cool 4 deg C	EART12	D11	CONCRETE-GRAB	10/11/2021	Chemtech -SO
10/29/2021	Solid	M4351-01	Percent Solids	Cool 4 deg C	EART12	D11	ASPHALT-GRAB	10/19/2021	Chemtech -SO
10/29/2021	Solid	M4352-01	Percent Solids	Cool 4 deg C	EART12	D11	CONCRETE-GRAB	10/19/2021	Chemtech -SO
10/28/2021	Solid	M4353-01	Percent Solids	Cool 4 deg C	EART12	D11	CONCRETE-GRAB-II	10/18/2021	Chemtech -SO
10/23/2021	Solid	M4354-01	Percent Solids	Cool 4 deg C	EART12	D11	ASPHALT-GRAB	10/13/2021	Chemtech -SO
10/28/2021	Solid	M4355-01	Percent Solids	Cool 4 deg C	PSEG05	K11	OK-2-102521	10/25/2021	Chemtech -SO
10/28/2021	Solid	M4355-02	Percent Solids	Cool 4 deg C	PSEG05	K11	OK-2-102521-EPH-2	10/25/2021	Chemtech -SO
10/28/2021	Solid	M4356-01	Percent Solids	Cool 4 deg C	PSEG03	K11	ETGI-282	10/25/2021	Chemtech -SO
10/28/2021	Solid	M4356-02	Percent Solids	Cool 4 deg C	PSEG03	K11	ETGI-282-EPH-2	10/25/2021	Chemtech -SO
10/28/2021	Solid	M4356-03	Percent Solids	Cool 4 deg C	PSEG03	K11	VNJ-221	10/25/2021	Chemtech -SO
10/28/2021	Solid	M4356-04	Percent Solids	Cool 4 deg C	PSEG03	K11	VNJ-221-EPH-2	10/25/2021	Chemtech -SO
10/28/2021	Solid	M4357-01	Percent Solids	Cool 4 deg C	PSEG05	K11	CL-01-102521	10/25/2021	Chemtech -SO
10/28/2021	Solid	M4357-02	Percent Solids	Cool 4 deg C	PSEG05	K11	CL-01-102521-EPH-2	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4358-01	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-FS	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4358-03	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-FS-EPH-2	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4358-04	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-FD	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4358-06	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-FD-EPH-2	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4358-07	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-GS	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4358-09	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-GS-EPH-2	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4358-10	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-GD	10/25/2021	Chemtech -SO
11/01/2021	Solid	M4358-12	Percent Solids	Cool 4 deg C	PSEG03	K11	TH-GD-EPH-2	10/25/2021	Chemtech -SO

Date/Time 10/25/21 15:40
 Raw Sample Received by: JO CWC
 Raw Sample Relinquished by: CL SM

Date/Time 10/25/21 17:20
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JO CWC