



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117017

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.0 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.7	4.01	10/25/2021	15:15
2	CAL2	1	Water	NA	NA	20.7	7.00	10/25/2021	15:16
3	CAL3	1	Water	NA	NA	20.7	10.03	10/25/2021	15:18
4	ICV	1	Water	NA	NA	20.8	6.99	10/25/2021	15:20
5	CCV1	1	Water	NA	NA	20.7	2.01	10/25/2021	15:21
6	M4347-01	1	Water	NA	NA	21.1	6.69	10/25/2021	15:24
7	M4347-01DUP	1	Water	NA	NA	21.2	6.70	10/25/2021	15:25
8	CCV2	1	Water	NA	NA	20.8	12.03	10/25/2021	15:26

WORKLIST(Hardcopy Internal Chain)

LB 117017

WorkList Name : PH M4347 WorkList ID : 153637 Department : Wet-Chemistry Date : 10-25-2021 15:01:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/25/2021	Water	M4347-01	pH	Cool 4 deg C	PSEG03	K51	TOTE-2223	10/25/2021	9040C

Date/Time 10/25/21 15:10
 Raw Sample Received by: JP CWC
 Raw Sample Relinquished by: JP SM

Date/Time 10/25/21 16:15
 Raw Sample Received by: JP SM
 Raw Sample Relinquished by: JP CWC