


L3117021

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.007	0.007	31	0.000	31					1
0.2	0.200	0.204	0.004	745	0.0 %	745					2
2.5	2.500	2.504	0.004	9081	0.0 %	9081					0
5.0	5.000	5.083	0.083	18430	0.0 %	18430					2
7.5	7.500	7.299	-0.201	26465	0.0 %	26465					3
10.0	10.000	10.102	0.102	36627	0.0 %	36627					1

LB117021 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 26 Oct 2021 12:06:05					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	7470A	PPB	26 Oct 2021 12:16:38	S	Std
0.2 - 1	50.2	745	-	0.2000	7470A	PPB	26 Oct 2021 12:26:38	S	Std
2.5 - 1	52.5	9081	-	2.5000	7470A	PPB	26 Oct 2021 12:28:55	S	Std
5.0 - 1	55	18430	-	5.0000	7470A	PPB	26 Oct 2021 12:31:11	S	Std
7.5 - 1	57.5	26465	-	7.5000	7470A	PPB	26 Oct 2021 12:33:28	S	Std
10.0 - 1	510	36627	-	10.0000	7470A	PPB	26 Oct 2021 12:38:56	S	Std
ICV58 - 1	ICV58	15522	4.2806	-	7470A	PPB	26 Oct 2021 12:41:42	U	SMPL
ICB58 - 1	ICB58	-50	-0.0149	-	7470A	PPB	26 Oct 2021 12:43:57	U	SMPL
CCV56 - 1	CCV56	19007	5.2419	-	7470A	PPB	26 Oct 2021 12:46:11	U	SMPL
CCB56 - 1	CCB56	-53	-0.0158	-	7470A	PPB	26 Oct 2021 12:48:26	U	SMPL
CRA - 1	CRA	725	0.1989	-	7470A	PPB	26 Oct 2021 12:50:41	U	SMPL
HighStd - 1	HighStd	36672	10.1148	-	7470A	PPB	26 Oct 2021 12:52:56	U	SMPL
ChkStd - 1	ChkStd	25147	6.9356	-	7470A	PPB	26 Oct 2021 12:55:11	U	SMPL
PB140231BL - 1	PBW	-107	-0.0307	-	7470A	PPB	26 Oct 2021 12:57:28	U	SMPL
PB140231BS - 1	LCSW	15583	4.2974	-	7470A	PPB	26 Oct 2021 12:59:43	U	SMPL
M4325-02 - 1	RED-BOX-1-225057	-75	-0.0218	-	7470A	PPB	26 Oct 2021 13:02:00	U	SMPL
M4325-04 - 1	RED-BOX-2-1934	21	0.0047	-	7470A	PPB	26 Oct 2021 13:04:16	U	SMPL
M4326-01 - 1	GARAGE-DEBRIS	20	0.0044	-	7470A	PPB	26 Oct 2021 13:06:33	U	SMPL
M4326-02 - 1	XFMRSHP1	42	0.0105	-	7470A	PPB	26 Oct 2021 13:08:50	U	SMPL
M4326-03 - 1	XFMRSHP2	33	0.0080	-	7470A	PPB	26 Oct 2021 13:11:05	U	SMPL
CCV57 - 1	CCV57	18788	5.1815	-	7470A	PPB	26 Oct 2021 13:13:20	U	SMPL
CCB57 - 1	CCB57	-51	-0.0152	-	7470A	PPB	26 Oct 2021 13:15:35	U	SMPL
M4329-02 - 1	PSC124019	1	-0.0009	-	7470A	PPB	26 Oct 2021 13:17:50	U	SMPL
M4329-04 - 1	PSC124042	39	0.0096	-	7470A	PPB	26 Oct 2021 13:20:05	U	SMPL
M4330-02 - 1	JET-GROUT-SPOILS	43	0.0107	-	7470A	PPB	26 Oct 2021 13:22:21	U	SMPL
M4332-03 - 1	BB-SAMPLE-1	36	0.0088	-	7470A	PPB	26 Oct 2021 13:24:37	U	SMPL
M4332-07 - 1	BB-SAMPLE-2	41	0.0102	-	7470A	PPB	26 Oct 2021 13:26:53	U	SMPL
M4332-11 - 1	BB-SAMPLE-3	31	0.0074	-	7470A	PPB	26 Oct 2021 13:29:10	U	SMPL
M4332-15 - 1	BB-SAMPLE-4	36	0.0088	-	7470A	PPB	26 Oct 2021 13:31:26	U	SMPL
M4332-15DUP - 1	BB-SAMPLE-4DUP	42	0.0105	-	7470A	PPB	26 Oct 2021 13:33:43	U	SMPL
M4332-15MS - 1	BB-SAMPLE-4MS	14349	3.9570	-	7470A	PPB	26 Oct 2021 13:35:59	U	SMPL
M4332-15MSD - 1	BB-SAMPLE-4MSD	13720	3.7835	-	7470A	PPB	26 Oct 2021 13:38:15	U	SMPL
CCV58 - 1	CCV58	18853	5.1994	-	7470A	PPB	26 Oct 2021 13:40:30	U	SMPL
CCB58 - 1	CCB58	-52	-0.0155	-	7470A	PPB	26 Oct 2021 13:42:46	U	SMPL
PB140232BL - 1	PBW	8	0.0011	-	7470A	PPB	26 Oct 2021 13:45:02	U	SMPL
PB140232BS - 1	LCSW	14306	3.9452	-	7470A	PPB	26 Oct 2021 13:47:17	U	SMPL
M4326-04 - 1	17066	-51	-0.0152	-	7470A	PPB	26 Oct 2021 13:49:33	U	SMPL
M4326-04DUP - 1	17066DUP	6	0.0005	-	7470A	PPB	26 Oct 2021 13:51:48	U	SMPL
M4326-04MS - 1	17066MS	13742	3.7896	-	7470A	PPB	26 Oct 2021 13:54:04	U	SMPL
M4326-04MSD - 1	17066MSD	14246	3.9286	-	7470A	PPB	26 Oct 2021 13:56:20	U	SMPL
M4326-05 - 1	17060	285	0.0775	-	7470A	PPB	26 Oct 2021 13:58:36	U	SMPL
PB140231TB - 1	PB140150TB	-45	-0.0135	-	7470A	PPB	26 Oct 2021 14:00:53	U	SMPL
PB140232TB - 1	PB140155TB	1	-0.0009	-	7470A	PPB	26 Oct 2021 14:03:10	U	SMPL
CCV59 - 1	CCV59	18895	5.2110	-	7470A	PPB	26 Oct 2021 14:05:26	U	SMPL
CCB59 - 1	CCB59	-43	-0.0130	-	7470A	PPB	26 Oct 2021 14:07:41	U	SMPL
M4332-15LX5 - 1		35	0.0085	-	7470A	PPB	26 Oct 2021 14:09:57	U	SMPL
M4332-15A - 1		15919	4.3901	-	7470A	PPB	26 Oct 2021 14:12:12	U	SMPL
M4326-04LX5 - 1		-19	-0.0064	-	7470A	PPB	26 Oct 2021 14:14:28	U	SMPL
M4326-04A - 1		15797	4.3565	-	7470A	PPB	26 Oct 2021 14:16:43	U	SMPL
CCV60 - 1	CCV60	19665	5.4234	-	7470A	PPB	26 Oct 2021 14:25:32	U	SMPL
CCB60 - 1	CCB60	-21	-0.0069	-	7470A	PPB	26 Oct 2021 14:27:47	U	SMPL