

Clear table

Instrument ID: IC-2												
Analyst : AP Method: 300.0												
ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst	
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-10-11_	10/11/21 8:44		AP/MA	
STD2	0.4300	0.6780	0.6690	2.2590	0.5740	0.8870	3.5680	2021-10-11_	10/11/21 9:18		AP/MA	
STD3	0.8200	1.1940	1.1860	4.0070	1.0020	2.0340	6.0080	2021-10-11_	10/11/21 9:46		AP/MA	
STD4	0.9190	1.4550	1.4640	4.8490	1.2160	2.3710	7.2560	2021-10-11_	10/11/21 10:14		AP/MA	
STD5	1.9910	2.9530	2.9660	9.8230	2.4380	5.2410	14.5130	2021-10-11_	10/11/21 10:42		AP/MA	
STD6	4.2170	5.9970	5.9890	19.9420	4.9810	10.2720	29.8810	2021-10-11_	10/11/21 11:10		AP/MA	
STD7	4.8230	7.5230	7.5250	25.1200	6.2890	12.1950	37.7730	2021-10-11_	10/11/21 11:38		AP/MA	
ICV	2.0280	2.8130	2.8470	9.4620	2.3390	5.3320	13.9310	2021-10-11_	10/11/21 12:07		AP/MA	
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-10-11_	10/11/21 12:35		AP/MA	
CCV	2.1840	2.8010	2.7850	9.3560	2.3240	4.9830	13.6930	2021-10-26_	10/26/21 8:48		AP/MA	
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-10-26_	10/26/21 9:22		AP/MA	
LB117022BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-10-26_	10/26/21 9:50		AP/MA	
LB117022BSW	2.1700	2.8280	2.8430	9.4600	2.3620	5.2660	13.9470	2021-10-26_	10/26/21 10:18		AP/MA	
M4362-01	0.0000	0.0000	0.2550	0.0000	0.5070	0.0000	0.0000	2021-10-26_	10/26/21 12:02		AP/MA	
M4362-02	0.0000	0.0000	0.0000	0.0000	0.3700	0.0000	0.0000	2021-10-26_	10/26/21 12:30		AP/MA	
M4362-03	0.0000	0.3760	0.0000	0.0000	0.2950	0.0000	0.0000	2021-10-26_	10/26/21 12:58		AP/MA	
M4362-03MS	2.1710	3.2240	3.0250	10.0030	2.6500	5.3770	14.7580	2021-10-26_	10/26/21 13:26		AP/MA	
M4362-03MSD	2.1410	3.0780	2.8480	9.4110	2.5180	5.1940	13.8630	2021-10-26_	10/26/21 13:54		AP/MA	
M4372-15	0.4670	289.9090	0.0000	0.0000	1.4340	0.1800	31.1380	2021-10-26_	10/26/21 15:19		AP/MA	
CCV	1.9660	2.7710	2.8310	9.3210	2.3140	5.3640	13.7380	2021-10-26_	10/26/21 16:52		AP/MA	
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-10-26_	10/26/21 17:21		AP/MA	


Clear table

Instrument ID: IC-2 Analyst: AP Method: 300 / 9056A

ident	concentra tion F-	concentratio n CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2021-10-11_	10/11/21 8:44		AP/MA
STD2	0.4300	0.6780	0.6690	2.2590	0.5740	0.8870	3.5680	2021-10-11_	10/11/21 9:18		AP/MA
STD3	0.8200	1.1940	1.1860	4.0070	1.0020	2.0340	6.0080	2021-10-11_	10/11/21 9:46		AP/MA
STD4	0.9190	1.4550	1.4640	4.8490	1.2160	2.3710	7.2560	2021-10-11_	10/11/21 10:14		AP/MA
STD5	1.9910	2.9530	2.9660	9.8230	2.4380	5.2410	14.5130	2021-10-11_	10/11/21 10:42		AP/MA
STD6	4.2170	5.9970	5.9890	19.9420	4.9810	10.2720	29.8810	2021-10-11_	10/11/21 11:10		AP/MA
STD7	4.8230	7.5230	7.5250	25.1200	6.2890	12.1950	37.7730	2021-10-11_	10/11/21 11:38		AP/MA
ident	True Value F-	True Value CL-	True Value NO2	True Value BR-	True Value NO3	True Value HPO4	True Value SO4				
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000				
STD2	0.4000	0.6000	0.6000	2.0000	0.5000	1.0000	3.0000				
STD3	0.8000	1.2000	1.2000	4.0000	1.0000	2.0000	6.0000				
STD4	1.0000	1.5000	1.5000	5.0000	1.2500	2.5000	7.5000				
STD5	2.0000	3.0000	3.0000	10.0000	2.5000	5.0000	15.0000				
STD6	4.0000	6.0000	6.0000	20.0000	5.0000	10.0000	30.0000				
STD7	5.0000	7.5000	7.5000	25.0000	6.2500	12.5000	37.5000				
ident	Relative Error F-	Relative Error CL-	Relative Error NO2	Relative Error BR-	Relative Error NO3	Relative Error HPO4	Relative Error SO4				
STD1											
STD2	7.5000	13.0000	11.5000	12.9500	14.8000	-11.3000	18.9333				
STD3	2.5000	-0.5000	-1.1667	0.1750	0.2000	1.7000	0.1333				
STD4	-8.1000	-3.0000	-2.4000	-3.0200	-2.7200	-5.1600	-3.2533				
STD5	-0.4500	-1.5667	-1.1333	-1.7700	-2.4800	4.8200	-3.2467				
STD6	5.4250	-0.0500	-0.1833	-0.2900	-0.3800	2.7200	-0.3967				
STD7	-3.5400	0.3067	0.3333	0.4800	0.6240	-2.4400	0.7280				

CALIBRATION OF COMPONENT F-

Method: AnionsIC2-101121.mtw
 Equation: $Q = 0.0744588 \cdot A + 1.6749$
 RSD: 6.687 %
 Correlation coefficient: 0.997560



K3 = 0 K2 = 0 K1 = 0.0744588 K0 = 1.6749

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	1.78	92.91	0.4	20	5.681	
3	3.828	197.9	0.8	20	5.681	
4	4.665	224.2	1	20	5.681	
5	11.97	512.3	2	20	5.681	
6	24.28	1110	4	20	5.681	
7	29.23	1273	5	20	5.681	

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-101121.mtw
 Equation: $Q = 0.110027 \cdot A + 3.0598$
 RSD: 1.585 %
 Correlation coefficient: 0.999863



K3 = 0 K2 = 0 K1 = 0.110027 K0 = 3.0598

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1		0	0	0	0	0
2	6.9	95.4	0.6	20	8.597	
3	13.49	189.3	1.2	20	8.597	
4	17.36	236.6	1.5	20	8.597	
5	37.9	508.9	3	20	8.597	
6	80.05	1062	6	20	8.597	
7	102.5	1340	7.5	20	8.597	

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-101121.mtw
 Equation: $Q = 0.0487979 \cdot A + 3.14089$
 RSD: 1.366 %
 Correlation coefficient: 0.999898



K3 = 0 K2 = 0 K1 = 0.0487979 K0 = 3.14089

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	12.43	209.9	0.6	20	10.39	
3	24.79	421.8	1.2	20	10.39	
4	32.09	535.8	1.5	20	10.39	
5	67.48	1151	3	20	10.39	
6	137.9	2390	6	20	10.39	
7	172	3020	7.5	20	10.39	

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-101121.mtw
 Equation: $Q = 0.245006 \cdot A + 12.0971$
 RSD: 1.694 %
 Correlation coefficient: 0.999844




K3 = 0 K2 = 0 K1 = 0.245006 K0 = 12.0971
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	-0.00298	-0.000298		0	20		13.42
2	7.066	135		2	20		13.42
3	14.35	277.7		4	20		13.42
4	18.53	346.4		5	20		13.42
5	41.08	752.5		10	20		13.42
6	86.24	1579		20	20		13.42
7	109.2	2001		25	20		13.42

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-101121.mtw
 Equation: $Q = 0.0424654 \cdot A + 3.07921$
 RSD: 2.014 %
 Correlation coefficient: 0.999779



K3 = 0 K2 = 0 K1 = 0.0424654 K0 = 3.07921

Base: Area

Ref.channel: ch1

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	8.74	197.8	0.5	20	15.44	
3	17.58	399.6	1	20	15.44	
4	22.63	500.2	1.25	20	15.44	
5	49.87	1076	2.5	20	15.44	
6	106.7	2273	5	20	15.44	
7	137.2	2889	6.25	20	15.44	

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-101121.mtw
 Equation: $Q = 0.0894674 \cdot A - 6.00166$
 RSD: 4.593 %
 Correlation coefficient: 0.998849



K3 = 0 K2 = 0 K1 = 0.0894674 K0 = -6.00166

Base: Area

Ref.channel: chl

ISTD:


Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1	0	0	0	0	0	0
2	7.575	265.4	1	20	20.23	
3	15.5	521.8	2	20	20.23	
4	17.48	597.1	2.5	20	20.23	
5	39.73	1239	5	20	20.23	
6	78.86	2363	10	20	20.23	
7	92.96	2793	12.5	20	20.23	

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-101121.mtw
 Equation: $Q = 0.137023 \cdot A + 21.8065$
 RSD: 2.549 %
 Correlation coefficient: 0.999646



K3 = 0 K2 = 0 K1 = 0.137023 K0 = 21.8065

Base: Area

Ref.channel: chl

ISTD:

Formula: Linear

Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		0
2	12.98	361.7	3	20	23.11		
3	25.88	717.8	6	20	23.11		
4	33.03	899.9	7.5	20	23.11		
5	73.48	1959	15	20	23.11		
6	158.9	4202	30	20	23.11		
7	203.6	5354	37.5	20	23.11		

Report date: 10/13/2021 3:11:33 PM
Printed by: wet


Ident: STD1
Analysis from: 10/11/2021 8:44:43 AM
File: _2021-10-11_

Last save: 10/11/2021 12:37:14 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28563

Last save: 10/11/2021 11:3

SAMPLE:
: AP/MA
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/13/2021 3:11:40 PM
Printed by: wet


Ident: STD2
Analysis from: 10/11/2021 9:18:29 AM
File: _2021-10-11_

Last save: 10/11/2021 12:37:14 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28564

Last save: 10/11/2021 11:3

SAMPLE:
: AP/MA
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.73	0.645	1.78	3.09	92.906	6.84
2	8.64	0.191	6.90	12.00	95.396	7.02
3	10.42	0.230	12.43	21.63	209.868	15.45
4	13.49	0.279	7.07	12.30	135.041	9.94
5	15.47	0.324	8.74	15.21	197.760	14.56
6	20.44	0.449	7.58	13.18	265.362	19.54
7	23.63	0.414	12.98	22.58	361.696	26.63
7	25.50	0.362	57.47	99.99	1358.029	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/13/2021 3:11:46 PM
Printed by: wet


Ident: STD3
Analysis from: 10/11/2021 9:46:35 AM
File: _2021-10-11_

Last save: 10/11/2021 12:37:14 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28565

Last save: 10/11/2021 11:3

SAMPLE:
: AP/MA
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.78	0.701	3.83	3.31	197.851	7.26
2	8.63	0.193	13.49	11.69	189.272	6.94
3	10.41	0.229	24.79	21.48	421.823	15.48
4	13.47	0.280	14.35	12.43	277.703	10.19
5	15.44	0.324	17.58	15.23	399.569	14.66
6	20.51	0.437	15.50	13.43	521.785	19.14
7	23.71	0.412	25.88	22.42	717.810	26.33
7	25.50	0.368	115.43	100.00	2725.813	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/13/2021 3:11:52 PM
Printed by: wet


Ident: STD4
Analysis from: 10/11/2021 10:14:41 AM
File: _2021-10-11_

Last save: 10/11/2021 12:37:14 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28566

Last save: 10/11/2021 11:3

SAMPLE:
: AP/MA
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.73	0.622	4.66	3.20	224.221	6.71
2	8.58	0.189	17.36	11.91	236.601	7.08
3	10.32	0.227	32.09	22.01	535.827	16.04
4	13.33	0.271	18.53	12.71	346.427	10.37
5	15.25	0.316	22.63	15.52	500.211	14.97
6	20.38	0.439	17.47	11.99	597.117	17.88
7	23.57	0.407	33.03	22.66	899.950	26.94
7	25.50	0.353	145.77	100.00	3340.352	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/13/2021 3:11:57 PM
Printed by: wet


Ident: STD5
Analysis from: 10/11/2021 10:42:47 AM
File: _2021-10-11_

Last save: 10/11/2021 12:37:14 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28567

Last save: 10/11/2021 11:3

SAMPLE:
: AP/MA
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.72	0.554	11.97	3.72	512.343	7.12
2	8.57	0.187	37.90	11.79	508.883	7.07
3	10.29	0.234	67.48	20.99	1151.409	15.99
4	13.28	0.266	41.08	12.78	752.513	10.45
5	15.16	0.309	49.87	15.51	1075.917	14.95
6	20.36	0.411	39.73	12.36	1238.702	17.21
7	23.58	0.398	73.47	22.85	1959.243	27.22
7	25.50	0.337	321.50	100.00	7199.009	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/13/2021 3:12:03 PM
Printed by: wet


Ident: STD6
Analysis from: 10/11/2021 11:10:54 AM
File: _2021-10-11_

Last save: 10/11/2021 12:37:14 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28568

Last save: 10/11/2021 11:3

SAMPLE:
: AP/MA
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.76	0.627	24.28	3.61	1110.305	7.41
2	8.56	0.183	80.05	11.89	1062.370	7.09
3	10.26	0.237	137.89	20.49	2390.088	15.95
4	13.24	0.266	86.24	12.81	1578.541	10.54
5	15.07	0.303	106.74	15.86	2273.277	15.18
6	20.33	0.398	78.86	11.72	2363.242	15.78
7	23.56	0.393	158.91	23.61	4202.375	28.05
7	25.50	0.344	672.96	100.00	14980.198	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/13/2021 3:12:10 PM
 Printed by: wet


Ident: STD7
 Analysis from: 10/11/2021 11:38:59 AM
 File: _2021-10-11_

Last save: 10/11/2021 12:37:14 PM

Method: AnionsIC2-101121.mtw
 Run operator: wet
 Analysis number: 28569

Last save: 10/11/2021 11:3

SAMPLE:
 : AP/MA
 Vial number: 8
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.72	0.579	29.23	3.45	1272.991	6.82
2	8.51	0.180	102.53	12.11	1339.725	7.18
3	10.19	0.240	171.97	20.31	3019.898	16.17
4	13.17	0.266	109.19	12.90	2001.159	10.72
5	14.98	0.298	137.25	16.21	2889.232	15.47
6	20.26	0.398	92.96	10.98	2793.275	14.96
7	23.49	0.391	203.59	24.04	5354.184	28.68
7	25.50	0.336	846.71	100.00	18670.463	100.00

This report has been created by IC Net
 METROHM LTD

Report date: 10/13/2021 3:12:31 PM
Printed by: wet


Ident: ICV
Analysis from: 10/11/2021 12:07:05 PM
File: _2021-10-11_

Last save: 10/11/2021 12:32:35 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28570

Last save: 10/11/2021 12:0

SAMPLE:
: AP/MA
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.71	0.561	12.10	3.88	522.118	7.47
2	8.53	0.184	36.36	11.65	483.549	6.91
3	10.23	0.230	65.03	20.85	1102.404	15.76
4	13.19	0.262	39.95	12.81	723.012	10.34
5	15.04	0.304	48.29	15.48	1028.903	14.71
6	20.32	0.411	40.02	12.83	1258.914	18.00
7	23.53	0.398	70.20	22.50	1874.230	26.80
7	25.50	0.336	311.95	100.00	6993.130	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/13/2021 3:12:40 PM
Printed by: wet


Ident: ICB
Analysis from: 10/11/2021 12:35:11 PM
File: _2021-10-11_

Last save: 10/11/2021 1:00:41 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28571

Last save: 10/11/2021 12:0

SAMPLE:
: AP/MA
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/26/2021 3:10:23 PM
Printed by: wet


Ident: CCV
Analysis from: 10/26/2021 8:48:12 AM
File: _2021-10-26_

Last save: 10/26/2021 9:13:42 AM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28716

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.706	11.08	3.90	564.222	8.20
2	8.67	0.193	34.62	12.19	481.263	7.00
3	10.44	0.234	62.85	22.14	1076.919	15.65
4	13.53	0.287	36.30	12.79	714.376	10.38
5	15.47	0.333	44.02	15.50	1022.174	14.86
6	20.80	0.471	33.44	11.78	1181.085	17.17
7	24.00	0.448	61.61	21.70	1839.483	26.74
7	25.50	0.382	283.92	100.00	6879.524	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/26/2021 3:10:31 PM
Printed by: wet


Ident: CCB
Analysis from: 10/26/2021 9:22:15 AM
File: _2021-10-26_

Last save: 10/26/2021 9:47:45 AM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28717

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 12
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/26/2021 3:10:39 PM
Printed by: wet


Ident: LB117022BLW
Analysis from: 10/26/2021 9:50:23 AM
File: _2021-10-26_

Last save: 10/26/2021 1:03:49 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28718

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/26/2021 3:10:45 PM
Printed by: wet


Ident: LB117022BSW
Analysis from: 10/26/2021 10:18:31 AM
File: _2021-10-26_

Last save: 10/26/2021 1:03:50 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28719

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.78	0.670	11.44	3.78	560.351	7.97
2	8.59	0.188	35.87	11.87	486.305	6.92
3	10.32	0.228	65.74	21.75	1100.916	15.66
4	13.34	0.276	38.26	12.66	722.877	10.28
5	15.22	0.319	46.81	15.49	1039.774	14.79
6	20.72	0.441	37.34	12.35	1244.321	17.70
7	23.90	0.422	66.76	22.09	1876.591	26.69
7	25.50	0.364	302.22	100.00	7031.136	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/26/2021 3:10:52 PM
Printed by: wet


Ident: M4362-01
Analysis from: 10/26/2021 12:02:02 PM
File: _2021-10-26_

Last save: 10/26/2021 12:27:32 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28722

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	0.00	0.000	0.00	0.00	0.000	0.00
3	10.32	0.234	2.00	21.40	39.995	19.38
4	0.00	0.000	0.00	0.00	0.000	0.00
5	15.27	0.328	7.34	78.56	166.345	80.62
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	25.50	0.080	9.35	99.96	206.340	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/26/2021 3:10:59 PM
Printed by: wet


Ident: M4362-02
Analysis from: 10/26/2021 12:30:10 PM
File: _2021-10-26_

Last save: 10/26/2021 12:55:40 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28723

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	0.00	0.000	0.00	0.00	0.000	0.00
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	15.32	0.329	4.45	99.97	101.662	100.00
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	25.50	0.047	4.45	99.97	101.662	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/26/2021 3:11:05 PM
Printed by: wet


Ident: M4362-03
Analysis from: 10/26/2021 12:58:18 PM
File: _2021-10-26_

Last save: 10/26/2021 1:23:48 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28724

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	0.00	0.000	0.00	0.00	0.000	0.00
2	8.56	0.186	3.11	51.36	40.596	37.93
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	15.31	0.330	2.94	48.63	66.446	62.07
6	0.00	0.000	0.00	0.00	0.000	0.00
7	0.00	0.000	0.00	0.00	0.000	0.00
7	25.50	0.074	6.06	99.98	107.042	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/26/2021 3:11:21 PM
Printed by: wet


Ident: M4362-03MS
Analysis from: 10/26/2021 1:26:26 PM
File: _2021-10-26_

Last save: 10/26/2021 1:51:57 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28725

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 18
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.71	0.579	12.66	3.84	560.648	7.47
2	8.54	0.183	42.06	12.75	558.314	7.44
3	10.27	0.229	70.02	21.22	1175.521	15.67
4	13.26	0.272	41.02	12.43	767.191	10.23
5	15.12	0.312	53.69	16.27	1175.375	15.67
6	20.64	0.430	38.83	11.77	1269.179	16.92
7	23.82	0.416	71.72	21.73	1994.894	26.59
7	25.50	0.346	330.00	100.00	7501.123	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/26/2021 3:11:28 PM
Printed by: wet


Ident: M4362-03MSD
Analysis from: 10/26/2021 1:54:35 PM
File: _2021-10-26_

Last save: 10/26/2021 2:20:05 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28726

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 19
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.77	0.654	11.52	3.72	552.581	7.77
2	8.58	0.187	39.46	12.74	531.607	7.48
3	10.31	0.226	66.55	21.49	1102.755	15.51
4	13.34	0.274	38.25	12.35	718.814	10.11
5	15.21	0.316	50.37	16.27	1113.340	15.66
6	20.84	0.444	36.45	11.77	1228.183	17.27
7	24.07	0.417	67.01	21.64	1864.251	26.21
7	25.50	0.360	309.60	100.00	7111.530	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/26/2021 3:55:26 PM
 Printed by: wet


Ident: M4372-15
 Analysis from: 10/26/2021 3:19:47 PM
 File: _2021-10-26_

Last save: 10/26/2021 3:45:17 PM

Method: AnionsIC2-101121.mtw
 Run operator: wet
 Analysis number: 28729

Last save: 10/11/2021 12:3

SAMPLE:
 : AP/MA
 Vial number: 20
 Volume: 20.0 µL
 Dilution: 1.00
 Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.82	0.901	1.63	0.09	102.959	0.18
2	8.86	0.531	1613.09	89.23	52669.708	91.02
3	0.00	0.000	0.00	0.00	0.000	0.00
4	0.00	0.000	0.00	0.00	0.000	0.00
5	15.23	0.299	29.06	1.61	602.867	1.04
6	20.83	0.557	2.41	0.13	107.388	0.19
7	23.89	0.405	161.50	8.93	4385.795	7.58
7	25.50	0.385	1807.69	100.00	57868.717	100.00

This report has been created by IC Net
 METROHM LTD

Report date: 10/27/2021 9:10:16 AM
Printed by: wet


Ident: CCV
Analysis from: 10/26/2021 4:52:56 PM
File: _2021-10-26_

Last save: 10/26/2021 5:18:26 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28732

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.75	0.639	10.76	3.58	505.565	7.31
2	8.55	0.189	35.17	11.69	475.848	6.88
3	10.26	0.228	65.90	21.90	1096.114	15.84
4	13.23	0.276	37.79	12.56	711.465	10.28
5	15.10	0.316	46.37	15.41	1017.412	14.71
6	20.67	0.435	38.39	12.76	1266.068	18.30
7	23.86	0.417	66.51	22.10	1846.113	26.68
7	25.50	0.357	300.89	100.00	6918.586	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/27/2021 9:10:25 AM
Printed by: wet


Ident: CCB
Analysis from: 10/26/2021 5:21:04 PM
File: _2021-10-26_

Last save: 10/26/2021 5:46:34 PM

Method: AnionsIC2-101121.mtw
Run operator: wet
Analysis number: 28733

Last save: 10/11/2021 12:3

SAMPLE:
: AP/MA
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD