



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117026

Slope : 98.6

BalanceID: WC SC-6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/26/2021	14:25
2	CAL2	1	Water	NA	NA	20.3	7.00	10/26/2021	14:26
3	CAL3	1	Water	NA	NA	20.3	10.03	10/26/2021	14:27
4	ICV	1	Water	NA	NA	20.4	7.01	10/26/2021	14:30
5	CCV1	1	Water	NA	NA	20.4	2.01	10/26/2021	14:32
6	M4366-01	1	Solid	20.03	20	21.1	6.80	10/26/2021	14:35
7	M4366-01DUP	1	Solid	20.04	20	21.2	6.81	10/26/2021	14:36
8	M4366-02	1	Solid	20.01	20	20.9	7.29	10/26/2021	14:39
9	CCV2	1	Water	NA	NA	20.4	12.03	10/26/2021	14:40

10/26/2021

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4366 S WorkList ID : 153574 Department : Wet-Chemistry Date : 10-26-2021 14:15:50

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/26/2021	Solid	M4366-01	pH	Cool 4 deg C	PSEG03	K11	TRE-1520	10/26/2021	9045D
10/26/2021	Solid	M4366-02	pH	Cool 4 deg C	PSEG03	K11	TRE-1521	10/26/2021	9045D

Date/Time 10/26/21 14:20

Raw Sample Received by: JP (WQ)

Raw Sample Relinquished by: JP (WQ)

Date/Time 10/26/21 15:25

Raw Sample Received by: JP (WQ)

Raw Sample Relinquished by: JP (WQ)