



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 10/27/2021OVEN IN TEMP1: 103 °COVEN OUT TEMP1: 103 °CDATE & TIME IN1: 10/26/2021 16:00 DATE & TIME OUT1: 10/27/2021 08:00OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 10/27/2021 08:30 DATE & TIME OUT2: 10/27/2021 10:00Run Number: LB117033BalanceID: WC SC-6OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB117033BL	LB117033BL	76.0327	76.0327	100	76.0327	76.0327	76.0327	0.0000	0
2	M4372-15	GW-1	84.7575	84.7575	100	84.8435	84.8435	84.8435	0.0860	860
3	M4372-15DUP	GW-1DUP	68.8291	68.8291	100	68.9146	68.9146	68.9146	0.0855	855

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

VB117033

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS M4372 WorkList ID : 153677 Department : Wet-Chemistry Date : 10-26-2021 15:17:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/01/2021	Water	M4372-15	✓TS	Cool 4 deg C	PSEG03	K11	GW-1	10/26/2021	SM2540 B

Date/Time 10/26/21 15:20
Raw Sample Received by: JF (WC)
Raw Sample Relinquished by: KS (WC)

Date/Time 10/26/21 16:30
Raw Sample Received by: KS (WC)
Raw Sample Relinquished by: JF (WC)