



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 103 °C 10/26/2021 16:00 TEMP1 OUT: 103 °C 10/27/2021 08:00
TEMP2 IN: 103 °C 10/27/2021 08:30 TEMP2 OUT: 103 °C 10/27/2021 10:00
TEMP3 IN: 540 °C 10/27/2021 10:30 TEMP3 OUT: 540 °C 10/27/2021 12:00
TEMP4 IN: 560 °C 10/27/2021 12:30 TEMP4 OUT: 560 °C 10/27/2021 14:00

Run Number: LB117034

SUPERVISOR: apatel

ANALYST: jignesh

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB117034BL	76.0327	76.0327	100	76.0327	76.0327	76.0327	76.0327	0.0000	0
2	M4372-15	84.7575	84.7575	100	84.8435	84.8435	84.8355	84.8355	0.0080	80
3	M4372-15DUP	68.8291	68.8291	100	68.9146	68.9146	68.9068	68.9068	0.0078	78

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \times 1000000$

WORKLIST(Hardcopy Internal Chain)

28117034

WorkList Name : TVS M4372 WorkList ID : 153678 Department : Wet-Chemistry Date : 10-26-2021 15:18:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/01/2021	Water	M4372-15	TVS	Cool 4 deg C	PSEG03	K11	GW-1	10/26/2021	SM2540 G

Date/Time 10/26/21 15:20
 Raw Sample Received by: JS (WC)
 Raw Sample Relinquished by: JS (WC)

Date/Time 10/26/22 16:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JS (WC)