



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117041

Ambient Barometric Pressure (mmHg): 755.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.9	10/27/2021	09:15
2	M4359-06		1	10	100.00	>212.0	>212.0	10/27/2021	12:15
3	M4359-06DUP		1	10	100.00	>212.0	>212.0	10/27/2021	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117041

WorkList Name : fp102621-1 WorkList ID : 153681 Department : Wet-Chemistry Date : 10-26-2021 16:26:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/02/2021	Water	M4359-06 B	Flash Point	Cool 4 deg C	PSEG03	K11	33508	10/26/2021	1010B

Date/Time 10/27/21 08:45
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: [Signature]

Date/Time 10/27/21 12:50
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: KS (WC)