

LB117043


7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

3242E



A= 0.0000e+000

B= 3.0888e-004

C= -6.2625e-002

Rho= 0.9998610

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.0	0.000	-0.019	-0.019	140	0.000	140					1
0.2	0.200	0.220	0.020	914	0.0 %	914					10
2.5	2.500	2.423	-0.077	8047	0.0 %	8047					-3
5.0	5.000	5.118	0.118	16771	0.0 %	16771					2
7.5	7.500	7.506	0.006	24502	0.0 %	24502					0
10.0	10.000	9.954	-0.046	32428	0.0 %	32428					0

LB117043 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Oct 2021 11:50:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	140	-	0.0000	7470A	PPB	27 Oct 2021 11:53:02	S	Std
0.2 - 1	50.2	914	-	0.2000	7470A	PPB	27 Oct 2021 11:55:19	S	Std
2.5 - 1	52.5	8047	-	2.5000	7470A	PPB	27 Oct 2021 11:57:36	S	Std
5.0 - 1	55	16771	-	5.0000	7470A	PPB	27 Oct 2021 11:59:53	S	Std
7.5 - 1	57.5	24502	-	7.5000	7470A	PPB	27 Oct 2021 12:04:44	S	Std
10.0 - 1	510	32428	-	10.0000	7470A	PPB	27 Oct 2021 12:11:26	S	Std
ICV60 - 1	ICV60	13835	4.2107	-	7470A	PPB	27 Oct 2021 12:14:25	U	SMPL
ICB60 - 1	ICB60	166	-0.0114	-	7470A	PPB	27 Oct 2021 12:16:40	U	SMPL
CCV66 - 1	CCV66	17072	5.2105	-	7470A	PPB	27 Oct 2021 12:18:55	U	SMPL
CCB66 - 1	CCB66	162	-0.0126	-	7470A	PPB	27 Oct 2021 12:21:10	U	SMPL
CRA - 1	CRA	937	0.2268	-	7470A	PPB	27 Oct 2021 12:23:25	U	SMPL
HighStd - 1	HighStd	33564	10.3046	-	7470A	PPB	27 Oct 2021 12:25:40	U	SMPL
ChkStd - 1	ChkStd	22936	7.0218	-	7470A	PPB	27 Oct 2021 12:27:56	U	SMPL
PB140262BL - 1	PBW	-54	-0.0793	-	7470A	PPB	27 Oct 2021 12:30:12	U	SMPL
PB140262BS - 1	LCSW	13964	4.2506	-	7470A	PPB	27 Oct 2021 12:32:31	U	SMPL
M4340-02 - 1	TH-ES	17	-0.0574	-	7470A	PPB	27 Oct 2021 12:34:47	U	SMPL
M4340-05 - 1	TH-ED	67	-0.0419	-	7470A	PPB	27 Oct 2021 12:37:03	U	SMPL
M4340-08 - 1	TH-EES	86	-0.0361	-	7470A	PPB	27 Oct 2021 12:39:20	U	SMPL
M4340-11 - 1	TH-EED	141	-0.0191	-	7470A	PPB	27 Oct 2021 12:41:37	U	SMPL
M4358-02 - 1	TH-FS	158	-0.0138	-	7470A	PPB	27 Oct 2021 12:43:52	U	SMPL
CCV67 - 1	CCV67	16517	5.0391	-	7470A	PPB	27 Oct 2021 12:46:07	U	SMPL
CCB67 - 1	CCB67	160	-0.0132	-	7470A	PPB	27 Oct 2021 12:48:23	U	SMPL
M4358-05 - 1	TH-FD	205	0.0007	-	7470A	PPB	27 Oct 2021 12:50:38	U	SMPL
M4358-08 - 1	TH-GS	200	-0.0008	-	7470A	PPB	27 Oct 2021 12:52:53	U	SMPL
M4358-11 - 1	TH-GD	191	-0.0036	-	7470A	PPB	27 Oct 2021 12:55:09	U	SMPL
M4358-11DUP - 1	TH-GDDUP	216	0.0041	-	7470A	PPB	27 Oct 2021 12:57:25	U	SMPL
M4358-11MS - 1	TH-GDMS	12500	3.7984	-	7470A	PPB	27 Oct 2021 12:59:40	U	SMPL
M4358-11MSD - 1	TH-GDMSD	13412	4.0801	-	7470A	PPB	27 Oct 2021 13:01:57	U	SMPL
PB140263BL - 1	PBW	24	-0.0552	-	7470A	PPB	27 Oct 2021 13:04:13	U	SMPL
PB140263BS - 1	LCSW	13515	4.1119	-	7470A	PPB	27 Oct 2021 13:06:30	U	SMPL
M4336-01 - 1	MW-7	113	-0.0277	-	7470A	PPB	27 Oct 2021 13:08:47	U	SMPL
M4336-01DUP - 1	MW-7DUP	194	-0.0027	-	7470A	PPB	27 Oct 2021 13:11:02	U	SMPL
CCV68 - 1	CCV68	17010	5.1914	-	7470A	PPB	27 Oct 2021 13:13:18	U	SMPL
CCB68 - 1	CCB68	161	-0.0129	-	7470A	PPB	27 Oct 2021 13:15:33	U	SMPL
M4336-01MS - 1	MW-7MS	13388	4.0726	-	7470A	PPB	27 Oct 2021 13:17:49	U	SMPL
M4336-01MSD - 1	MW-7MSD	12634	3.8397	-	7470A	PPB	27 Oct 2021 13:20:04	U	SMPL
M4347-01 - 1	TOTE-2223	2114	0.5903	-	7470A	PPB	27 Oct 2021 13:22:20	U	SMPL
PB140262TB - 1	PB140224TB	19	-0.0568	-	7470A	PPB	27 Oct 2021 13:24:36	U	SMPL
M4358-11X5 - 1		76	-0.0391	-	7470A	PPB	27 Oct 2021 13:26:51	U	SMPL
M4358-11A - 1		14165	4.3126	-	7470A	PPB	27 Oct 2021 13:29:08	U	SMPL
M4336-01LX5 - 1		50	-0.0472	-	7470A	PPB	27 Oct 2021 13:31:24	U	SMPL
M4336-01A - 1		14007	4.2638	-	7470A	PPB	27 Oct 2021 13:33:41	U	SMPL
CCV69 - 1	CCV69	17266	5.2705	-	7470A	PPB	27 Oct 2021 13:35:58	U	SMPL
CCB69 - 1	CCB69	164	-0.0120	-	7470A	PPB	27 Oct 2021 13:38:14	U	SMPL