



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117044

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/27/2021	11:15
2	CAL2	1	Water	NA	NA	20.6	7.00	10/27/2021	11:16
3	CAL3	1	Water	NA	NA	20.6	10.03	10/27/2021	11:19
4	ICV	1	Water	NA	NA	20.7	6.99	10/27/2021	11:20
5	CCV1	1	Water	NA	NA	20.7	2.01	10/27/2021	11:23
6	M4359-06	1	Water	NA	NA	20.1	7.40	10/27/2021	11:25
7	M4383-01	1	Water	NA	NA	20.4	5.10	10/27/2021	11:30
8	M4383-01DUP	1	Water	NA	NA	20.5	5.11	10/27/2021	11:31
9	CCV2	1	Water	NA	NA	20.6	12.03	10/27/2021	11:35

WORKLIST(Hardcopy Internal Chain)

2B117045

WorkList Name : ph m4383 WorkList ID : 153707 Department : Wet-Chemistry Date : 10-27-2021 11:09:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/26/2021	Water	M4359-06	✓ pH	Cool 4 deg C	PSEG03	K11	33508	10/26/2021	9040C
10/22/2021	Water	M4383-01	✗ pH	Cool 4 deg C	STER14	D11	JS10000	10/22/2021	9040C

Date/Time 10/27/21 11:10
 Raw Sample Received by: SO (0207)
 Raw Sample Relinquished by: SO (0207)

Date/Time 10/27/21 12:20
 Raw Sample Received by: SO (0207)
 Raw Sample Relinquished by: SO (0207)