



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/28/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 10/27/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:16
Out Date: 10/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117048

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M4298-04	NB-08-102021	1	1.18	9.64	8.3	84.2	
M4375-01	SB-6(1.0-1.5)	2	1.14	9.67	8.6	87.5	
M4375-02	SB-7(1.0-1.5)	3	1.16	9.73	8.89	90.2	
M4375-03	SB-8(1.0-1.5)	4	1.12	9.79	9.14	92.5	
M4375-04	SB-9(2.0-2.5)	5	1.15	9.96	9.41	93.8	
M4375-05	SB-10(1.0-1.5)	6	1.12	9.66	8.64	88.1	
M4375-06	SB-11(1.0-1.5)	7	1.13	9.6	8.77	90.2	
M4375-07	SB-12(1.0-1.5)	8	1.16	9.88	8.85	88.2	
M4375-08	SB-13(2.5-3.0)	9	1.15	9.83	8.94	89.7	
M4375-09	SB-14(17-17.5)	10	1.18	9.97	9.27	92.0	
M4375-11	SB-15(1.0-1.5)	11	1.17	9.77	8.97	90.7	
M4376-01	DRUM-1	12	1.00	2.00	2.00	100.0	OILY DEBRIS
M4378-01	WASTE-CHAR- (EXCESS)	13	1.15	9.58	8.56	87.9	
M4380-01	WASTE-CHAR- (SUSPECT)	14	1.16	9.76	9.3	94.7	
M4384-01	COMP-A	45	1.19	9.74	9.17	93.3	
M4384-02	COMP-B	46	1.16	9.96	8.58	84.3	
M4384-03	COMP-C	47	1.17	9.65	7.78	77.9	
M4384-04	1890	48	1.00	2.00	2.00	100.0	OIL SAMPLE,
M4386-01	PT-AN-SOIL	15	1.00	2.00	2.00	100.0	
M4386-02	PT-CORR-SOIL	16	1.00	2.00	2.00	100.0	
M4386-03	PT-CORR-SOIL	17	1.00	2.00	2.00	100.0	
M4386-04	PT-CN-SOIL	18	1.00	2.00	2.00	100.0	
M4386-05	PT-CN-SOIL	19	1.00	2.00	2.00	100.0	
M4386-06	PT-FP-SOIL	20	1.00	2.00	2.00	100.0	
M4386-07	PT-CR6-SOIL	21	1.00	2.00	2.00	100.0	
M4386-08	PT-NUT-SOIL	22	1.00	2.00	2.00	100.0	
M4386-09	PT-NUT-SOIL	23	1.00	2.00	2.00	100.0	
M4386-10	PT-OGR-SOIL	24	1.00	2.00	2.00	100.0	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/28/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:30
In Date: 10/27/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:16
Out Date: 10/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117048

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4386-11	PT-MET-SOIL	25	1.00	2.00	2.00	100.0	
M4386-12	PT-BNA-SOIL	26	1.00	2.00	2.00	100.0	
M4386-13	PT-TRIAZINE-SOIL	27	1.00	2.00	2.00	100.0	
M4386-14	PT-PAH-SOIL	28	1.00	2.00	2.00	100.0	
M4386-15	PT-DIES-SOIL	29	1.00	2.00	2.00	100.0	
M4386-16	PT-GAS-SOIL	30	1.00	2.00	2.00	100.0	
M4386-17	PT-NJEPH-SOIL	31	1.00	2.00	2.00	100.0	
M4386-18	PT-HERB-SOIL	32	1.00	2.00	2.00	100.0	
M4386-19	PT-PCB-SOIL	33	1.00	2.00	2.00	100.0	
M4386-20	PT-PCBO-SOIL	34	1.00	2.00	2.00	100.0	
M4386-21	PT-PEST-SOIL	35	1.00	2.00	2.00	100.0	
M4386-22	PT-CHLR-COIL	36	1.00	2.00	2.00	100.0	
M4386-23	PT-TXP-SOIL	37	1.00	2.00	2.00	100.0	
M4386-24	PT-VOA-SOIL	38	1.00	2.00	2.00	100.0	
M4386-25	PT-SOL-SOIL	39	1.19	9.7	7.66	76.0	
M4387-01	SP-1	42	1.14	9.82	9.13	92.1	
M4387-02	SP-2	43	1.16	9.92	9.16	91.3	
M4387-03	SP-3	44	1.11	9.98	9.25	91.8	
M4388-01	72-12002	49	1.18	9.77	8.57	86.0	
M4388-02	72-12002-EPH-2	50	1.18	9.77	8.57	86.0	
M4389-01	IRV-2021-1A	51	1.00	2.00	2.00	100.0	PILC REEL
M4389-02	IRV-2021-2A	52	1.00	2.00	2.00	100.0	PILC REEL
M4389-03	IRV-2021-3A	53	1.00	2.00	2.00	100.0	PILC REEL
M4389-04	IRV-2021-3B	54	1.00	2.00	2.00	100.0	PILC REEL
M4389-05	IRV-2021-4A	55	1.00	2.00	2.00	100.0	PILC REEL
M4389-06	IRV-2021-4B	56	1.00	2.00	2.00	100.0	PILC REEL
M4391-01	PLA-135-TOP	40	1.00	2.00	2.00	100.0	WIPE SAMPLE
M4391-02	PLA-135-BOTTOM	41	1.00	2.00	2.00	100.0	WIPE SAMPLE

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$