


LB117049

7471B

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.1840e-004

C= -4.3934e-002

Rho= 0.9995022

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	-0.034	-0.034	46	0.000	46					-
0.2	0.200	0.161	-0.039	937	0.0 %	937					-20
2.5	2.500	2.445	-0.055	11395	0.0 %	11395					-2
5.0	5.000	5.253	0.253	24252	0.0 %	24252					5
7.5	7.500	7.482	-0.018	34459	0.0 %	34459					0
10.0	10.000	9.893	-0.107	45500	0.0 %	45500					-1

LB117049 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Oct 2021 14:54:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	46	-	0.0000	7471B	PPB	27 Oct 2021 14:56:14	S	Std
0.2 - 1	50.2	937	-	0.2000	7471B	PPB	27 Oct 2021 14:58:32	S	Std
2.5 - 1	52.5	11395	-	2.5000	7471B	PPB	27 Oct 2021 15:00:49	S	Std
5.0 - 1	55	24252	-	5.0000	7471B	PPB	27 Oct 2021 15:03:05	S	Std
7.5 - 1	57.5	34459	-	7.5000	7471B	PPB	27 Oct 2021 15:05:22	S	Std
10.0 - 1	510	45500	-	10.0000	7471B	PPB	27 Oct 2021 15:18:34	S	Std
ICV61 - 1	ICV61	19344	4.1809	-	7471B	PPB	27 Oct 2021 15:21:11	U	SMPL
ICB61 - 1	ICB61	-67	-0.0586	-	7471B	PPB	27 Oct 2021 15:23:27	U	SMPL
CCV70 - 1	CCV70	23101	5.0014	-	7471B	PPB	27 Oct 2021 15:25:42	U	SMPL
CCB70 - 1	CCB70	-75	-0.0603	-	7471B	PPB	27 Oct 2021 15:27:57	U	SMPL
CRA - 1	CRA	1190	0.2160	-	7471B	PPB	27 Oct 2021 15:35:22	U	SMPL
HighStd - 1	HighStd	45608	9.9171	-	7471B	PPB	27 Oct 2021 15:37:37	U	SMPL
ChkStd - 1	ChkStd	33983	7.3781	-	7471B	PPB	27 Oct 2021 15:39:52	U	SMPL
PB140295BL - 1	PBS	-116	-0.0693	-	7471B	PPB	27 Oct 2021 15:42:09	U	SMPL
PB140295BS - 1	LCSS	20764	4.4910	-	7471B	PPB	27 Oct 2021 15:44:27	U	SMPL
M4367-01 - 1	TR-05-102621	1322	0.2448	-	7471B	PPB	27 Oct 2021 15:46:43	U	SMPL
M4368-01 - 1	HD-02-102621	1501	0.2839	-	7471B	PPB	27 Oct 2021 15:48:59	U	SMPL
M4372-01 - 1	SB-1	388	0.0408	-	7471B	PPB	27 Oct 2021 15:51:16	U	SMPL
M4372-03 - 1	SB-2	2260	0.4497	-	7471B	PPB	27 Oct 2021 15:53:33	U	SMPL
M4372-05 - 1	SB-3	5013	1.0509	-	7471B	PPB	27 Oct 2021 15:55:48	U	SMPL
CCV71 - 1	CCV71	23845	5.1639	-	7471B	PPB	27 Oct 2021 15:58:03	U	SMPL
CCB71 - 1	CCB71	-40	-0.0527	-	7471B	PPB	27 Oct 2021 16:00:19	U	SMPL
M4372-07 - 1	SB-4	153	-0.0105	-	7471B	PPB	27 Oct 2021 16:02:34	U	SMPL
M4372-09 - 1	SB-5	9066	1.9361	-	7471B	PPB	27 Oct 2021 16:04:50	U	SMPL
M4372-11 - 1	SB-6	1941	0.3800	-	7471B	PPB	27 Oct 2021 16:07:05	U	SMPL
M4372-13 - 1	SB-7	2367	0.4730	-	7471B	PPB	27 Oct 2021 16:09:21	U	SMPL
M4373-01 - 1	SU-01-102621	915	0.1559	-	7471B	PPB	27 Oct 2021 16:11:36	U	SMPL
M4373-03 - 1	SU-04-102621	915	0.1559	-	7471B	PPB	27 Oct 2021 16:13:53	U	SMPL
M4373-03DUP - 1	SU-04-102621DUP	1041	0.1834	-	7471B	PPB	27 Oct 2021 16:18:44	U	SMPL
M4373-03MS - 1	SU-04-102621MS	20738	4.4853	-	7471B	PPB	27 Oct 2021 16:21:01	U	SMPL
M4373-03MSD - 1	SU-04-102621MSD	19841	4.2894	-	7471B	PPB	27 Oct 2021 16:23:18	U	SMPL
CCV72 - 1	CCV72	23192	5.0213	-	7471B	PPB	27 Oct 2021 16:25:34	U	SMPL
CCB72 - 1	CCB72	-48	-0.0544	-	7471B	PPB	27 Oct 2021 16:27:49	U	SMPL
M4373-03LX5 - 1		261	0.0131	-	7471B	PPB	27 Oct 2021 16:30:05	U	SMPL
M4373-03A - 1		20387	4.4087	-	7471B	PPB	27 Oct 2021 16:32:21	U	SMPL
CCV73 - 1	CCV73	23814	5.1572	-	7471B	PPB	27 Oct 2021 16:34:36	U	SMPL
CCB73 - 1	CCB73	-58	-0.0566	-	7471B	PPB	27 Oct 2021 16:36:52	U	SMPL