



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: jigneshDate: 10/28/2021OVEN IN TEMP1: 104 °COVEN OUT TEMP1: 104 °CDATE & TIME IN1: 10/27/2021 16:30 DATE & TIME OUT1: 10/28/2021 08:00OVEN IN TEMP2: 103 °COVEN OUT TEMP2: 103 °CDATE & TIME IN2: 10/28/2021 08:30 DATE & TIME OUT2: 10/28/2021 10:00Run Number: LB117054BalanceID: WC SC-6OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB117054BL	LB117054BL	74.9613	74.9613	100	74.9613	74.9613	74.9613	0.0000	0
2	M4390-01	FRAC-COMP	85.4469	85.4469	100	85.6045	85.6045	85.6045	0.1576	1576
3	M4390-01DUP	FRAC-COMPDUP	85.0802	85.0802	100	85.2381	85.2381	85.2381	0.1579	1579

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

WORKLIST(Hardcopy Internal Chain)

NR 117054

WorkList Name : TS M4390 W WorkList ID : 153722 Department : Wet-Chemistry Date : 10-27-2021 15:28:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/03/2021	Water	M4390-01	TS	Cool 4 deg C	PSEG03	K11	FRAC-COMP	10/27/2021	SM2540 B

Date/Time 10/27/21 15:30
 Raw Sample Received by: JP (WCC)
 Raw Sample Relinquished by: SP (WCC)

Date/Time 10/27/21 16:35
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: SP (WCC)