

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: jignesh

Date: 10/29/2021

OVEN IN TEMP1: 103 °C

OVEN OUT TEMP1: 103 °C

DATE & TIME IN1: 10/28/2021 16:30

DATE & TIME OUT1: 10/29/2021 08:00

OVEN IN TEMP2: 103 °C

OVEN OUT TEMP2: 104 °C

DATE & TIME IN2: 10/29/2021 08:30

DATE & TIME OUT2: 10/29/2021 10:00

Run Number: LB117065

BalanceID: WC SC-6

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB117065BL	LB117065BL	74.1361	74.1361	100	74.1361	74.1361	74.1361	0.0000	0
2	M4392-01	N41924	81.4879	81.4879	100	81.5137	81.5137	81.5137	0.0258	258
3	M4392-01DUP	N41924DUP	89.1712	89.1712	100	89.1972	89.1972	89.1972	0.0260	260

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

Result mg/L = ((C - B) / A) * 1000 * 1000

117065

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS M4392 WorkList ID : 153750 Department : Wet-Chemistry Date : 10-28-2021 12:26:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/02/2021	Water	M4392-01	TS	Cool 4 deg C	PSEG03	K11	N41924	10/28/2021	SM2540 B

Date/Time 10/28/21 13:00
Raw Sample Received by: JP (Cm)
Raw Sample Relinquished by: JP (Cm)

Date/Time 10/28/21 16:35
Raw Sample Received by: JP (Cm)
Raw Sample Relinquished by: JP (Cm)