



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117067

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/28/2021	13:00
2	CAL2	1	Water	NA	NA	20.3	7.00	10/28/2021	13:01
3	CAL3	1	Water	NA	NA	20.3	10.03	10/28/2021	13:02
4	ICV	1	Water	NA	NA	20.4	7.01	10/28/2021	13:05
5	CCV1	1	Water	NA	NA	20.4	2.01	10/28/2021	13:06
6	M4392-01	1	Water	NA	NA	20.6	7.40	10/28/2021	13:10
7	M4392-01DUP	1	Water	NA	NA	20.7	7.41	10/28/2021	13:11
8	CCV2	1	Water	NA	NA	20.4	12.03	10/28/2021	13:15

WORKLIST(Hardcopy Internal Chain)

13117067

WorkList Name : PH M4392 WorkList ID : 153749 Department : Wet-Chemistry Date : 10-28-2021 12:25:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/28/2021	Water	M4392-01	pH	Cool 4 deg C	PSEG03	K11	N41924	10/28/2021	9040C

Date/Time 10/28/21 12:30
Raw Sample Received by: JP (QC)
Raw Sample Relinquished by: JP (QC)

Date/Time 10/28/21 13:40
Raw Sample Received by: JP (QC)
Raw Sample Relinquished by: JP (QC)