



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/29/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 10/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:16
Out Date: 10/29/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117071

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4396-01	EX9S7	1	1.15	8.75	9.9	8.01	78.4	
M4396-02	EX9S8	2	1.16	8.81	9.97	8.74	86.0	
M4396-03	EX9S9	3	1.16	8.58	9.74	8.42	84.6	
M4396-04	EX9T0	4	1.15	8.81	9.96	8.4	82.3	
M4396-05	EX9T1	5	1.13	8.46	9.59	7.74	78.1	
M4396-06	EX9T2	6	1.11	8.55	9.66	7.93	79.8	
M4396-07	EX9T3	7	1.16	8.74	9.9	8.23	80.9	
M4396-08	EX9T4	8	1.19	8.34	9.53	7.89	80.3	
M4396-09	EX9T5	9	1.16	8.44	9.6	7.82	78.9	
M4396-10	EX9T6	10	1.16	8.80	9.96	7.95	77.2	
M4396-11	EX9T7	11	1.13	8.75	9.88	8.15	80.2	
M4396-12	EX9T8	12	1.11	8.54	9.65	7.75	77.8	
M4396-13	EX9T9	13	1.18	8.74	9.92	8.24	80.8	
M4396-14	EX9W0	14	1.16	8.76	9.92	7.86	76.5	
M4396-15	EX9W1	15	1.15	8.81	9.96	8.4	82.3	
M4396-16	EX9W2	16	1.16	8.48	9.64	7.71	77.2	
M4396-17	EX9W3	17	1.1	8.53	9.63	8.02	81.1	
M4396-18	EX9W4	18	1.16	8.54	9.7	7.46	73.8	
M4396-19	EX9W5	19	1.17	8.81	9.98	7.99	77.4	
M4396-20	EX9W8	20	1.14	8.38	9.52	7.61	77.2	
M4396-21	EX9W8D	21	1.14	8.38	9.52	7.61	77.2	
M4396-22	EX9W8S	22	1.14	8.38	9.52	7.61	77.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$