



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/29/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:35
In Date: 10/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:20
Out Date: 10/29/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117072

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4397-01	EX9W6	1	1.15	8.82	9.97	8.29	81.0	
M4397-02	EX9W7	2	1.19	8.69	9.88	7.8	76.1	
M4397-03	EX9W9	3	1.16	8.50	9.66	7.49	74.5	
M4397-04	EX9X6	4	1.19	8.53	9.72	7.55	74.6	
M4397-05	EX9X7	5	1.15	8.40	9.55	7.42	74.6	
M4397-06	EX9X8	6	1.15	8.42	9.57	7.44	74.7	
M4397-07	EX9X9	7	1.16	8.76	9.92	7.89	76.8	
M4397-08	EX9Y0	8	1.1	8.62	9.72	7.79	77.6	
M4397-09	EX9Y1	9	1.19	8.78	9.97	7.9	76.4	
M4397-10	EX9Y2	10	1.13	8.37	9.5	7.75	79.1	
M4397-11	EX9Y3	11	1.16	8.55	9.71	7.73	76.8	
M4397-12	EX9Y4	12	1.16	8.65	9.81	7.86	77.5	
M4397-13	EX9Y5	13	1.16	8.61	9.77	7.67	75.6	
M4397-14	EX9Y6	14	1.15	8.81	9.96	8.27	80.8	
M4397-15	EX9Y7	15	1.14	8.80	9.94	8.03	78.3	
M4397-16	EX9Y8	16	1.12	8.56	9.68	7.41	73.5	
M4397-17	EX9Y9	17	1.16	8.61	9.77	7.49	73.5	
M4397-18	EX9Z0	18	1.18	8.79	9.97	7.55	72.5	
M4397-19	EX9Z1	19	1.16	8.34	9.5	7.45	75.4	
M4397-20	EX9Z7	20	1.16	8.47	9.63	8.35	84.9	
M4397-21	EX9Z7D	21	1.16	8.47	9.63	8.35	84.9	
M4397-22	EX9Z7S	22	1.16	8.47	9.63	8.35	84.9	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$