

LB117081


SFAM01.1

INSTRUMENT ID: CV2

Linear

 μ Abs.:

36467



A= 0.0000e+000

B= 2.7851e-004 slope

C= -6.7763e-002 y-intercept

Rho= 0.9997922

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.040	-0.040	101	0.000	101					=
0.05	0.050					205	0				19
0.20	0.200	0.237	0.037	1093	0.0 %	1093					2
2.50	2.500	2.561	0.061	9438	0.0 %	9438					0
5.00	5.000	4.993	-0.007	18169	0.0 %	18169					-2
7.50	7.500	7.361	-0.139	26672	0.0 %	26672					1
10.0	10.000	10.089	0.089	36467	0.0 %	36467					

LB117081
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	101	-		0 SFAM01.1	PPB	10/28/2021 15:16	Std
	0.2 S01	1093	-		0.2 SFAM01.1	PPB	10/28/2021 15:18	Std
	2.5 S02	9438	-		2.5 SFAM01.1	PPB	10/28/2021 15:21	Std
	5 S03	18169	-		5 SFAM01.1	PPB	10/28/2021 15:25	Std
	7.5 S04	26672	-		7.5 SFAM01.1	PPB	10/28/2021 15:27	Std
	10 S05	36467	-		10 SFAM01.1	PPB	10/28/2021 15:32	Std
ICV087	ICV087	14210	3.8899 -		SFAM01.1	PPB	10/28/2021 15:35	SMPL
ICB087	ICB087	110	-0.0371 -		SFAM01.1	PPB	10/28/2021 15:37	SMPL
CCV074	CCV074	17904	4.9188 -		SFAM01.1	PPB	10/28/2021 15:39	SMPL
CCB074	CCB074	110	-0.0371 -		SFAM01.1	PPB	10/28/2021 15:41	SMPL
PB140317BL	PBW317	177	-0.0185 -		SFAM01.1	PPB	10/28/2021 15:44	SMPL
M4239-01	MH1U13	146	-0.0271 -		SFAM01.1	PPB	10/28/2021 15:46	SMPL
M4239-02	MH1U15	92	-0.0421 -		SFAM01.1	PPB	10/28/2021 15:48	SMPL
M4239-03	MH1U17	115	-0.0357 -		SFAM01.1	PPB	10/28/2021 15:50	SMPL
M4239-04	MH1U19	127	-0.0324 -		SFAM01.1	PPB	10/28/2021 15:53	SMPL
M4239-05	MH1U21	26	-0.0605 -		SFAM01.1	PPB	10/28/2021 15:55	SMPL
M4239-06	MH1U23	121	-0.0341 -		SFAM01.1	PPB	10/28/2021 15:57	SMPL
M4239-07	MH1U25	160	-0.0232 -		SFAM01.1	PPB	10/28/2021 15:59	SMPL
M4239-08	MH1U27	220	-0.0065 -		SFAM01.1	PPB	10/28/2021 16:02	SMPL
M4239-09	MH1U29	205	-0.0107 -		SFAM01.1	PPB	10/28/2021 16:04	SMPL
M4239-10	MH1U31	124	-0.0332 -		SFAM01.1	PPB	10/28/2021 16:06	SMPL
M4239-11	MH1U33	125	-0.0329 -		SFAM01.1	PPB	10/28/2021 16:08	SMPL
M4239-12	MH1U35	120	-0.0343 -		SFAM01.1	PPB	10/28/2021 16:11	SMPL
M4239-13	MH1U37	100	-0.0399 -		SFAM01.1	PPB	10/28/2021 16:13	SMPL
M4239-14	MH1U39	108	-0.0377 -		SFAM01.1	PPB	10/28/2021 16:15	SMPL
M4239-15	MH1U41	99	-0.0402 -		SFAM01.1	PPB	10/28/2021 16:17	SMPL
M4239-16	MH1U43	103	-0.0391 -		SFAM01.1	PPB	10/28/2021 16:20	SMPL
M4239-17	MH1U45	122	-0.0338 -		SFAM01.1	PPB	10/28/2021 16:22	SMPL
M4239-18	MH1V34	112	-0.0366 -		SFAM01.1	PPB	10/28/2021 16:24	SMPL
M4239-19	MH1V36	134	-0.0304 -		SFAM01.1	PPB	10/28/2021 16:27	SMPL
CCV075	CCV075	17479	4.8004 -		SFAM01.1	PPB	10/28/2021 16:29	SMPL
CCB075	CCB075	16	-0.0633 -		SFAM01.1	PPB	10/28/2021 16:31	SMPL
M4239-20	MH1V38	110	-0.0371 -		SFAM01.1	PPB	10/28/2021 16:33	SMPL
M4239-21	MH1V38D	131	-0.0313 -		SFAM01.1	PPB	10/28/2021 16:36	SMPL
M4239-22	MH1V38S	3499	0.9068 -		SFAM01.1	PPB	10/28/2021 16:38	SMPL
PB140318BL	PBW318	98	-0.0405 -		SFAM01.1	PPB	10/28/2021 16:40	SMPL
M4257-01	YAT26	126	-0.0327 -		SFAM01.1	PPB	10/28/2021 16:42	SMPL
M4257-02	YAT52	94	-0.0416 -		SFAM01.1	PPB	10/28/2021 16:45	SMPL
M4257-03	YAT52D	93	-0.0419 -		SFAM01.1	PPB	10/28/2021 16:47	SMPL
M4257-04	YAT52S	3898	1.0179 -		SFAM01.1	PPB	10/28/2021 16:49	SMPL
M4257-05	YAT54	78	-0.046 -		SFAM01.1	PPB	10/28/2021 16:51	SMPL
M4257-06	YAT22	367	0.0345 -		SFAM01.1	PPB	10/28/2021 16:54	SMPL
M4257-07	YAT24	514	0.0754 -		SFAM01.1	PPB	10/28/2021 16:56	SMPL
M4257-08	YAT46	364	0.0336 -		SFAM01.1	PPB	10/28/2021 16:58	SMPL
M4257-09	YAT48	357	0.0317 -		SFAM01.1	PPB	10/28/2021 17:01	SMPL
M4257-10	YAT50	374	0.0364 -		SFAM01.1	PPB	10/28/2021 17:03	SMPL

LB117081
INSTRUMENT ID : CV1

PB140319BL	PBW319	357	0.0317 -	SFAM01.1	PPB	10/28/2021 17:05 SMPL
M4258-01	YAT23	467	0.0623 -	SFAM01.1	PPB	10/28/2021 17:07 SMPL
M4258-02	YAT25	514	0.0754 -	SFAM01.1	PPB	10/28/2021 17:10 SMPL
M4258-03	YAT27	402	0.0442 -	SFAM01.1	PPB	10/28/2021 17:12 SMPL
M4258-04	YAT53	414	0.0475 -	SFAM01.1	PPB	10/28/2021 17:14 SMPL
M4258-05	YAT53D	431	0.0523 -	SFAM01.1	PPB	10/28/2021 17:16 SMPL
CCV076	CCV076	19475	5.3563 -	SFAM01.1	PPB	10/28/2021 17:19 SMPL
CCB076	CCB076	119	-0.0346 -	SFAM01.1	PPB	10/28/2021 17:21 SMPL
M4258-06	YAT53S	4533	1.1947 -	SFAM01.1	PPB	10/28/2021 17:23 SMPL
M4258-07	YAT55	663	0.1169 -	SFAM01.1	PPB	10/28/2021 17:25 SMPL
M4258-08	YAT47	601	0.0996 -	SFAM01.1	PPB	10/28/2021 17:28 SMPL
M4258-09	YAT49	414	0.0475 -	SFAM01.1	PPB	10/28/2021 17:30 SMPL
M4258-10	YAT51	68	-0.0488 -	SFAM01.1	PPB	10/28/2021 17:32 SMPL
CCV077	CCV077	19395	5.334 -	SFAM01.1	PPB	10/28/2021 17:35 SMPL
CCB077	CCB077	131	-0.0313 -	SFAM01.1	PPB	10/28/2021 17:37 SMPL