



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117084

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/28/2021	16:35
2	CAL2	1	Water	NA	NA	20.6	7.00	10/28/2021	16:36
3	CAL3	1	Water	NA	NA	20.6	10.02	10/28/2021	16:37
4	ICV	1	Water	NA	NA	20.6	6.98	10/28/2021	16:40
5	CCV1	1	Water	NA	NA	20.5	2.01	10/28/2021	16:42
6	M4406-07	1	Solid	20.03	20	20.2	6.66	10/28/2021	17:05
7	M4406-07DUP	1	Solid	20.04	20	20.3	6.67	10/28/2021	17:06
8	CCV2	1	Water	NA	NA	20.6	12.03	10/28/2021	17:10

WORKLIST(Hardcopy Internal Chain)

117084

WorkList Name : ph m4406 WorkList ID : 153768 Department : Wet-Chemistry Date : 10-28-2021 16:24:40

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/28/2021	Solid	M4406-07	pH	Cool 4 deg C	PSEG03	K11	CHRT-25062	10/28/2021	9045D

Date/Time 10/28/21 16:25
 Raw Sample Received by: JR 1601
 Raw Sample Relinquished by: RC CO'S-Lab

Date/Time 10/28/21 14:20
 Raw Sample Received by: RC CO'S-Lab
 Raw Sample Relinquished by: JR 1601