



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117085

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	10/28/2021	16:35
2	CAL2	1	Water	NA	NA	20.6	7.00	10/28/2021	16:36
3	CAL3	1	Water	NA	NA	20.6	10.02	10/28/2021	16:37
4	ICV	1	Water	NA	NA	20.6	6.98	10/28/2021	16:40
5	CCV1	1	Water	NA	NA	20.5	2.01	10/28/2021	16:42
6	M4405-01	1	Water	NA	NA	20.7	4.64	10/28/2021	16:45
7	M4405-01DUP	1	Water	NA	NA	20.8	4.65	10/28/2021	16:46
8	CCV2	1	Water	NA	NA	20.6	12.02	10/28/2021	16:50

114085

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m4405 WorkList ID : 153769 Department : Wet-Chemistry Date : 10-28-2021 16:25:19

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
10/22/2021	Water	M4405-01	pH	Cool 4 deg C	STER14	D11	MINI-CELL	10/22/2021	9040C

Date/Time 10/28/21 16:26
Raw Sample Received by: JP COIG
Raw Sample Relinquished by: JP COIG

Date/Time 10/28/21 17:20
Raw Sample Received by: JP COIG
Raw Sample Relinquished by: JP COIG