



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/29/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:30
In Date: 10/28/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:26
Out Date: 10/29/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117087

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4398-01	EXA01	1	1.14	8.49	9.63	7.85	79.0	
M4398-02	EXA02	2	1.15	8.80	9.95	8.49	83.4	
M4398-03	EXA03	3	1.16	8.81	9.97	8.2	79.9	
M4398-04	EXA04	4	1.16	8.63	9.79	7.93	78.4	
M4398-05	EXA05	5	1.19	8.52	9.71	7.98	79.7	
M4398-06	EXA06	6	1.14	8.80	9.94	8.27	81.0	
M4398-07	EXA07	7	1.12	8.38	9.5	7.42	75.2	
M4398-08	EXA08	8	1.19	8.74	9.93	7.78	75.4	
M4398-09	EXA09	9	1.16	8.47	9.63	7.71	77.3	
M4398-10	EXA10	10	1.19	8.78	9.97	7.78	75.1	
M4398-11	EXA11	11	1.16	8.81	9.97	7.75	74.8	
M4398-12	EXA13	12	1.12	8.44	9.56	7.57	76.4	
M4398-13	EXA14	13	1.15	8.62	9.77	7.62	75.1	
M4398-14	EXA15	14	1.12	8.46	9.58	7.5	75.4	
M4398-15	EXA16	15	1.19	8.51	9.7	7.84	78.1	
M4398-16	EXA17	16	1.16	8.42	9.58	7.62	76.7	
M4398-17	EXA18	17	1.15	8.56	9.71	7.71	76.6	
M4398-18	EXA19	18	1.16	8.82	9.98	7.64	73.5	
M4398-19	EXA20	19	1.16	8.62	9.78	7.42	72.6	
M4398-20	EXA21	20	1.19	8.51	9.7	7.75	77.1	
M4398-21	EXA21D	21	1.19	8.51	9.7	7.75	77.1	
M4398-22	EXA21S	22	1.19	8.51	9.7	7.75	77.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$