


LB117099

7471B

INSTRUMENT ID: CV1

 μ Abs.:

44384



Linear

A= 0.0000e+000

B= 2.2418e-004

C= -3.7655e-003

Rho= 0.9999381

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.006	0.006	43	0.000	43					-10
0.2	0.200	0.181	-0.019	824	0.0 %	824					-1
2.5	2.500	2.464	-0.036	11006	0.0 %	11006					-1
5.0	5.000	5.061	0.061	22593	0.0 %	22593					-1
7.5	7.500	7.542	0.042	33659	0.0 %	33659					-1
10.0	10.000	9.946	-0.054	44384	0.0 %	44384					-1

LB117099 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 29 Oct 2021 14:45:45

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	43	-	0.0000	7471B	PPB	29 Oct 2021 14:47:06	S	Std
0.2 - 1	50.2	824	-	0.2000	7471B	PPB	29 Oct 2021 14:49:24	S	Std
2.5 - 1	52.5	11006	-	2.5000	7471B	PPB	29 Oct 2021 14:51:41	S	Std
5.0 - 1	55	22593	-	5.0000	7471B	PPB	29 Oct 2021 14:53:58	S	Std
7.5 - 1	57.5	33659	-	7.5000	7471B	PPB	29 Oct 2021 14:56:14	S	Std
10.0 - 1	50	44384	-	10.0000	7471B	PPB	29 Oct 2021 14:58:31	S	Std
ICV63 - 1	ICV63	16731	3.7470	-	7471B	PPB	29 Oct 2021 15:01:22	U	SMPL
ICB63 - 1	ICB63	-75	-0.0206	-	7471B	PPB	29 Oct 2021 15:03:36	U	SMPL
CCV83 - 1	CCV83	23230	5.2040	-	7471B	PPB	29 Oct 2021 15:05:53	U	SMPL
CCB83 - 1	CCB83	-79	-0.0215	-	7471B	PPB	29 Oct 2021 15:08:08	U	SMPL
CRA - 1	CRA	808	0.1774	-	7471B	PPB	29 Oct 2021 15:10:25	U	SMPL
HighStd - 1	HighStd	44575	9.9892	-	7471B	PPB	29 Oct 2021 15:12:40	U	SMPL
ChkStd - 1	ChkStd	29358	6.5778	-	7471B	PPB	29 Oct 2021 15:19:26	U	SMPL
PB140365BL - 1	PBS	-131	-0.0331	-	7471B	PPB	29 Oct 2021 15:21:41	U	SMPL
PB140365BS - 1	LCSS	16329	3.6569	-	7471B	PPB	29 Oct 2021 15:24:00	U	SMPL
M4376-01 - 1	DRUM-1	93	0.0171	-	7471B	PPB	29 Oct 2021 15:26:16	U	SMPL
M4388-01 - 1	72-12002	7153	1.5998	-	7471B	PPB	29 Oct 2021 15:28:33	U	SMPL
M4395-01 - 1	MH-FFS	261	0.0547	-	7471B	PPB	29 Oct 2021 15:30:50	U	SMPL
M4395-04 - 1	MH-FFD	492	0.1065	-	7471B	PPB	29 Oct 2021 15:33:07	U	SMPL
M4395-07 - 1	MH-GGS	4326	0.9660	-	7471B	PPB	29 Oct 2021 15:35:22	U	SMPL
CCV84 - 1	CCV84	21281	4.7671	-	7471B	PPB	29 Oct 2021 15:37:37	U	SMPL
CCB84 - 1	CCB84	-82	-0.0221	-	7471B	PPB	29 Oct 2021 15:39:52	U	SMPL
M4395-10 - 1	MH-GGD	1431	0.3170	-	7471B	PPB	29 Oct 2021 15:42:09	U	SMPL
M4402-01 - 1	EO-01-102821	1857	0.4125	-	7471B	PPB	29 Oct 2021 15:44:24	U	SMPL
M4406-01 - 1	CHRT-20310	1756	0.3899	-	7471B	PPB	29 Oct 2021 15:46:39	U	SMPL
M4406-03 - 1	CHRT-27334	1462	0.3240	-	7471B	PPB	29 Oct 2021 15:48:55	U	SMPL
M4406-05 - 1	280243	1685	0.3740	-	7471B	PPB	29 Oct 2021 15:51:11	U	SMPL
M4406-05DUP - 1	280243DUP	1868	0.4150	-	7471B	PPB	29 Oct 2021 15:53:27	U	SMPL
M4406-05MS - 1	280243MS	18315	4.1021	-	7471B	PPB	29 Oct 2021 15:55:43	U	SMPL
M4406-05MSD - 1	280243MSD	17655	3.9542	-	7471B	PPB	29 Oct 2021 15:58:00	U	SMPL
M4406-05LX5 - 1		266	0.0559	-	7471B	PPB	29 Oct 2021 16:00:18	U	SMPL
M4406-05A - 1		19328	4.3292	-	7471B	PPB	29 Oct 2021 16:02:35	U	SMPL
CCV85 - 1	CCV85	21659	4.8518	-	7471B	PPB	29 Oct 2021 16:04:51	U	SMPL
CCB85 - 1	CCB85	-117	-0.0300	-	7471B	PPB	29 Oct 2021 16:07:08	U	SMPL