



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 11/1/2021

OVENTEMP IN Celsius(°C): 107

Time IN: 16:40

In Date: 10/29/2021

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:24

Out Date: 10/30/2021

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB117105

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4415-01	EXA41	1	1.15	8.78	9.93	8.45	83.1	
M4415-02	EXA42	2	1.16	8.82	9.98	8.25	80.4	
M4415-03	EXA43	3	1.18	8.54	9.72	7.79	77.4	
M4415-04	EXA45	4	1.16	8.81	9.97	8.26	80.6	
M4415-05	EXA46	5	1.15	8.81	9.96	8.26	80.7	
M4415-06	EXA47	6	1.15	8.80	9.95	8.35	81.8	
M4415-07	EXA48	7	1.16	8.70	9.86	7.84	76.8	
M4415-08	EXA49	8	1.16	8.81	9.97	8.5	83.3	
M4415-09	EXA50	9	1.12	8.57	9.69	7.7	76.8	
M4415-10	EXA51	10	1.16	8.81	9.97	8.24	80.4	
M4415-11	EXA52	11	1.16	8.56	9.72	7.98	79.7	
M4415-12	EXA53	12	1.19	8.66	9.85	7.95	78.1	
M4415-13	EXA54	13	1.15	8.81	9.96	8.05	78.3	
M4415-14	EXA55	14	1.1	8.67	9.77	7.91	78.5	
M4415-15	EXA67	15	1.17	8.41	9.58	7.86	79.5	
M4415-16	EXA67D	16	1.17	8.41	9.58	7.86	79.5	
M4415-17	EXA67S	17	1.17	8.41	9.58	7.86	79.5	
M4415-18	EXA68	18	1.15	8.78	9.93	8.08	78.9	
M4415-19	EXA69	19	1.17	8.43	9.6	7.96	80.5	
M4415-20	EXA70	20	1.19	8.60	9.79	7.96	78.7	
M4415-21	EXA71	21	1.15	8.83	9.98	8.55	83.8	
M4415-22	EXA72	22	1.16	8.81	9.97	8.32	81.3	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$