



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 11/1/2021

OVENTEMP IN Celsius(°C): 107

Time IN: 16:35

In Date: 10/29/2021

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:15

Out Date: 10/30/2021

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB117107

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4416-01	EXA56	1	1.15	8.83	9.98	8.71	85.6	
M4416-02	EXA57	2	1.14	8.80	9.94	7.89	76.7	
M4416-03	EXA58	3	1.14	8.81	9.95	7.66	74.0	
M4416-04	EXA59	4	1.16	8.77	9.93	7.93	77.2	
M4416-05	EXA60	5	1.16	8.81	9.97	8.56	84.0	
M4416-06	EXA61	6	1.15	8.81	9.96	7.64	73.7	
M4416-07	EXA62	7	1.16	8.62	9.78	7.23	70.4	
M4416-08	EXA63	8	1.13	8.56	9.69	7.65	76.2	
M4416-09	EXA64	9	1.17	8.81	9.98	7.44	71.2	
M4416-10	EXA65	10	1.14	8.84	9.98	8.12	79.0	
M4416-11	EXA66	11	1.14	8.76	9.9	8.52	84.2	
M4416-14	EXAB2	12	1.14	8.71	9.85	7.91	77.7	
M4416-15	EXAB3	13	1.16	8.79	9.95	7.47	71.8	
M4416-16	EXAB4	14	1.14	8.53	9.67	8.07	81.2	
M4416-17	EXAB5	15	1.13	8.59	9.72	8.19	82.2	
M4416-18	EXAB5D	16	1.13	8.59	9.72	8.19	82.2	
M4416-19	EXAB5S	17	1.13	8.59	9.72	8.19	82.2	
M4416-20	EXAB6	18	1.14	8.83	9.97	7.61	73.3	
M4416-21	EXAB7	19	1.12	8.83	9.95	8.14	79.5	
M4416-22	EXAB8	20	1.14	8.76	9.9	8.71	86.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$