



PERCENT SOLID

Supervisor: mohammad

Analyst: jignesh

Date: 11/1/2021

OVENTEMP IN Celsius(°C): 107

Time IN: 17:35

In Date: 10/29/2021

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103

Time OUT: 08:33

Out Date: 10/30/2021

Weight Check 1.0g: 1.00

Weight Check 10g: 10.00

BalanceID: M SC-4

Thermometer ID: %SOLIDS-OVEN

QC:LB117108

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4417-01	EXAC9	1	1.14	8.46	9.6	7.67	77.2	
M4417-02	EXAD0	2	1.13	8.43	9.56	7.52	75.8	
M4417-03	EXAD1	3	1.17	8.45	9.62	7.63	76.4	
M4417-04	EXAD2	4	1.13	8.41	9.54	7.44	75.0	
M4417-05	EXAD3	5	1.16	8.66	9.82	7.78	76.4	
M4417-06	EXAD4	6	1.14	8.84	9.98	8.04	78.1	
M4417-07	EXAD5	7	1.17	8.55	9.72	7.6	75.2	
M4417-08	EXAD6	8	1.16	8.53	9.69	7.98	80.0	
M4417-09	EXAD7	9	1.11	8.51	9.62	8.02	81.2	
M4417-10	EXAD8	10	1.11	8.59	9.7	7.88	78.8	
M4417-11	EXAD8D	11	1.11	8.59	9.7	7.88	78.8	
M4417-12	EXAD8S	12	1.11	8.59	9.7	7.88	78.8	
M4417-13	EXAD9	13	1.19	8.61	9.8	8.11	80.4	
M4417-14	EXAE0	14	1.17	8.63	9.8	7.97	78.8	
M4417-15	EXAE1	15	1.16	8.65	9.81	7.77	76.4	
M4417-16	EXAE2	16	1.17	8.80	9.97	8.18	79.7	
M4417-17	EXAE3	17	1.17	8.79	9.96	7.94	77.0	
M4417-18	EXAE4	18	1.15	8.80	9.95	7.75	75.0	
M4417-19	EXAE5	19	1.11	8.81	9.92	8.22	80.7	
M4417-20	EXAE6	20	1.15	8.81	9.96	8.43	82.6	
M4417-21	EXAE7	21	1.15	8.80	9.95	8.23	80.5	
M4417-22	EXAE8	22	1.17	8.39	9.56	7.67	77.5	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$