



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117122

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/01/2021	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	11/01/2021	16:15
3	CAL3	1	Water	NA	NA	20.4	10.03	11/01/2021	16:19
4	ICV	1	Water	NA	NA	20.3	7.02	11/01/2021	16:20
5	CCV1	1	Water	NA	NA	20.4	2.01	11/01/2021	16:22
6	M4434-01	1	Water	NA	NA	20.2	6.28	11/01/2021	16:25
7	M4434-01DUP	1	Water	NA	NA	20.3	6.29	11/01/2021	16:26
8	M4434-02	1	Water	NA	NA	20.5	6.61	11/01/2021	16:30
9	M4434-03	1	Water	NA	NA	20.6	6.44	11/01/2021	16:32
10	M4434-04	1	Water	NA	NA	20.5	6.60	11/01/2021	16:35
11	M4434-05	1	Water	NA	NA	20.5	8.42	11/01/2021	16:37
12	M4434-06	1	Water	NA	NA	20.6	6.48	11/01/2021	16:40
13	M4434-07	1	Water	NA	NA	20.5	6.32	11/01/2021	16:43
14	M4434-08	1	Water	NA	NA	20.4	6.60	11/01/2021	16:45
15	M4434-09	1	Water	NA	NA	20.2	6.16	11/01/2021	16:50
16	CCV2	1	Water	NA	NA	20.6	12.03	11/01/2021	17:00
17	M4434-10	1	Water	NA	NA	20.4	8.98	11/01/2021	17:01
18	M4434-11	1	Water	NA	NA	20.1	8.97	11/01/2021	17:05
19	M4434-12	1	Water	NA	NA	20.4	6.98	11/01/2021	17:06
20	M4434-13	1	Water	NA	NA	20.1	6.90	11/01/2021	17:10

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	M4435-01	1	Water	NA	NA	20.1	7.49	11/01/2021	17:15
22	M4435-02	1	Water	NA	NA	20.1	7.65	11/01/2021	17:20
23	M4435-05	1	Water	NA	NA	20.4	6.29	11/01/2021	17:22
24	M4435-06	1	Water	NA	NA	20.4	6.29	11/01/2021	17:23
25	CCV3	1	Water	NA	NA	20.6	2.01	11/01/2021	17:25

WORKLIST(Hardcopy Internal Chain)

VB 11/12/22

WorkList Name : PH M4434

WorkList ID : 153852

Department : Wet-Chemistry

Date : 11-01-2021 16:06:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/01/2021	Water	M4434-01	pH	Cool 4 deg C	PSEG03	L11	2220	11/01/2021	9040C
11/01/2021	Water	M4434-02	pH	Cool 4 deg C	PSEG03	L11	2221	11/01/2021	9040C
11/01/2021	Water	M4434-03	pH	Cool 4 deg C	PSEG03	L11	2222	11/01/2021	9040C
11/01/2021	Water	M4434-04	pH	Cool 4 deg C	PSEG03	L11	2223	11/01/2021	9040C
11/01/2021	Water	M4434-05	pH	Cool 4 deg C	PSEG03	L11	2224	11/01/2021	9040C
11/01/2021	Water	M4434-06	pH	Cool 4 deg C	PSEG03	L11	2225	11/01/2021	9040C
11/01/2021	Water	M4434-07	pH	Cool 4 deg C	PSEG03	L11	2226	11/01/2021	9040C
11/01/2021	Water	M4434-08	pH	Cool 4 deg C	PSEG03	L11	2227	11/01/2021	9040C
11/01/2021	Water	M4434-09	pH	Cool 4 deg C	PSEG03	L11	2228	11/01/2021	9040C
11/01/2021	Water	M4434-10	pH	Cool 4 deg C	PSEG03	L11	2230	11/01/2021	9040C
11/01/2021	Water	M4434-11	pH	Cool 4 deg C	PSEG03	L11	2231	11/01/2021	9040C
11/01/2021	Water	M4434-12	pH	Cool 4 deg C	PSEG03	L11	2229	11/01/2021	9040C
11/01/2021	Water	M4434-13	pH	Cool 4 deg C	PSEG03	L11	2004	11/01/2021	9040C
11/01/2021	Water	M4435-01	pH	Cool 4 deg C	PSEG03	L11	2233	11/01/2021	9040C
11/01/2021	Water	M4435-02	pH	Cool 4 deg C	PSEG03	L11	2234	11/01/2021	9040C
11/01/2021	Water	M4435-05	pH	Cool 4 deg C	PSEG03	L11	2239	11/01/2021	9040C
11/01/2021	Water	M4435-06	pH	Cool 4 deg C	PSEG03	L11	2240	11/01/2021	9040C

Date/Time 11/01/21 16:10
 Raw Sample Received by: JB 160C1
 Raw Sample Relinquished by: KS (WC)

Date/Time 11/01/21 17:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JB (WC)