



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117129

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/02/2021	11:25
2	CAL2	1	Water	NA	NA	20.3	7.00	11/02/2021	11:26
3	CAL3	1	Water	NA	NA	20.3	10.03	11/02/2021	11:27
4	ICV	1	Water	NA	NA	20.4	6.99	11/02/2021	11:30
5	CCV1	1	Water	NA	NA	20.4	2.01	11/02/2021	11:32
6	M4442-01	1	Water	NA	NA	20.2	4.53	11/02/2021	11:35
7	M4442-01DUP	1	Water	NA	NA	20.3	4.54	11/02/2021	11:36
8	CCV2	1	Water	NA	NA	20.4	12.02	11/02/2021	11:38

11-11-2021

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4442 WorkList ID : 153874 Department : Wet-Chemistry Date : 11-02-2021 11:15:29

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/01/2021	Water	M4442-01	pH	Cool 4 deg C	STER04	D11	POOL-WATER	11/01/2021	9040C

Date/Time 11/02/21 11:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 11/02/21 12:20
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]