



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117132

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2743

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	11/02/2021	14:15
2	M4434-14		1	12	100.00	>212.0	>212.0	11/02/2021	14:40
3	M4434-14DUP		1	12	100.00	>212.0	>212.0	11/02/2021	15:05
4	M4434-16		1	10	100.00	>212.0	>212.0	11/02/2021	15:35
5	M4435-03		1	15	100.00	>212.0	>212.0	11/02/2021	15:58
6	M4435-07		1	13	100.00	>212.0	>212.0	11/02/2021	16:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB17132

WorkList Name : fp11221 WorkList ID : 153879 Department : Wet-Chemistry Date : 11-02-2021 12:07:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/08/2021	Water	M4434-14	Flash Point	Cool 4 deg C	PSEG03	L11	20-28	11/01/2021	1010B
11/08/2021	Water	M4434-16	Flash Point	Cool 4 deg C	PSEG03	L11	30-31	11/01/2021	1010B
11/04/2021	Water	M4435-03	Flash Point	Cool 4 deg C	PSEG03	L11	33-34	11/01/2021	1010B
11/04/2021	Water	M4435-07	Flash Point	Cool 4 deg C	PSEG03	L11	39-40	11/01/2021	1010B

Date/Time 11/02/21 13:55
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: SP (WC)

Date/Time 11/02/21 16:25
 Raw Sample Received by: SP (WC)
 Raw Sample Relinquished by: KS (WC)