

LB117133


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

47240



A= 0.0000e+000

B= 2.1152e-004

C= 2.7088e-002

Rho= 0.9999806

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/00
0.0	0.000	0.034	0.034	34	0.000	34					-
0.2	0.200	0.194	-0.006	791	0.0 %	791					-3
2.5	2.500	2.464	-0.036	11523	0.0 %	11523					-1
5.0	5.000	5.005	0.005	23535	0.0 %	23535					0
7.5	7.500	7.483	-0.017	35248	0.0 %	35248					0
10.0	10.000	10.019	0.019	47240	0.0 %	47240					0

LB117133 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 02 Nov 2021 13:36:06

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7471B	PPB	02 Nov 2021 13:59:20	S	Std
0.2 - 1	50.2	791	-	0.2000	7471B	PPB	02 Nov 2021 14:01:37	S	Std
2.5 - 1	52.5	11523	-	2.5000	7471B	PPB	02 Nov 2021 14:03:53	S	Std
5.0 - 1	55	23535	-	5.0000	7471B	PPB	02 Nov 2021 14:06:10	S	Std
7.5 - 1	57.5	35248	-	7.5000	7471B	PPB	02 Nov 2021 14:08:27	S	Std
10.0 - 1	510	47240	-	10.0000	7471B	PPB	02 Nov 2021 14:10:43	S	Std
ICV65 - 1	ICV65	18621	3.9657	-	7471B	PPB	02 Nov 2021 14:13:22	U	SMPL
ICB65 - 1	ICB65	55	0.0387	-	7471B	PPB	02 Nov 2021 14:17:58	U	SMPL
CCV89 - 1	CCV89	22963	4.8842	-	7471B	PPB	02 Nov 2021 14:20:13	U	SMPL
CCB89 - 1	CCB89	-13	0.0243	-	7471B	PPB	02 Nov 2021 14:22:28	U	SMPL
CRA - 1	CRA	754	0.1866	-	7471B	PPB	02 Nov 2021 14:24:43	U	SMPL
HighStd - 1	HighStd	45196	9.5868	-	7471B	PPB	02 Nov 2021 14:26:58	U	SMPL
ChkStd - 1	ChkStd	31968	6.7889	-	7471B	PPB	02 Nov 2021 14:29:13	U	SMPL
PB140442BL - 1	PBS	-102	0.0055	-	7471B	PPB	02 Nov 2021 14:31:30	U	SMPL
PB140442BS - 1	LCSS	17895	3.8122	-	7471B	PPB	02 Nov 2021 14:36:27	U	SMPL
M4353-02 - 1	CONCRETE-GRAB-II	4915	1.0667	-	7471B	PPB	02 Nov 2021 14:38:43	U	SMPL
M4409-01 - 1	VNJ-207-ETGI-340	3122	0.6874	-	7471B	PPB	02 Nov 2021 14:40:59	U	SMPL
M4411-01 - 1	MH-HHS	2291	0.5117	-	7471B	PPB	02 Nov 2021 14:43:16	U	SMPL
M4411-04 - 1	MH-HHD	873	0.2117	-	7471B	PPB	02 Nov 2021 14:45:33	U	SMPL
M4422-01 - 1	OR-03-102921	17498	3.7282	-	7471B	PPB	02 Nov 2021 14:47:48	U	SMPL
CCV90 - 1	CCV90	24920	5.2981	-	7471B	PPB	02 Nov 2021 14:53:19	U	SMPL
CCB90 - 1	CCB90	12	0.0296	-	7471B	PPB	02 Nov 2021 14:57:30	U	SMPL
M4428-01 - 1	HD-01-102921	686	0.1722	-	7471B	PPB	02 Nov 2021 15:04:36	U	SMPL
M4428-01DUP - 1	HD-01-102921DUP	562	0.1460	-	7471B	PPB	02 Nov 2021 15:06:51	U	SMPL
M4428-01MS - 1	HD-01-102921MS	19540	4.1601	-	7471B	PPB	02 Nov 2021 15:09:06	U	SMPL
M4428-01MSD - 1	HD-01-102921MSD	19842	4.2240	-	7471B	PPB	02 Nov 2021 15:11:22	U	SMPL
M4429-01 - 1	OK-03-102921	910	0.2196	-	7471B	PPB	02 Nov 2021 15:13:38	U	SMPL
M4433-01 - 1	PSC-124055	2511	0.5582	-	7471B	PPB	02 Nov 2021 15:15:54	U	SMPL
M4433-03 - 1	PSC-124035	2047	0.4601	-	7471B	PPB	02 Nov 2021 15:18:11	U	SMPL
M4436-01 - 1	S15-SP	16465	3.5097	-	7471B	PPB	02 Nov 2021 15:20:28	U	SMPL
M4439-01 - 1	EO-01-110221	1609	0.3674	-	7471B	PPB	02 Nov 2021 15:22:44	U	SMPL
CCV91 - 1	CCV91	25207	5.3588	-	7471B	PPB	02 Nov 2021 15:25:00	U	SMPL
CCB91 - 1	CCB91	-23	0.0222	-	7471B	PPB	02 Nov 2021 15:27:16	U	SMPL
M4428-01LX5 - 1		89	0.0459	-	7471B	PPB	02 Nov 2021 15:29:31	U	SMPL
M4428-01A - 1		19862	4.2282	-	7471B	PPB	02 Nov 2021 15:31:47	U	SMPL
CCV92 - 1	CCV92	25232	5.3641	-	7471B	PPB	02 Nov 2021 15:34:02	U	SMPL
CCB92 - 1	CCB92	-34	0.0199	-	7471B	PPB	02 Nov 2021 15:36:18	U	SMPL