



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117135

Slope : 98.8

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.02	11/02/2021	16:55
2	CAL2	1	Water	NA	NA	20.3	7.00	11/02/2021	16:56
3	CAL3	1	Water	NA	NA	20.2	10.02	11/02/2021	16:57
4	ICV	1	Water	NA	NA	20.3	6.99	11/02/2021	16:58
5	CCV1	1	Water	NA	NA	20.4	2.01	11/02/2021	17:00
6	M4454-01	1	Solid	20	20	20.2	5.22	11/02/2021	17:01
7	M4454-01DUP	1	Solid	20	20	20.2	5.20	11/02/2021	17:02
8	CCV2	1	Water	NA	NA	20.3	12.01	11/02/2021	17:03

WORKLIST(Hardcopy Internal Chain)

LB 117135

WorkList Name : PH11221 WorkList ID : 153887 Department : Wet-Chemistry Date : 11-02-2021 16:45:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/02/2021	Solid	M4454-01	pH	Cool 4 deg C	PSEG03	L11	30860-OIL	11/02/2021	9045D

Date/Time 11/02/21 16:50
Raw Sample Received by: KS(WC)
Raw Sample Relinquished by: KS(WC)

Date/Time 11/02/21 17:05
Raw Sample Received by: KS(WC)
Raw Sample Relinquished by: KS(WC)