

LB 117146


7470A

INSTRUMENT ID: CVL

Linear

 μ Abs.:

36480



A= 0.0000e+000

B= 2.7495e-004

C= 3.6729e-002

Rho= 0.9998330

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	90D
0.0	0.000	0.051	0.051	51	0.000	51					2
0.2	0.200	0.203	0.003	603	0.0 %	603					1
2.5	2.500	2.529	0.029	9063	0.0 %	9063					1
5.0	5.000	4.861	-0.139	17546	0.0 %	17546					-3
7.5	7.500	7.491	-0.009	27110	0.0 %	27110					0
10.0	10.000	10.067	0.067	36480	0.0 %	36480					1

LB117146 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 03 Nov 2021 10:54:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	51	-	0.0000	7470A	PPB	03 Nov 2021 11:06:42	S	Std
0.2 - 1	50.2	603	-	0.2000	7470A	PPB	03 Nov 2021 11:08:59	S	Std
2.5 - 1	52.5	9063	-	2.5000	7470A	PPB	03 Nov 2021 11:11:16	S	Std
5.0 - 1	55	17546	-	5.0000	7470A	PPB	03 Nov 2021 11:13:32	S	Std
7.5 - 1	57.5	27110	-	7.5000	7470A	PPB	03 Nov 2021 11:15:49	S	Std
10.0 - 1	510	36480	-	10.0000	7470A	PPB	03 Nov 2021 11:20:37	S	Std
ICV66 - 1	ICV66	13302	3.6941	-	7470A	PPB	03 Nov 2021 11:23:17	U	SMPL
ICB66 - 1	ICB66	-276	-0.0392	-	7470A	PPB	03 Nov 2021 11:25:32	U	SMPL
CCV93 - 1	CCV93	18870	5.2250	-	7470A	PPB	03 Nov 2021 11:27:47	U	SMPL
CCB93 - 1	CCB93	-101	0.0090	-	7470A	PPB	03 Nov 2021 11:30:02	U	SMPL
CRA - 1	CRA	606	0.2033	-	7470A	PPB	03 Nov 2021 11:32:17	U	SMPL
HighStd - 1	HighStd	36149	9.9757	-	7470A	PPB	03 Nov 2021 11:34:33	U	SMPL
ChkStd - 1	ChkStd	25302	6.9934	-	7470A	PPB	03 Nov 2021 11:36:48	U	SMPL
PB140432BL - 1	PBW	-141	-0.0020	-	7470A	PPB	03 Nov 2021 11:39:05	U	SMPL
PB140432BS - 1	LCSW	13593	3.7741	-	7470A	PPB	03 Nov 2021 11:41:22	U	SMPL
M4409-03 - 1	VNJ-207-ETGI-340	-23	0.0304	-	7470A	PPB	03 Nov 2021 11:43:38	U	SMPL
M4411-02 - 1	MH-HHS	-16	0.0323	-	7470A	PPB	03 Nov 2021 11:45:55	U	SMPL
M4411-05 - 1	MH-HHD	-3	0.0359	-	7470A	PPB	03 Nov 2021 11:48:12	U	SMPL
M4411-05DUP - 1	MH-HHDDUP	7	0.0387	-	7470A	PPB	03 Nov 2021 11:50:29	U	SMPL
M4411-05MS - 1	MH-HHDMS	13311	3.6965	-	7470A	PPB	03 Nov 2021 11:52:44	U	SMPL
CCV94 - 1	CCV94	18538	5.1337	-	7470A	PPB	03 Nov 2021 11:54:59	U	SMPL
CCB94 - 1	CCB94	-96	0.0103	-	7470A	PPB	03 Nov 2021 11:57:14	U	SMPL
M4411-05MSD - 1	MH-HHDMSD	13420	3.7265	-	7470A	PPB	03 Nov 2021 11:59:30	U	SMPL
PB140434BL - 1	PBW	-82	0.0142	-	7470A	PPB	03 Nov 2021 12:01:45	U	SMPL
PB140434BS - 1	LCSW	13008	3.6132	-	7470A	PPB	03 Nov 2021 12:04:01	U	SMPL
M4433-02 - 1	PSC-124055	-80	0.0147	-	7470A	PPB	03 Nov 2021 12:06:16	U	SMPL
M4433-04 - 1	PSC-124035	2	0.0373	-	7470A	PPB	03 Nov 2021 12:08:32	U	SMPL
M4433-04DUP - 1	PSC-124035DUP	8	0.0389	-	7470A	PPB	03 Nov 2021 12:10:48	U	SMPL
M4433-04MS - 1	PSC-124035MS	13263	3.6833	-	7470A	PPB	03 Nov 2021 12:13:04	U	SMPL
M4433-04MSD - 1	PSC-124035MSD	13722	3.8095	-	7470A	PPB	03 Nov 2021 12:15:21	U	SMPL
PB140435BL - 1	PBW	-81	0.0145	-	7470A	PPB	03 Nov 2021 12:17:38	U	SMPL
PB140435BS - 1	LCSW	13433	3.7301	-	7470A	PPB	03 Nov 2021 12:19:55	U	SMPL
CCV95 - 1	CCV95	18741	5.1895	-	7470A	PPB	03 Nov 2021 12:22:10	U	SMPL
CCB95 - 1	CCB95	-113	0.0057	-	7470A	PPB	03 Nov 2021 12:24:26	U	SMPL
M4434-12 - 1	2229	1	0.0370	-	7470A	PPB	03 Nov 2021 12:26:41	U	SMPL
M4434-12DUP - 1	2229DUP	1	0.0370	-	7470A	PPB	03 Nov 2021 12:28:57	U	SMPL
M4434-12MS - 1	2229MS	13163	3.6558	-	7470A	PPB	03 Nov 2021 12:31:12	U	SMPL
M4434-12MSD - 1	2229MSD	13340	3.7045	-	7470A	PPB	03 Nov 2021 12:33:28	U	SMPL
M4434-13 - 1	2004	-90	0.0120	-	7470A	PPB	03 Nov 2021 12:35:45	U	SMPL
PB140436BL - 1	PBW	7	0.0387	-	7470A	PPB	03 Nov 2021 12:38:02	U	SMPL
PB140436BS - 1	LCSW	13137	3.6487	-	7470A	PPB	03 Nov 2021 12:40:19	U	SMPL
M4434-14 - 1	20-28	346	0.1319	-	7470A	PPB	03 Nov 2021 12:42:36	U	SMPL
M4434-16 - 1	30-31	1869	0.5506	-	7470A	PPB	03 Nov 2021 12:44:53	U	SMPL
CCV96 - 1	CCV96	18268	5.0594	-	7470A	PPB	03 Nov 2021 12:47:09	U	SMPL
CCB96 - 1	CCB96	-58	0.0208	-	7470A	PPB	03 Nov 2021 12:49:25	U	SMPL
M4435-03 - 1	33-34	374	0.1396	-	7470A	PPB	03 Nov 2021 12:51:41	U	SMPL
M4435-03DUP - 1	33-34DUP	319	0.1244	-	7470A	PPB	03 Nov 2021 12:53:56	U	SMPL
M4435-03MS - 1	33-34MS	16241	4.5021	-	7470A	PPB	03 Nov 2021 12:56:12	U	SMPL
M4435-03MSD - 1	33-34MSD	16101	4.4636	-	7470A	PPB	03 Nov 2021 12:58:27	U	SMPL
M4435-07 - 1	39-40	1259	0.3829	-	7470A	PPB	03 Nov 2021 13:00:43	U	SMPL
PB140432TB - 1	PB140369TB	-52	0.0224	-	7470A	PPB	03 Nov 2021 13:02:59	U	SMPL
PB140434TB - 1	PB140415TB	-2	0.0362	-	7470A	PPB	03 Nov 2021 13:05:15	U	SMPL
PB140435TB - 1	PB140416TB	2	0.0373	-	7470A	PPB	03 Nov 2021 13:07:31	U	SMPL
CCV97 - 1	CCV97	17468	4.8395	-	7470A	PPB	03 Nov 2021 13:09:48	U	SMPL
CCB97 - 1	CCB97	-54	0.0219	-	7470A	PPB	03 Nov 2021 13:12:05	U	SMPL
M4411-05LX5 - 1		13	0.0403	-	7470A	PPB	03 Nov 2021 13:14:21	U	SMPL
M4411-05A - 1		13258	3.6820	-	7470A	PPB	03 Nov 2021 13:16:36	U	SMPL
M4433-04LX5 - 1		-17	0.0321	-	7470A	PPB	03 Nov 2021 13:18:52	U	SMPL
M4433-04A - 1		13243	3.6778	-	7470A	PPB	03 Nov 2021 13:21:07	U	SMPL
M4434-12LX5 - 1		-25	0.0299	-	7470A	PPB	03 Nov 2021 13:23:23	U	SMPL
M4434-12A - 1		13480	3.7430	-	7470A	PPB	03 Nov 2021 13:25:38	U	SMPL
M4435-03LX5 - 1		0	0.0367	-	7470A	PPB	03 Nov 2021 13:27:54	U	SMPL
M4435-03A - 1		13598	3.7754	-	7470A	PPB	03 Nov 2021 13:30:10	U	SMPL
CCV98 - 1	CCV98	18374	5.0886	-	7470A	PPB	03 Nov 2021 13:32:26	U	SMPL
CCB98 - 1	CCB98	-87	0.0128	-	7470A	PPB	03 Nov 2021 13:34:42	U	SMPL