



PERCENT SOLID

Supervisor: apatel
Analyst: ketankumar
Date: 11/4/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:25
In Date: 11/03/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 09:30
Out Date: 11/04/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117154

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4450-03	EXB59	1	1.16	8.40	9.56	7.42	74.5	
M4450-04	EXB60	2	1.14	8.48	9.62	8.05	81.5	
M4450-05	EXB61	3	1.16	8.63	9.79	7.89	78.0	
M4450-06	EXB62	4	1.14	8.36	9.5	7.54	76.6	
M4450-07	EXB63	5	1.16	8.45	9.61	7.58	76.0	
M4450-08	EXB64	6	1.14	8.83	9.97	8.47	83.0	
M4450-09	EXB65	7	1.14	8.44	9.58	7.74	78.2	
M4450-10	EXB66	8	1.13	8.82	9.95	8.05	78.5	
M4450-11	EXB67	9	1.16	8.58	9.74	7.54	74.4	
M4450-12	EXB68	10	1.13	8.43	9.56	7.66	77.5	
M4450-13	EXB69	11	1.16	8.41	9.57	7.32	73.2	
M4450-14	EXB76	12	1.17	8.58	9.75	8.09	80.7	
M4450-15	EXB77	13	1.15	8.71	9.86	8.45	83.8	
M4450-16	EXB78	14	1.16	8.78	9.94	8.44	82.9	
M4450-17	EXB79	15	1.14	8.61	9.75	8.27	82.8	
M4450-18	EXB81	16	1.14	8.47	9.61	7.82	78.9	
M4450-19	EXB82	17	1.16	8.49	9.65	7.77	77.9	
M4450-20	EXB83	18	1.14	8.57	9.71	8.00	80.0	
M4450-21	EXB83D	19	1.14	8.57	9.71	8.00	80.0	
M4450-22	EXB83S	20	1.14	8.57	9.71	8.00	80.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$