



PERCENT SOLID

Supervisor: apatel
Analyst: ketankumar
Date: 11/4/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 14:35
In Date: 11/03/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 108
Time OUT: 09:00
Out Date: 11/03/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB117155

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M4451-01	EXB84	1	1.15	8.81	9.96	7.82	75.7	
M4451-02	EXB85	2	1.12	8.65	9.77	7.75	76.6	
M4451-03	EXB86	3	1.16	8.36	9.52	7.51	76.0	
M4451-04	EXB87	4	1.15	8.83	9.98	7.84	75.8	
M4451-05	EXB88	5	1.13	8.68	9.81	7.91	78.1	
M4451-06	EXB89	6	1.15	8.62	9.77	7.83	77.5	
M4451-07	EXB90	7	1.17	8.57	9.74	7.64	75.5	
M4451-08	EXB91	8	1.14	8.82	9.96	7.82	75.7	
M4451-09	EXB92	9	1.13	8.55	9.68	7.51	74.6	
M4451-10	EXB93	10	1.14	8.68	9.82	7.83	77.1	
M4451-11	EXB94	11	1.15	8.51	9.66	7.89	79.2	
M4451-12	EXB95	12	1.15	8.47	9.62	7.59	76.0	
M4451-13	EXB96	13	1.14	8.83	9.97	7.91	76.7	
M4451-14	EXB97	14	1.14	8.51	9.65	7.54	75.2	
M4451-15	EXB98	15	1.13	8.57	9.7	7.82	78.1	
M4451-16	EXB99	16	1.15	8.70	9.85	7.81	76.6	
M4451-17	EXBA0	17	1.15	8.64	9.79	7.84	77.4	
M4451-18	EXBA1	18	1.13	8.82	9.95	8.03	78.2	
M4451-19	EXBA2	19	1.13	8.80	9.93	7.88	76.7	
M4451-20	EXBA3	20	1.15	8.37	9.52	7.6	77.1	
M4451-21	EXBA3D	21	1.15	8.37	9.52	7.6	77.1	
M4451-22	EXBA3S	22	1.15	8.37	9.52	7.6	77.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$