



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117157

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	11/03/2021	11:25
2	CAL2	1	Water	NA	NA	20.3	7.00	11/03/2021	11:26
3	CAL3	1	Water	NA	NA	20.3	10.03	11/03/2021	11:27
4	ICV	1	Water	NA	NA	20.4	7.00	11/03/2021	11:30
5	CCV1	1	Water	NA	NA	20.4	2.01	11/03/2021	11:33
6	M4478-01	1	Water	NA	NA	20.6	6.49	11/03/2021	11:35
7	M4478-01DUP	1	Water	NA	NA	20.7	6.50	11/03/2021	11:36
8	M4479-01	1	Water	NA	NA	20.4	5.77	11/03/2021	11:37
9	CCV2	1	Water	NA	NA	20.4	12.02	11/03/2021	11:40

VB 117157

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M4476 WorkList ID : 153916 Department : Wet-Chemistry Date : 11-03-2021 11:12:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/02/2021	Water	M4478-01	E pH	Cool 4 deg C	loui01	L11	TWP02	11/02/2021	9040C
10/28/2021	Water	M4479-01	A pH	Cool 4 deg C	STER05	D11	BOTTLE-1	10/28/2021	9040C

Date/Time 11/03/21 11:20
 Raw Sample Received by: JO (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 11/03/21 12:30
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: JP (WC)