

LB117163


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

50224



A= 0.0000e+000

B= 1.9955e-004

C= 4.9217e-002

Rho= 0.9999018

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.055	0.055	29	0.000	29					
0.2	0.200	0.210	0.010	808	0.0 %	808					
2.5	2.500	2.449	-0.051	12026	0.0 %	12026					
5.0	5.000	4.984	-0.016	24731	0.0 %	24731					
7.5	7.500	7.430	-0.070	36985	0.0 %	36985					
10.0	10.000	10.072	0.072	50224	0.0 %	50224					

5
-2
0
-1
1

LB117163 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 03 Nov 2021 15:01:49

Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	80	29	-	0.00007471B	PPB	03 Nov 2021 15:06:41	S	Std
0.2 - 1	80.2	808	-	0.20007471B	PPB	03 Nov 2021 15:08:58	S	Std
2.5 - 1	52.5	12026	-	2.50007471B	PPB	03 Nov 2021 15:11:16	S	Std
5.0 - 1	55	24731	-	5.00007471B	PPB	03 Nov 2021 15:13:33	S	Std
7.5 - 1	57.5	36985	-	7.50007471B	PPB	03 Nov 2021 15:15:49	S	Std
10.0 - 1	510	50224	-	10.00007471B	PPB	03 Nov 2021 15:18:06	S	Std
ICV67 - 1	ICV67	19069	3.8545	-7471B	PPB	03 Nov 2021 15:21:04	U	SMPL
ICB67 - 1	ICB67	-68	0.0356	-7471B	PPB	03 Nov 2021 15:23:18	U	SMPL
CCV01 - 1	CCV01	24858	5.0097	-7471B	PPB	03 Nov 2021 15:25:33	U	SMPL
CCB01 - 1	CCB01	10	0.0512	-7471B	PPB	03 Nov 2021 15:27:48	U	SMPL
CRA - 1	CRA	780	0.2049	-7471B	PPB	03 Nov 2021 15:30:04	U	SMPL
HighStd - 1	HighStd	49299	9.8869	-7471B	PPB	03 Nov 2021 15:32:19	U	SMPL
ChkStd - 1	ChkStd	35098	7.0531	-7471B	PPB	03 Nov 2021 15:34:34	U	SMPL
PB140479BL - 1	PBS	-115	0.0263	-7471B	PPB	03 Nov 2021 15:36:52	U	SMPL
PB140479BS - 1	LCSS	19026	3.8459	-7471B	PPB	03 Nov 2021 15:39:09	U	SMPL
M4448-01 - 1	VNJ-224	6438	1.3339	-7471B	PPB	03 Nov 2021 15:41:25	U	SMPL
M4448-01DUP - 1	VNJ-224DUP	5234	1.0937	-7471B	PPB	03 Nov 2021 15:43:41	U	SMPL
M4448-01MS - 1	VNJ-224MS	23662	4.7710	-7471B	PPB	03 Nov 2021 15:45:58	U	SMPL
M4448-01MSD - 1	VNJ-224MSD	26799	5.3970	-7471B	PPB	03 Nov 2021 15:48:15	U	SMPL
M4449-01 - 1	EXTERIOR-WALLS	610	0.1709	-7471B	PPB	03 Nov 2021 15:50:30	U	SMPL
CCV02 - 1	CCV02	24750	4.9881	-7471B	PPB	03 Nov 2021 15:52:46	U	SMPL
CCB02 - 1	CCB02	-51	0.0390	-7471B	PPB	03 Nov 2021 15:55:01	U	SMPL
M4449-02 - 1	CONCRETE-WALLS	1847	0.4178	-7471B	PPB	03 Nov 2021 15:57:16	U	SMPL
M4449-03 - 1	PAINTED-WALLS	1739	0.3962	-7471B	PPB	03 Nov 2021 15:59:32	U	SMPL
M4449-04 - 1	FLOORS	30796	6.1946	-7471B	PPB	03 Nov 2021 16:01:47	U	SMPL
M4454-01 - 1	30860-OIL	34	0.0560	-7471B	PPB	03 Nov 2021 16:04:03	U	SMPL
M4454-03 - 1	30859	459	0.1408	-7471B	PPB	03 Nov 2021 16:06:20	U	SMPL
M4454-04 - 1	30858	2688	0.5856	-7471B	PPB	03 Nov 2021 16:08:36	U	SMPL
M4454-05 - 1	11221-A	45	0.0582	-7471B	PPB	03 Nov 2021 16:10:52	U	SMPL
M4454-06 - 1	33104	334	0.1159	-7471B	PPB	03 Nov 2021 16:13:09	U	SMPL
M4470-01 - 1	SP-3	12421	2.5279	-7471B	PPB	03 Nov 2021 16:15:26	U	SMPL
M4470-04 - 1	SP-2	13428	2.7288	-7471B	PPB	03 Nov 2021 16:17:42	U	SMPL
CCV03 - 1	CCV03	24707	4.9796	-7471B	PPB	03 Nov 2021 16:19:57	U	SMPL
CCB03 - 1	CCB03	-50	0.0392	-7471B	PPB	03 Nov 2021 16:22:13	U	SMPL
M4472-01 - 1	WC-2-JS	435	0.1360	-7471B	PPB	03 Nov 2021 16:26:44	U	SMPL
M4472-04 - 1	WC-2-KS	1468	0.3422	-7471B	PPB	03 Nov 2021 16:29:00	U	SMPL
M4470-07 - 1	TP-1	9707	1.9863	-7471B	PPB	03 Nov 2021 16:35:41	U	SMPL
M4472-07 - 1	WC-2-KD	468	0.1426	-7471B	PPB	03 Nov 2021 16:37:57	U	SMPL
M4448-01LX5 - 1		1121	0.2729	-7471B	PPB	03 Nov 2021 16:40:14	U	SMPL
M4448-01A - 1		25386	5.1150	-7471B	PPB	03 Nov 2021 16:42:31	U	SMPL
CCV04 - 1	CCV04	24852	5.0085	-7471B	PPB	03 Nov 2021 16:44:48	U	SMPL
CCB04 - 1	CCB04	-55	0.0382	-7471B	PPB	03 Nov 2021 16:47:04	U	SMPL