



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB117166
Thermometer ID: flashpoint

Reviewed By: ketankumar
Supervisor Review By: apatel
Ambient Barometric Pressure (mmHg): 760.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	11/03/2021	13:20
2	M4454-07		1	10	100.00	>212.0	>212.0	11/03/2021	13:45
3	M4454-07DUP		1	10	100.00	>212.0	>212.0	11/03/2021	14:15
4	M4454-09		1	15	100.00	>212.0	>212.0	11/03/2021	14:40
5	M4454-11		1	12	100.00	>212.0	>212.0	11/03/2021	15:10

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$

WORKLIST(Hardcopy Internal Chain)

LB 117166

WorkList Name : fp 11321 WorkList ID : 153907 Department : Wet-Chemistry Date : 11-03-2021 09:08:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/05/2021	Water	M4454-07	Flash Point	Cool 4 deg C	PSEG03	L11	30860	11/02/2021	1010B
11/05/2021	Water	M4454-09	Flash Point	Cool 4 deg C	PSEG03	L11	31073	11/02/2021	1010B
11/05/2021	Water	M4454-11	Flash Point	Cool 4 deg C	PSEG03	L11	11221-B	11/02/2021	1010B

Date/Time 11/03/21 13:00
 Raw Sample Received by: FS (WC)
 Raw Sample Relinquished by: FS (WC)

Date/Time 11/03/21 15:15
 Raw Sample Received by: FS (WC)
 Raw Sample Relinquished by: FS (WC)