



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB117167

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	11/03/2021	13:20
2	M4386-06		1	9	60.00	140.0	140.0	11/03/2021	15:40
3	M4454-01		1	15	100.00	>212.0	>212.0	11/03/2021	16:05
4	M4454-01DUP		1	15	100.00	>212.0	>212.0	11/03/2021	16:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB117167

WorkList Name : fp11321-1 WorkList ID : 153908 Department : Wet-Chemistry Date : 11-03-2021 09:09:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/05/2021	Solid	M4454-01	Flash Point	Cool 4 deg C	PSEG03	L11	30860-OIL	11/02/2021	1010A

Date/Time 11/03/21 13:00
Raw Sample Received by: KS (wc)
Raw Sample Relinquished by: SP (wc)

Date/Time 11/03/21 16:40
Raw Sample Received by: SP (wc)
Raw Sample Relinquished by: KS (wc)