



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB117168

Slope : 98.6

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.8	4.01	11/03/2021	17:15
2	CAL2	1	Water	NA	NA	21.6	7.00	11/03/2021	17:16
3	CAL3	1	Water	NA	NA	21.4	10.02	11/03/2021	17:17
4	ICV	1	Water	NA	NA	21.7	7.00	11/03/2021	17:18
5	CCV1	1	Water	NA	NA	21.5	2.01	11/03/2021	17:19
6	M4499-01	1	Solid	20	20	20.3	11.74	11/03/2021	17:20
7	M4499-01DUP	1	Solid	20	20	20.3	11.73	11/03/2021	17:21
8	CCV2	1	Water	NA	NA	21.4	12.02	11/03/2021	17:22

WORKLIST(Hardcopy Internal Chain)

LB117168

WorkList Name : PH11321 WorkList ID : 153944 Department : Wet-Chemistry Date : 11-03-2021 17:05:59

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/03/2021	Solid	M4499-01	pH	Cool 4 deg C	TULL02	D11	WEL-BRIDGE	11/03/2021	9045D

Date/Time 11/03/21 17:10
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)

Date/Time 11/03/21 17:25
 Raw Sample Received by: KS (WC)
 Raw Sample Relinquished by: KS (WC)