



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB117172

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2810
buffer solution pH 7 yellow	W2813
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.8	4.01	11/03/2021	17:15
2	CAL2	1	Water	NA	NA	21.6	7.00	11/03/2021	17:16
3	CAL3	1	Water	NA	NA	21.4	10.02	11/03/2021	17:17
4	ICV	1	Water	NA	NA	21.7	7.00	11/03/2021	17:18
5	CCV1	1	Water	NA	NA	21.5	2.01	11/03/2021	17:19
6	M4480-02	1	Water	NA	NA	20.8	8.44	11/03/2021	17:41
7	M4480-02DUP	1	Water	NA	NA	20.8	8.43	11/03/2021	17:42
8	M4480-03	1	Water	NA	NA	20.1	4.55	11/03/2021	17:43
9	CCV2	1	Water	NA	NA	21.6	12.02	11/03/2021	17:45

WORKLIST(Hardcopy Internal Chain)

NR 11/1/22

WorkList Name : CORROSIVITY M4480 V WorkList ID : 153945 Department : Wet-Chemistry Date : 11-03-2021 17:08:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/03/2021	Water	M4480-02	Corrosivity	Cool 4 deg C	PSEG03	L11	1102	11/03/2021	9040C
11/03/2021	Water	M4480-03	Corrosivity	Cool 4 deg C	PSEG03	L11	1106	11/03/2021	9040C

Date/Time 11/03/21 17:10
 Raw Sample Received by: JO (WC)
 Raw Sample Relinquished by: RC (Dig. Lab)

Date/Time 11/09/21 18:30
 Raw Sample Received by: RC (Dig. Lab)
 Raw Sample Relinquished by: JO (WC)