


LB117195

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.7930e-004

C= 7.8041e-002

Rho= 0.9995575

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.082	0.082	14	0.000	14					19
0.2	0.200	0.237	0.037	569	0.0 %	569					-6
2.5	2.500	2.346	-0.154	8121	0.0 %	8121					-2
5.0	5.000	4.921	-0.079	17341	0.0 %	17341					2
7.5	7.500	7.675	0.175	27200	0.0 %	27200					-1
10.0	10.000	9.939	-0.061	35305	0.0 %	35305					

LB117175 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 04 Nov 2021 11:07:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	14	-	0.0000	7470A	PPB	04 Nov 2021 11:12:40	S	Std
0.2 - 1	50.2	569	-	0.2000	7470A	PPB	04 Nov 2021 11:14:57	S	Std
2.5 - 1	52.5	8121	-	2.5000	7470A	PPB	04 Nov 2021 11:17:14	S	Std
5.0 - 1	55	17341	-	5.0000	7470A	PPB	04 Nov 2021 11:19:31	S	Std
7.5 - 1	57.5	27200	-	7.5000	7470A	PPB	04 Nov 2021 11:24:21	S	Std
10.0 - 1	510	35305	-	10.0000	7470A	PPB	04 Nov 2021 11:29:12	S	Std
ICV68 - 1	ICV68	14233	4.0533	-	7470A	PPB	04 Nov 2021 11:32:27	U	SMPL
ICB68 - 1	ICB68	-98	0.0507	-	7470A	PPB	04 Nov 2021 11:34:42	U	SMPL
CCV05 - 1	CCV05	17747	5.0347	-	7470A	PPB	04 Nov 2021 11:36:57	U	SMPL
CCB05 - 1	CCB05	-106	0.0484	-	7470A	PPB	04 Nov 2021 11:39:12	U	SMPL
CRA - 1	CRA	547	0.2308	-	7470A	PPB	04 Nov 2021 11:41:27	U	SMPL
HighStd - 1	HighStd	36150	10.1746	-	7470A	PPB	04 Nov 2021 11:43:42	U	SMPL
ChkStd - 1	ChkStd	25236	7.1264	-	7470A	PPB	04 Nov 2021 11:45:57	U	SMPL
PB140480BL - 1	PBW	-137	0.0398	-	7470A	PPB	04 Nov 2021 11:48:14	U	SMPL
PB140480BS - 1	LCSW	14008	3.9904	-	7470A	PPB	04 Nov 2021 11:53:14	U	SMPL
M4470-02 - 1	SP-3	-112	0.0468	-	7470A	PPB	04 Nov 2021 11:55:30	U	SMPL
M4470-05 - 1	SP-2	16	0.0825	-	7470A	PPB	04 Nov 2021 11:57:47	U	SMPL
M4470-08 - 1	TP-1	53	0.0928	-	7470A	PPB	04 Nov 2021 12:00:03	U	SMPL
M4472-02 - 1	WC-2-JS	3	0.0789	-	7470A	PPB	04 Nov 2021 12:08:34	U	SMPL
M4472-05 - 1	WC-2-KS	16	0.0825	-	7470A	PPB	04 Nov 2021 12:10:50	U	SMPL
CCV06 - 1	CCV06	17368	4.9289	-	7470A	PPB	04 Nov 2021 12:13:05	U	SMPL
CCB06 - 1	CCB06	-109	0.0476	-	7470A	PPB	04 Nov 2021 12:15:20	U	SMPL
M4472-08 - 1	WC-2-KD	-6	0.0764	-	7470A	PPB	04 Nov 2021 12:17:35	U	SMPL
M4472-08DUP - 1	WC-2-KDDUP	5	0.0794	-	7470A	PPB	04 Nov 2021 12:19:50	U	SMPL
M4472-08MS - 1	WC-2-KDMS	14599	4.1555	-	7470A	PPB	04 Nov 2021 12:22:05	U	SMPL
M4472-08MSD - 1	WC-2-KDMSD	14502	4.1284	-	7470A	PPB	04 Nov 2021 12:24:21	U	SMPL
PB140480TB - 1	PB140445TB	-85	0.0543	-	7470A	PPB	04 Nov 2021 12:26:38	U	SMPL
M4472-08LX5 - 1		-22	0.0719	-	7470A	PPB	04 Nov 2021 12:28:54	U	SMPL
M4472-08A - 1		13939	3.9712	-	7470A	PPB	04 Nov 2021 12:31:10	U	SMPL
CCV07 - 1	CCV07	17519	4.9710	-	7470A	PPB	04 Nov 2021 12:33:27	U	SMPL
CCB07 - 1	CCB07	-132	0.0412	-	7470A	PPB	04 Nov 2021 12:35:44	U	SMPL