



Analytical Summary Report

Analysis Method: SM2120 B

Reviewed By: apatel

Parameter: Color

Supervisor Review By: ketankumar

Run Number: LB117177

pH Meter ID: WC pH Meter-1

Reagent/Standard	Lot/Log #
Calibration std, 0 color unit	WP93724
Calibration std, 5 color unit	WP93725
Calibration std, 10 color unit	WP93726
Calibration std, 15 color unit	WP93727
Calibration std, 20 color unit	WP93728
Calibration std, 25 color unit	WP93729
Calibration std, 30 color unit	WP93730
Calibration std, 40 color unit	WP93731
Calibration std, 50 color unit	WP93732
Calibration std, 60 color unit	WP93733
Calibration std, 70 color unit	WP93734

Seq	LabID	ClientID	DL	pH Reading	pH Adjustment	Result (cu)	Anal Date	Anal Time
1	LB117177BL	LB117177BL	1	5.15	N/A	0.00	11/04/2021	09:30
2	M4482-01	MONTHLY-OUTFALL-002	20	8.42	N/A	50.00	11/04/2021	09:35
3	M4482-01DUP	MONTHLY-OUTFALL-002	20	8.42	N/A	50.00	11/04/2021	09:40

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Color-110321 WorkList ID : 153923 Department : Wet-Chemistry Date : 11-03-2021 12:01:02

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
11/05/2021	Water	M4482-01	F Color	Cool 4 deg C	INFR01	D11	MONTHLY-OUTFALL-002	11/03/2021	SM2120 B

Date/Time 11/4/21 9:00
 Raw Sample Received by: AP (W.C.)
 Raw Sample Relinquished by: AP (W.C.)

Date/Time 11/4/21 10:30
 Raw Sample Received by: AP (W.C.)
 Raw Sample Relinquished by: AP (W.C.)