


LB117178

7471B



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/01
0.0	0.000	0.010	0.010	67	0.000	67					-
0.2	0.200	0.214	0.014	1078	0.0 %	1078					7
2.5	2.500	2.470	-0.030	12287	0.0 %	12287					-1
5.0	5.000	5.076	0.076	25234	0.0 %	25234					2
7.5	7.500	7.359	-0.141	36578	0.0 %	36578					-2
10.0	10.000	10.072	0.072	50058	0.0 %	50058					1

LB117178 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 04 Nov 2021 14:28:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	67	-	0.00007471B	PPB		04 Nov 2021 14:31:36	S	Std
0.2 - 1	50.2	1078	-	0.20007471B	PPB		04 Nov 2021 14:33:53	S	Std
2.5 - 1	52.5	12287	-	2.50007471B	PPB		04 Nov 2021 14:36:10	S	Std
5.0 - 1	55	25234	-	5.00007471B	PPB		04 Nov 2021 14:38:27	S	Std
7.5 - 1	57.5	36578	-	7.50007471B	PPB		04 Nov 2021 14:40:43	S	Std
10.0 - 1	510	50058	-	10.00007471B	PPB		04 Nov 2021 14:43:00	S	Std
ICV69 - 1	ICV69	20023	4.0268	-7471B	PPB		04 Nov 2021 14:45:42	U	SMPL
ICB69 - 1	ICB69	-25	-0.0084	-7471B	PPB		04 Nov 2021 14:47:56	U	SMPL
CCV08 - 1	CCV08	25048	5.0382	-7471B	PPB		04 Nov 2021 14:50:12	U	SMPL
CCB08 - 1	CCB08	-16	-0.0066	-7471B	PPB		04 Nov 2021 14:52:27	U	SMPL
CRA - 1	CRA	1061	0.2102	-7471B	PPB		04 Nov 2021 14:54:42	U	SMPL
HighStd - 1	HighStd	49703	10.0007	-7471B	PPB		04 Nov 2021 14:56:57	U	SMPL
ChkStd - 1	ChkStd	34394	6.9193	-7471B	PPB		04 Nov 2021 14:59:12	U	SMPL
PB140511BL - 1	PBS	-100	-0.0235	-7471B	PPB		04 Nov 2021 15:01:30	U	SMPL
PB140511BS - 1	LCSS	20037	4.0296	-7471B	PPB		04 Nov 2021 15:03:47	U	SMPL
M4475-01 - 1	SB01A	5426	1.0887	-7471B	PPB		04 Nov 2021 15:06:04	U	SMPL
M4475-02 - 1	SB01B	567	0.1107	-7471B	PPB		04 Nov 2021 15:08:20	U	SMPL
M4475-03 - 1	SB02A	4445	0.8913	-7471B	PPB		04 Nov 2021 15:10:37	U	SMPL
M4475-04 - 1	SB02B	593	0.1160	-7471B	PPB		04 Nov 2021 15:12:55	U	SMPL
M4475-05 - 1	SB03	6162	1.2369	-7471B	PPB		04 Nov 2021 15:15:10	U	SMPL
CCV09 - 1	CCV09	24859	5.0001	-7471B	PPB		04 Nov 2021 15:17:25	U	SMPL
CCB09 - 1	CCB09	-33	-0.0100	-7471B	PPB		04 Nov 2021 15:19:40	U	SMPL
M4475-06 - 1	DUP01	754	0.1484	-7471B	PPB		04 Nov 2021 15:21:56	U	SMPL
M4497-01 - 1	RO-COMP-1	972	0.1922	-7471B	PPB		04 Nov 2021 15:24:12	U	SMPL
M4497-04 - 1	VNJ-203	1429	0.2842	-7471B	PPB		04 Nov 2021 15:26:27	U	SMPL
M4497-07 - 1	346	1167	0.2315	-7471B	PPB		04 Nov 2021 15:28:43	U	SMPL
M4498-01 - 1	WC-2-MS	4763	0.9553	-7471B	PPB		04 Nov 2021 15:30:58	U	SMPL
M4498-04 - 1	WC-2-MD	7018	1.4092	-7471B	PPB		04 Nov 2021 15:33:15	U	SMPL
M4498-07 - 1	WC-2-LS	1849	0.3688	-7471B	PPB		04 Nov 2021 15:35:31	U	SMPL
M4500-01 - 1	TR-4-110321	892	0.1761	-7471B	PPB		04 Nov 2021 15:37:48	U	SMPL
M4500-03 - 1	TR-8-110321	920	0.1818	-7471B	PPB		04 Nov 2021 15:40:05	U	SMPL
M4500-03DUP - 1	TR-8-110321DUP	887	0.1751	-7471B	PPB		04 Nov 2021 15:42:21	U	SMPL
CCV10 - 1	CCV10	25313	5.0915	-7471B	PPB		04 Nov 2021 15:44:36	U	SMPL
CCB10 - 1	CCB10	-27	-0.0088	-7471B	PPB		04 Nov 2021 15:46:52	U	SMPL
M4500-03MS - 1	TR-8-110321MS	21203	4.2643	-7471B	PPB		04 Nov 2021 15:49:08	U	SMPL
M4500-03MSD - 1	TR-8-110321MSD	21518	4.3277	-7471B	PPB		04 Nov 2021 15:51:24	U	SMPL
M4501-01 - 1	72-11929	27297	5.4909	-7471B	PPB		04 Nov 2021 15:53:40	U	SMPL
M4500-03LX5 - 1		92	0.0151	-7471B	PPB		04 Nov 2021 15:55:55	U	SMPL
M4500-03A - 1		21308	4.2854	-7471B	PPB		04 Nov 2021 15:58:12	U	SMPL
CCV11 - 1	CCV11	25030	5.0346	-7471B	PPB		04 Nov 2021 16:00:28	U	SMPL
CCB11 - 1	CCB11	-41	-0.0116	-7471B	PPB		04 Nov 2021 16:02:45	U	SMPL