



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 11/8/2021

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:30  
In Date: 11/05/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:25  
Out Date: 11/06/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB117180

| Lab ID   | Client SampleID    | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|--------------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------|
| M4477-01 | T1-EP01-N          | 1      | 1.14            | 8.67          | 9.81                     | 9.36                       | 94.8    |          |
| M4477-02 | T1-EP02-NE         | 2      | 1.12            | 8.56          | 9.68                     | 9.33                       | 95.9    |          |
| M4477-03 | T1-EP03-S          | 3      | 1.15            | 8.49          | 9.64                     | 9.21                       | 94.9    |          |
| M4477-04 | T1-B1-S-N          | 4      | 1.1             | 8.51          | 9.61                     | 9.1                        | 94.0    |          |
| M4477-05 | T1-B2-NW           | 5      | 1.13            | 8.73          | 9.86                     | 9.34                       | 94.0    |          |
| M4477-06 | T5-EP1-S           | 6      | 1.15            | 8.81          | 9.96                     | 9.42                       | 93.9    |          |
| M4477-07 | T5-EP2-NW          | 7      | 1.13            | 8.44          | 9.57                     | 8.76                       | 90.4    |          |
| M4477-08 | T1-B3-N            | 8      | 1.13            | 8.73          | 9.86                     | 9.44                       | 95.2    |          |
| M4477-09 | T5-EP3-N           | 9      | 1.16            | 8.44          | 9.6                      | 8.98                       | 92.7    |          |
| M4477-10 | T5-B1-NE           | 10     | 1.13            | 8.55          | 9.68                     | 9.06                       | 92.7    |          |
| M4477-11 | T5-B2-NW           | 11     | 1.11            | 8.77          | 9.88                     | 9.37                       | 94.2    |          |
| M4485-01 | AG-2               | 12     | 1.14            | 8.81          | 9.95                     | 9.47                       | 94.6    |          |
| M4485-02 | AG-3               | 13     | 1.13            | 8.73          | 9.86                     | 9.23                       | 92.8    |          |
| M4485-03 | AG-4               | 14     | 1.12            | 8.62          | 9.74                     | 9.26                       | 94.4    |          |
| M4485-04 | AG-5               | 15     | 1.15            | 8.58          | 9.73                     | 8.98                       | 91.3    |          |
| M4485-05 | AG-6               | 16     | 1.18            | 8.79          | 9.97                     | 9.16                       | 90.8    |          |
| M4485-06 | AG-7               | 17     | 1.13            | 8.68          | 9.81                     | 9.13                       | 92.2    |          |
| M4485-07 | AG-8               | 18     | 1.18            | 8.42          | 9.6                      | 9.03                       | 93.2    |          |
| M4485-08 | AG-9               | 19     | 1.18            | 8.46          | 9.64                     | 8.66                       | 88.4    |          |
| M4485-09 | AG-10              | 20     | 1.18            | 8.45          | 9.63                     | 8.87                       | 91.0    |          |
| M4497-07 | 346                | 21     | 1.14            | 8.39          | 9.53                     | 8.6                        | 88.9    |          |
| M4519-01 | SB-1               | 22     | 1.14            | 8.56          | 9.7                      | 9.06                       | 92.5    |          |
| M4519-02 | SB-2               | 23     | 1.16            | 8.79          | 9.95                     | 8.44                       | 82.8    |          |
| M4519-03 | SB-3               | 24     | 1.16            | 8.34          | 9.5                      | 8.83                       | 92.0    |          |
| M4519-04 | SB-4               | 25     | 1.16            | 8.77          | 9.93                     | 9.04                       | 89.9    |          |
| M4519-05 | SB-5               | 26     | 1.15            | 8.69          | 9.84                     | 9.06                       | 91.0    |          |
| M4520-01 | 40540              | 27     | 1.15            | 8.60          | 9.75                     | 8.93                       | 90.5    |          |
| M4523-01 | VWE-B4-110421-0-10 | 28     | 1.15            | 8.36          | 9.51                     | 8.82                       | 91.7    |          |



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 11/8/2021

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:30  
In Date: 11/05/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:25  
Out Date: 11/06/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB117180

| Lab ID   | Client SampleID     | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments    |
|----------|---------------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|-------------|
| M4523-02 | VWE-B4-110421-10-20 | 29     | 1.15            | 8.15          | 9.3                      | 8.32                       | 88.0    |             |
| M4523-03 | VWE-B4-110421-20-30 | 30     | 1.15            | 8.54          | 9.69                     | 8.45                       | 85.5    |             |
| M4523-04 | VWE-B4-110421-0-1   | 31     | 1.15            | 8.62          | 9.77                     | 9.08                       | 92.0    |             |
| M4523-05 | VWE-B4-110421-14-15 | 32     | 1.15            | 8.70          | 9.85                     | 8.89                       | 89.0    |             |
| M4523-06 | VWE-B4-110421-29-30 | 33     | 1.15            | 8.45          | 9.6                      | 8.87                       | 91.4    |             |
| M4523-09 | VWE-B4-110421-BD    | 34     | 1.15            | 8.65          | 9.8                      | 8.43                       | 84.2    |             |
| M4523-10 | VWE-B4-110421-BD    | 35     | 1.15            | 8.82          | 9.97                     | 9.24                       | 91.7    |             |
| M4525-01 | SEC-2021-26A        | 36     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-02 | SEC-2021-26B        | 37     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-03 | SEC-2021-27A        | 38     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-04 | SEC-2021-27B        | 39     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-05 | SEC-2021-28A        | 40     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-06 | SEC-2021-28B        | 41     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-07 | SEC-2021-29A        | 42     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-08 | SEC-2021-29B        | 43     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-09 | SEC-2021-30A        | 44     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-10 | SEC-2021-30B        | 45     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-11 | SEC-2021-31A        | 46     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-12 | SEC-2021-31B        | 47     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-13 | SEC-2021-32A        | 48     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4525-14 | SEC-2021-32B        | 49     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4529-01 | 6-INCH-STEEL        | 50     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | WIPE SAMPLE |
| M4529-02 | 12-INCH-STEEL       | 51     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | WIPE SAMPLE |
| M4536-01 | SEC-2021-33A        | 52     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4536-02 | SEC-2021-33B        | 53     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4536-03 | SEC-2021-34A        | 54     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4536-04 | SEC-2021-34B        | 55     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4536-05 | SEC-2021-35A        | 56     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 11/8/2021

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:30  
In Date: 11/05/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:25  
Out Date: 11/06/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB117180

| Lab ID   | Client SampleID | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments  |
|----------|-----------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|-----------|
| M4536-06 | SEC-2021-35B    | 57     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL |
| M4536-07 | SEC-2021-36B    | 58     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL |
| M4536-08 | SEC-2021-37A    | 59     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL |
| M4536-09 | SEC-2021-38B    | 60     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL |
| M4536-10 | SEC-2021-39A    | 61     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL |
| M4536-11 | SEC-2021-39B    | 62     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL |
| M4537-01 | S1-S2-S3        | 81     | 1.12            | 8.41          | 9.53                     | 8.16                       | 83.7    |           |
| M4538-01 | WC-C1-COMP      | 63     | 1.15            | 8.60          | 9.75                     | 8.76                       | 88.5    |           |
| M4539-01 | WC-C2-COMP      | 64     | 1.15            | 8.66          | 9.81                     | 8.84                       | 88.8    |           |
| M4539-03 | M4539-01MS      | 65     | 1.15            | 8.66          | 9.81                     | 8.84                       | 88.8    |           |
| M4539-04 | M4539-01MSD     | 66     | 1.15            | 8.66          | 9.81                     | 8.84                       | 88.8    |           |
| M4540-01 | TP-1            | 67     | 1.15            | 8.39          | 9.54                     | 8.48                       | 87.4    |           |
| M4540-02 | TP-1-EPH-2      | 68     | 1.15            | 8.45          | 9.6                      | 8.84                       | 91.0    |           |
| M4540-03 | TP-2            | 69     | 1.15            | 8.46          | 9.61                     | 8.29                       | 84.4    |           |
| M4540-04 | TP-2-EPH-2      | 70     | 1.15            | 8.61          | 9.76                     | 8.4                        | 84.2    |           |
| M4540-05 | TP-3            | 71     | 1.15            | 8.70          | 9.85                     | 8.59                       | 85.5    |           |
| M4540-06 | TP-3-EPH-2      | 72     | 1.15            | 8.37          | 9.52                     | 7.8                        | 79.5    |           |
| M4540-07 | TP-4            | 73     | 1.15            | 8.45          | 9.6                      | 8.65                       | 88.8    |           |
| M4540-08 | TP-4-EPH-2      | 74     | 1.15            | 8.59          | 9.74                     | 8.79                       | 88.9    |           |
| M4540-09 | TP-5            | 75     | 1.15            | 8.45          | 9.6                      | 8.67                       | 89.0    |           |
| M4540-10 | TP-5-EPH-2      | 76     | 1.15            | 8.79          | 9.94                     | 8.91                       | 88.3    |           |
| M4540-11 | TP-6            | 77     | 1.15            | 8.70          | 9.85                     | 8.09                       | 79.8    |           |
| M4540-12 | TP-6-EPH-2      | 78     | 1.15            | 8.72          | 9.87                     | 8.19                       | 80.7    |           |
| M4541-01 | WC-3-AS         | 79     | 1.15            | 8.73          | 9.88                     | 9.06                       | 90.6    |           |
| M4541-03 | WC-3-AS-EPH-2   | 80     | 1.15            | 8.72          | 9.87                     | 8.74                       | 87.0    |           |
| M4544-01 | MOO-2021-10A    | 98     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC      |
| M4544-02 | MOO-2021-10B    | 99     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL |
| M4544-03 | MOO-2021-10C    | 100    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL |



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 11/8/2021

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:30  
In Date: 11/05/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:25  
Out Date: 11/06/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB117180

| Lab ID   | Client SampleID    | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments    |
|----------|--------------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|-------------|
| M4544-04 | MOO-2021-11A       | 101    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4544-05 | MOO-2021-11B       | 102    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4544-06 | MOO-2021-11C       | 103    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4544-07 | MOO-2021-11D       | 104    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4548-01 | LAW-2021-1A        | 105    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4548-02 | LAW-2021-2A        | 106    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4548-03 | LAW-2021-2B        | 107    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4548-04 | LAW-2021-3A        | 108    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4548-05 | LAW-2021-3B        | 109    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4561-01 | ETGI-280           | 82     | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | WOOD SAMPLE |
| M4562-01 | T-1                | 83     | 1.18            | 8.50          | 9.68                     | 7.92                       | 79.3    |             |
| M4562-02 | T-1A               | 84     | 1.16            | 8.56          | 9.72                     | 8.18                       | 82.0    |             |
| M4563-01 | S-1                | 85     | 1.13            | 8.55          | 9.68                     | 8.37                       | 84.7    |             |
| M4569-01 | CL-02-110521       | 86     | 1.14            | 8.80          | 9.94                     | 9.23                       | 91.9    |             |
| M4569-02 | CL-02-110521-EPH-2 | 87     | 1.15            | 8.80          | 9.95                     | 9.13                       | 90.7    |             |
| M4570-01 | SOM-2021-1A        | 110    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-02 | SOM-2021-1B        | 111    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-03 | SOM-2021-2A        | 112    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-04 | SOM-2021-2B        | 113    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-05 | SOM-2021-3A        | 114    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-06 | SOM-2021-3B        | 115    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-07 | SOM-2021-4A        | 116    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-08 | SOM-2021-5A        | 117    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-09 | SOM-2021-5B        | 118    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-10 | SOM-2021-6A        | 119    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-11 | SOM-2021-6B        | 120    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-12 | KMA79SC-A          | 121    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |
| M4570-13 | KMA79SC-B          | 122    | 1.00            | 1.00          | 2.00                     | 2.00                       | 100.0   | PILC REEL   |



PERCENT SOLID

Supervisor: apatel  
Analyst: jignesh  
Date: 11/8/2021

OVENTEMP IN Celsius(°C): 107  
Time IN: 17:30  
In Date: 11/05/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103  
Time OUT: 08:25  
Out Date: 11/06/2021  
Weight Check 1.0g: 1.00  
Weight Check 10g: 10.00  
BalanceID: M SC-4  
Thermometer ID: %SOLIDS-OVEN

QC:LB117180

| Lab ID   | Client SampleID     | Dish # | Dish Wt (g) (A) | Sample Wt (g) | Dish + Sample Wt (g) (B) | Dish+Dry Sample Wt (g) (C) | % Solid | Comments |
|----------|---------------------|--------|-----------------|---------------|--------------------------|----------------------------|---------|----------|
| M4574-01 | VWE-B6-110521-0-10  | 88     | 1.12            | 8.76          | 9.88                     | 9.06                       | 90.6    |          |
| M4574-02 | VWE-B6-110521-10-20 | 89     | 1.1             | 8.69          | 9.79                     | 9.22                       | 93.4    |          |
| M4574-03 | VWE-B6-110521-20-30 | 90     | 1.17            | 8.73          | 9.9                      | 9.45                       | 94.8    |          |
| M4574-04 | VWE-B6-110521-0-1   | 91     | 1.14            | 8.59          | 9.73                     | 9.21                       | 93.9    |          |
| M4574-05 | VWE-B6-110521-14-15 | 92     | 1.13            | 8.50          | 9.63                     | 8.75                       | 89.6    |          |
| M4574-06 | VWE-B6-110521-29-30 | 93     | 1.13            | 8.61          | 9.74                     | 9.23                       | 94.1    |          |
| M4576-01 | SB-1- (12-12.5)     | 94     | 1.13            | 8.63          | 9.76                     | 8.98                       | 91.0    |          |
| M4576-03 | SB-2- (9-9.5)       | 95     | 1.16            | 8.44          | 9.6                      | 9.44                       | 98.1    |          |
| M4576-04 | DW-2                | 96     | 1.16            | 8.81          | 9.97                     | 7.78                       | 75.1    |          |
| M4576-05 | DW-3                | 97     | 1.18            | 8.60          | 9.78                     | 5.81                       | 53.8    |          |

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

# WORKLIST(Hardcopy Internal Chain)

W3117180

WorkList Name : %1-110521      WorkList ID : 153980      Department : Wet-Chemistry      Date : 11-05-2021 08:08:50

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|-----------------|--------------|--------------|
| 11/10/2021 | Solid  | M4477-01 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T1-EP01-N       | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4477-02 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T1-EP02-NE      | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4477-03 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T1-EP03-S       | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4477-04 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T1-B1-S-N       | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4477-05 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T1-B2-NW        | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4477-06 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T5-EP1-S        | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4477-07 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T5-EP2-NW       | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4477-08 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T1-B3-N         | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4477-09 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T5-EP3-N        | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4477-10 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T5-B1-NE        | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4477-11 | Percent Solids | Cool 4 deg C | LIRO01   | D11                         | T5-B2-NW        | 11/02/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4485-01 | Percent Solids | Cool 4 deg C | ANTE04   | L11                         | AG-2            | 11/01/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4485-02 | Percent Solids | Cool 4 deg C | ANTE04   | L11                         | AG-3            | 11/01/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4485-03 | Percent Solids | Cool 4 deg C | ANTE04   | L11                         | AG-4            | 11/01/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4485-04 | Percent Solids | Cool 4 deg C | ANTE04   | L11                         | AG-5            | 11/02/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4485-05 | Percent Solids | Cool 4 deg C | ANTE04   | L11                         | AG-6            | 11/02/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4485-06 | Percent Solids | Cool 4 deg C | ANTE04   | L11                         | AG-7            | 11/02/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4485-07 | Percent Solids | Cool 4 deg C | ANTE04   | L11                         | AG-8            | 11/02/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4485-08 | Percent Solids | Cool 4 deg C | ANTE04   | L11                         | AG-9            | 11/02/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4485-09 | Percent Solids | Cool 4 deg C | ANTE04   | L11                         | AG-10           | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4497-07 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | 346             | 11/03/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4519-01 | Percent Solids | Cool 4 deg C | POWE02   | D11                         | SB-1            | 11/02/2021   | Chemtech -SO |

Date/Time 11/05/21 15:40  
 Raw Sample Received by: Jo (WC)  
 Raw Sample Relinquished by: KS (WC)

Date/Time 11/05/21 17:35  
 Raw Sample Received by: KS (WC)  
 Raw Sample Relinquished by: Jo (WC)

# WORKLIST(Hardcopy Internal Chain)

18117180

WorkList Name : %1-110521      WorkList ID : 153980      Department : Wet-Chemistry      Date : 11-05-2021 08:08:50

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample     | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|---------------------|--------------|--------------|
| 11/11/2021 | Solid  | M4519-02 | Percent Solids | Cool 4 deg C | POWE02   | D11                         | SB-2                | 11/02/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4519-03 | Percent Solids | Cool 4 deg C | POWE02   | D11                         | SB-3                | 11/02/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4519-04 | Percent Solids | Cool 4 deg C | POWE02   | D11                         | SB-4                | 11/02/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4519-05 | Percent Solids | Cool 4 deg C | POWE02   | D11                         | SB-5                | 11/02/2021   | Chemtech -SO |
| 11/09/2021 | Solid  | M4520-01 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | 40540               | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4523-01 | Percent Solids | Cool 4 deg C | ENTE01   | D11                         | VWE-B4-110421-0-10  | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4523-02 | Percent Solids | Cool 4 deg C | ENTE01   | D11                         | VWE-B4-110421-10-20 | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4523-03 | Percent Solids | Cool 4 deg C | ENTE01   | D11                         | VWE-B4-110421-20-30 | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4523-04 | Percent Solids | Cool 4 deg C | ENTE01   | D11                         | VWE-B4-110421-0-1   | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4523-05 | Percent Solids | Cool 4 deg C | ENTE01   | D11                         | VWE-B4-110421-14-15 | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4523-06 | Percent Solids | Cool 4 deg C | ENTE01   | D11                         | VWE-B4-110421-29-30 | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4523-09 | Percent Solids | Cool 4 deg C | ENTE01   | D11                         | VWE-B4-110421-BD    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4523-10 | Percent Solids | Cool 4 deg C | ENTE01   | D11                         | VWE-B4-110421-BD    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-01 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-26A        | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-02 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-26B        | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-03 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-27A        | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-04 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-27B        | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-05 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-28A        | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-06 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-28B        | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-07 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-29A        | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-08 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-29B        | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-09 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-30A        | 11/04/2021   | Chemtech -SO |

Date/Time 11/05/21 15:40  
 Raw Sample Received by: JH 160C1  
 Raw Sample Relinquished by: KS (COC)

Date/Time 11/05/21 15:335  
 Raw Sample Received by: KS (COC)  
 Raw Sample Relinquished by: JH 160C1



## WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-110521

WorkList ID : 153980

Department : Wet-Chemistry

Date : 11-05-2021 08:08:50

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|-----------------|--------------|--------------|
| 11/11/2021 | Solid  | M4525-10 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-30B    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-11 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-31A    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-12 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-31B    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-13 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-32A    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4525-14 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-32B    | 11/04/2021   | Chemtech -SO |
| 11/05/2021 | Solid  | M4529-01 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | 6-INCH-STEEL    | 11/04/2021   | Chemtech -SO |
| 11/05/2021 | Solid  | M4529-02 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | 12-INCH-STEEL   | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-01 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-33A    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-02 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-33B    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-03 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-34A    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-04 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-34B    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-05 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-35A    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-06 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-35B    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-07 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-36B    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-08 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-37A    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-09 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-38B    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-10 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-39A    | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4536-11 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SEC-2021-39B    | 11/04/2021   | Chemtech -SO |
| 11/01/2021 | Solid  | M4537-01 | Percent Solids | Cool 4 deg C | EART12   | D11                         | S1-S2-S3        | 10/22/2021   | Chemtech -SO |
| 11/09/2021 | Solid  | M4538-01 | Percent Solids | Cool 4 deg C | EART12   | D11                         | WC-C1-COMP      | 11/03/2021   | Chemtech -SO |
| 11/09/2021 | Solid  | M4539-01 | Percent Solids | Cool 4 deg C | EART12   | L11                         | WC-C2-COMP      | 11/02/2021   | Chemtech -SO |
| 11/09/2021 | Solid  | M4539-03 | Percent Solids | Cool 4 deg C | EART12   | L11                         | M4539-01MS      | 11/02/2021   | Chemtech -SO |

Date/Time 11/05/21

Raw Sample Received by: 1540 30 (WC)

Raw Sample Relinquished by: 1540 30 (WC)

Date/Time 11/05/21

Raw Sample Received by: 17:35 30 (WC)

Raw Sample Relinquished by: 17:35 30 (WC)



# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-110521      WorkList ID : 153980      Department : Wet-Chemistry      Date : 11-05-2021 08:08:50

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|-----------------|--------------|--------------|
| 11/09/2021 | Solid  | M4539-04 | Percent Solids | Cool 4 deg C | EART12   | L11                         | M4539-01MSD     | 11/02/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-01 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-1            | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-02 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-1-EPH-2      | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-03 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-2            | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-04 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-2-EPH-2      | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-05 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-3            | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-06 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-3-EPH-2      | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-07 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-4            | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-08 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-4-EPH-2      | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-09 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-5            | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-10 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-5-EPH-2      | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-11 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-6            | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4540-12 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | TP-6-EPH-2      | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4541-01 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | WC-3-AS         | 11/04/2021   | Chemtech -SO |
| 11/11/2021 | Solid  | M4541-03 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | WC-3-AS-EPH-2   | 11/04/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4544-01 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | MOO-2021-10A    | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4544-02 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | MOO-2021-10B    | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4544-03 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | MOO-2021-10C    | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4544-04 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | MOO-2021-11A    | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4544-05 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | MOO-2021-11B    | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4544-06 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | MOO-2021-11C    | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4544-07 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | MOO-2021-11D    | 11/05/2021   | Chemtech -SO |

Date/Time 11/05/21 15:40      Date/Time 11/05/21 17:35  
 Raw Sample Received by: SD (WC)      Raw Sample Received by: KS (WC)  
 Raw Sample Relinquished by: KS (WC)      Raw Sample Relinquished by: JD (WC)

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-110521

WorkList ID : 153980

Department : Wet-Chemistry

Date : 11-05-2021 08:08:50

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample    | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|--------------------|--------------|--------------|
| 11/10/2021 | Solid  | M4548-01 | Percent Solids | Cool 4 deg C | PSEG03   | L12                         | LAW-2021-1A        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4548-02 | Percent Solids | Cool 4 deg C | PSEG03   | L12                         | LAW-2021-2A        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4548-03 | Percent Solids | Cool 4 deg C | PSEG03   | L12                         | LAW-2021-2B        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4548-04 | Percent Solids | Cool 4 deg C | PSEG03   | L12                         | LAW-2021-3A        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4548-05 | Percent Solids | Cool 4 deg C | PSEG03   | L12                         | LAW-2021-3B        | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4561-01 | Percent Solids | Cool 4 deg C | PSEG03   | L12                         | ETGI-280           | 11/05/2021   | Chemtech -SO |
| 11/13/2021 | Solid  | M4562-01 | Percent Solids | Cool 4 deg C | EARTH03  | L12                         | T-1                | 11/03/2021   | Chemtech -SO |
| 11/13/2021 | Solid  | M4562-02 | Percent Solids | Cool 4 deg C | EARTH03  | L12                         | T-1A               | 11/03/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4563-01 | Percent Solids | Cool 4 deg C | EARTH03  | L12                         | S-1                | 11/03/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4569-01 | Percent Solids | Cool 4 deg C | PSEG05   | L12                         | CL-02-110521       | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4569-02 | Percent Solids | Cool 4 deg C | PSEG05   | L12                         | CL-02-110521-EPH-2 | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-01 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-1A        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-02 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-1B        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-03 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-2A        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-04 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-2B        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-05 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-3A        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-06 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-3B        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-07 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-4A        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-08 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-5A        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-09 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-5B        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-10 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-6A        | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-11 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | SOM-2021-6B        | 11/05/2021   | Chemtech -SO |

Date/Time 11/05/21 15:40  
 Raw Sample Received by: KS (WC)  
 Raw Sample Relinquished by: KS (WC)

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-110521 WorkList ID : 153980 Department : Wet-Chemistry Date : 11-05-2021 08:08:50

| Due Date   | Matrix | Sample   | Test           | Preservative | Customer | Raw Sample Storage Location | Customer Sample     | Collect Date | Method       |
|------------|--------|----------|----------------|--------------|----------|-----------------------------|---------------------|--------------|--------------|
| 11/10/2021 | Solid  | M4570-12 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | KMA79SC-A           | 11/05/2021   | Chemtech -SO |
| 11/10/2021 | Solid  | M4570-13 | Percent Solids | Cool 4 deg C | PSEG03   | L11                         | KMA79SC-B           | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4574-01 | Percent Solids | Cool 4 deg C | ENTE01   | L12                         | VWE-B6-110521-0-10  | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4574-02 | Percent Solids | Cool 4 deg C | ENTE01   | L12                         | VWE-B6-110521-10-20 | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4574-03 | Percent Solids | Cool 4 deg C | ENTE01   | L12                         | VWE-B6-110521-20-30 | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4574-04 | Percent Solids | Cool 4 deg C | ENTE01   | L12                         | VWE-B6-110521-0-1   | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4574-05 | Percent Solids | Cool 4 deg C | ENTE01   | L12                         | VWE-B6-110521-14-15 | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4574-06 | Percent Solids | Cool 4 deg C | ENTE01   | L12                         | VWE-B6-110521-29-30 | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4576-01 | Percent Solids | Cool 4 deg C | AEIC01   | L12                         | SB-1-(12-12.5)      | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4576-03 | Percent Solids | Cool 4 deg C | AEIC01   | L12                         | SB-2-(9-9.5)        | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4576-04 | Percent Solids | Cool 4 deg C | AEIC01   | L12                         | DW-2                | 11/05/2021   | Chemtech -SO |
| 11/12/2021 | Solid  | M4576-05 | Percent Solids | Cool 4 deg C | AEIC01   | L12                         | DW-3                | 11/05/2021   | Chemtech -SO |

Date/Time 11/05/21 13:40  
 Raw Sample Received by: 30 (wcc)  
 Raw Sample Relinquished by: KS (wcc)

Date/Time 11/05/21 17:35  
 Raw Sample Received by: KS (wcc)  
 Raw Sample Relinquished by: 30 (wcc)