


LB117123

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/00
0.0	0.000	0.062	0.062	5	0.000	5					1
0.2	0.200	0.205	0.005	542	0.0 %	542					2
2.5	2.500	2.407	-0.093	8847	0.0 %	8847					-4
5.0	5.000	5.021	0.021	18708	0.0 %	18708					0
7.5	7.500	7.476	-0.024	27967	0.0 %	27967					0
10.0	10.000	10.030	0.030	37601	0.0 %	37601					0

LB117183 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Nov 2021 11:00:39

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	5	-	0.0000	7470A	PPB	05 Nov 2021 11:03:50	S	Std
0.2 - 1	50.2	542	-	0.2000	7470A	PPB	05 Nov 2021 11:06:07	S	Std
2.5 - 1	50.5	8847	-	2.5000	7470A	PPB	05 Nov 2021 11:08:24	S	Std
5.0 - 1	55	18708	-	5.0000	7470A	PPB	05 Nov 2021 11:10:40	S	Std
7.5 - 1	57.5	27967	-	7.5000	7470A	PPB	05 Nov 2021 11:12:56	S	Std
10.0 - 1	580	37601	-	10.0000	7470A	PPB	05 Nov 2021 11:15:12	S	Std
ICV70 - 1	ICV70	13600	3.6667	-7470A	PPB	05 Nov 2021 11:17:59	U	SMPL	
ICB70 - 1	ICB70	-114	0.0309	-7470A	PPB	05 Nov 2021 11:20:13	U	SMPL	
CCV12 - 1	CCV12	18650	5.0055	-7470A	PPB	05 Nov 2021 11:22:28	U	SMPL	
CCB12 - 1	CCB12	-110	0.0320	-7470A	PPB	05 Nov 2021 11:24:43	U	SMPL	
CRA - 1	CRA	537	0.2035	-7470A	PPB	05 Nov 2021 11:26:58	U	SMPL	
HighStd - 1	HighStd	35521	9.4782	-7470A	PPB	05 Nov 2021 11:29:13	U	SMPL	
ChkStd - 1	ChkStd	25885	6.9236	-7470A	PPB	05 Nov 2021 11:31:27	U	SMPL	
PB140509BL - 1	PBW	-104	0.0336	-7470A	PPB	05 Nov 2021 11:33:45	U	SMPL	
PB140509BS - 1	LCSW	12637	3.4114	-7470A	PPB	05 Nov 2021 11:38:37	U	SMPL	
M4475-07 - 1	WC01	-61	0.0450	-7470A	PPB	05 Nov 2021 11:40:52	U	SMPL	
M4475-08 - 1	WC02	24	0.0675	-7470A	PPB	05 Nov 2021 11:43:09	U	SMPL	
M4475-09 - 1	WC03	32	0.0696	-7470A	PPB	05 Nov 2021 11:45:26	U	SMPL	
M4480-01 - 1	1103-4-5	4	0.0622	-7470A	PPB	05 Nov 2021 11:47:43	U	SMPL	
M4497-02 - 1	RO-COMP-1	8	0.0632	-7470A	PPB	05 Nov 2021 11:49:59	U	SMPL	
CCV13 - 1	CCV13	17798	4.7796	-7470A	PPB	05 Nov 2021 11:52:14	U	SMPL	
CCB13 - 1	CCB13	-88	0.0378	-7470A	PPB	05 Nov 2021 11:54:29	U	SMPL	
M4497-05 - 1	VNJ-203	2	0.0617	-7470A	PPB	05 Nov 2021 11:56:45	U	SMPL	
M4497-08 - 1	346	15	0.0651	-7470A	PPB	05 Nov 2021 11:59:00	U	SMPL	
M4498-02 - 1	WC-2-MS	10	0.0638	-7470A	PPB	05 Nov 2021 12:01:16	U	SMPL	
M4498-05 - 1	WC-2-MD	30	0.0691	-7470A	PPB	05 Nov 2021 12:03:31	U	SMPL	
M4498-08 - 1	WC-2-LS	149	0.1006	-7470A	PPB	05 Nov 2021 12:05:47	U	SMPL	
M4499-02 - 1	WEL-BRIDGE	17	0.0656	-7470A	PPB	05 Nov 2021 12:08:03	U	SMPL	
M4501-02 - 1	72-11929	33	0.0699	-7470A	PPB	05 Nov 2021 12:10:19	U	SMPL	
M4501-02DUP - 1	72-11929DUP	36	0.0707	-7470A	PPB	05 Nov 2021 12:12:36	U	SMPL	
M4501-02MS - 1	72-11929MS	13440	3.6243	-7470A	PPB	05 Nov 2021 12:14:53	U	SMPL	
M4501-02MSD - 1	72-11929MSD	12083	3.2645	-7470A	PPB	05 Nov 2021 12:17:09	U	SMPL	
CCV14 - 1	CCV14	17652	4.7409	-7470A	PPB	05 Nov 2021 12:19:24	U	SMPL	
CCB14 - 1	CCB14	-93	0.0365	-7470A	PPB	05 Nov 2021 12:21:40	U	SMPL	
PB140510BL - 1	PBW	-9	0.0587	-7470A	PPB	05 Nov 2021 12:23:55	U	SMPL	
PB140510BS - 1	LCSW	12634	3.4106	-7470A	PPB	05 Nov 2021 12:26:11	U	SMPL	
M4454-11 - 1	11221-B	-52	0.0473	-7470A	PPB	05 Nov 2021 12:28:27	U	SMPL	
M4475-10 - 1	TWP02	5626	1.5527	-7470A	PPB	05 Nov 2021 12:35:49	U	SMPL	
M4475-10DUP - 1	TWP02DUP	4996	1.3856	-7470A	PPB	05 Nov 2021 12:38:04	U	SMPL	
M4475-10MS - 1	TWP02MS	18668	5.0103	-7470A	PPB	05 Nov 2021 12:40:21	U	SMPL	
M4475-10MSD - 1	TWP02MSD	18372	4.9318	-7470A	PPB	05 Nov 2021 12:42:37	U	SMPL	
M4475-11 - 1	TWP02	-35	0.0518	-7470A	PPB	05 Nov 2021 12:44:55	U	SMPL	
M4475-12 - 1	DUP02	5407	1.4946	-7470A	PPB	05 Nov 2021 12:53:13	U	SMPL	
M4475-13 - 1	DUP02	9	0.0635	-7470A	PPB	05 Nov 2021 12:55:28	U	SMPL	
CCV15 - 1	CCV15	17368	4.6656	-7470A	PPB	05 Nov 2021 13:24:46	U	SMPL	
CCB15 - 1	CCB15	-74	0.0415	-7470A	PPB	05 Nov 2021 13:27:01	U	SMPL	
M4475-14 - 1	FB01	-3	0.0603	-7470A	PPB	05 Nov 2021 13:29:16	U	SMPL	
M4485-10 - 1	AG-3	-24	0.0548	-7470A	PPB	05 Nov 2021 13:31:32	U	SMPL	
M4485-11 - 1	AG-7	5	0.0625	-7470A	PPB	05 Nov 2021 13:33:47	U	SMPL	
M4485-12 - 1	AG-9	0	0.0611	-7470A	PPB	05 Nov 2021 13:36:03	U	SMPL	
M4530-01 - 1	SW-P287	28	0.0686	-7470A	PPB	05 Nov 2021 13:38:19	U	SMPL	
PB140509TB - 1	PB140487TB	8	0.0632	-7470A	PPB	05 Nov 2021 13:40:35	U	SMPL	
M4501-02LX5 - 1		20	0.0664	-7470A	PPB	05 Nov 2021 13:42:51	U	SMPL	
M4501-02A - 1		12295	3.3207	-7470A	PPB	05 Nov 2021 13:45:08	U	SMPL	
M4475-10LX5 - 1		993	0.3244	-7470A	PPB	05 Nov 2021 13:47:25	U	SMPL	
M4475-10A - 1		18416	4.9435	-7470A	PPB	05 Nov 2021 13:49:41	U	SMPL	
CCV16 - 1	CCV16	17318	4.6524	-7470A	PPB	05 Nov 2021 13:51:57	U	SMPL	
CCB16 - 1	CCB16	-58	0.0458	-7470A	PPB	05 Nov 2021 13:54:13	U	SMPL	